

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040224\
 Data File : PP064026.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Apr 2024 16:30
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 02 17:22:37 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 19 06:23:02 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.487	3.637	114.2E6	119.8E6	47.862	50.195
2) SA Decachlor...	10.357	8.670	123.3E6	116.8E6	48.570	43.215
Target Compounds						
16) L4 AR-1242-1	5.668	4.728	30786052	29935159	510.402	507.807
17) L4 AR-1242-2	5.691	4.746	44959473	43093101	523.564	526.576
18) L4 AR-1242-3	5.754	4.923	28759731	23580993	511.675	525.397
19) L4 AR-1242-4	5.853	5.007	23383189	25428434	519.685	512.321
20) L4 AR-1242-5	6.597	5.533	26045580	32085319	581.583	549.797



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040224\
Data File : PP064026.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Apr 2024 16:30
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 02 17:22:37 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 19 06:23:02 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.487 min
Delta R.T.: -0.002 min
Response: 114205924
Conc: 47.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.637 min
Delta R.T.: -0.003 min
Response: 119841232
Conc: 50.19 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.357 min
Delta R.T.: -0.009 min
Response: 123319395
Conc: 48.57 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.670 min
Delta R.T.: -0.014 min
Response: 116790226
Conc: 43.22 ng/ml



#16 AR-1242-1


R.T.: 5.668 min
Delta R.T.: -0.005 min
Response: 30786052
Conc: 510.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.728 min
Delta R.T.: -0.006 min
Response: 29935159
Conc: 507.81 ng/ml




#17 AR-1242-2

R.T.: 5.691 min
Delta R.T.: -0.006 min
Response: 44959473
Conc: 523.56 ng/ml



#17 AR-1242-2


R.T.: 4.746 min
Delta R.T.: -0.006 min
Response: 43093101
Conc: 526.58 ng/ml



#18 AR-1242-3


R.T.: 5.754 min
Delta R.T.: -0.006 min
Response: 28759731
Conc: 511.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.923 min
Delta R.T.: -0.006 min
Response: 23580993
Conc: 525.40 ng/ml




#19 AR-1242-4

R.T.: 5.853 min
Delta R.T.: -0.006 min
Response: 23383189
Conc: 519.68 ng/ml



#19 AR-1242-4


R.T.: 5.007 min
Delta R.T.: -0.006 min
Response: 25428434
Conc: 512.32 ng/ml



#20 AR-1242-5

R.T.: 6.597 min
Delta R.T.: -0.006 min
Response: 26045580
Conc: 581.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.533 min
Delta R.T.: -0.007 min
Response: 32085319
Conc: 549.80 ng/ml