

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP040323.M
 Title : GC EXTRACTABLES
 Last Update : Tue Apr 04 02:33:08 2023
 Response Via : Initial Calibration

Calibration Files

500 =PP056617.D 1000 =PP056615.D 750 =PP056616.D
 250 =PP056618.D 50 =PP056619.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.066	2.002	2.029	2.056	1.895	2.010	E6	3.42
2) SA	Decachlorobiphenyl	1.034	0.960	0.993	1.024	0.945	0.991	E6	3.91
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1	6.633	6.107	6.353	6.940	6.878	6.582	E4	5.35
27) L6	AR-1254-2	9.447	8.731	9.020	9.845	9.612	9.331	E4	4.83
28) L6	AR-1254-3	8.986	8.481	8.681	9.165	8.851	8.833	E4	3.00
29) L6	AR-1254-4	5.861	5.488	5.624	5.976	5.743	5.738	E4	3.34
30) L6	AR-1254-5	6.374	5.971	6.126	6.489	6.106	6.213	E4	3.41
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

500 =PP056617.D 1000 =PP056615.D 750 =PP056616.D
 250 =PP056618.D 50 =PP056619.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.657	1.575	1.608	1.692	1.601	1.627	E6	2.88
2) SA	Decachlorobiphenyl	1.255	1.162	1.201	1.327	1.338	1.257	E6	6.12
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP040323.M
 Title : GC EXTRACTABLES
 Last Update : Tue Apr 04 02:33:08 2023
 Response Via : Initial Calibration

Calibration Files

500 =PP056617.D 1000 =PP056615.D 750 =PP056616.D
 250 =PP056618.D 50 =PP056619.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	8.370	7.683	7.963	8.850	8.802	8.334 E4	6.14
27) L6	AR-1254-2	7.271	6.603	6.891	7.738	7.587	7.218 E4	6.55
28) L6	AR-1254-3	1.116	1.031	1.064	1.158	1.112	1.096 E5	4.50
29) L6	AR-1254-4	5.993	5.452	5.644	6.286	5.879	5.851 E4	5.48
30) L6	AR-1254-5	9.121	8.387	8.698	9.617	9.277	9.020 E4	5.37
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range