

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP040323.M
 Title : GC EXTRACTABLES
 Last Update : Tue Apr 04 03:10:21 2023
 Response Via : Initial Calibration

Calibration Files

500 =PP056623.D 1000 =PP056621.D 750 =PP056622.D
 250 =PP056624.D 50 =PP056625.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.076	2.021	2.042	2.084	1.895	2.024 E6	3.78
2) SA	Decachlorobiphenyl	1.836	1.767	1.781	1.885	1.659	1.785 E6	4.76
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1	2.470					2.470 E4	0.00
9) L2	AR-1221-2	1.826					1.826 E4	0.00
10) L2	AR-1221-3	5.465					5.465 E4	0.00
11) L3	AR-1232-1	4.621					4.621 E4	0.00
12) L3	AR-1232-2	2.349					2.349 E4	0.00
13) L3	AR-1232-3	3.811					3.811 E4	0.00
14) L3	AR-1232-4	1.893					1.893 E4	0.00
15) L3	AR-1232-5	1.637					1.637 E4	0.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1	8.227					8.227 E4	0.00
37) L8	AR-1262-2	1.257					1.257 E5	0.00
38) L8	AR-1262-3	8.657					8.657 E4	0.00
39) L8	AR-1262-4	4.302					4.302 E4	0.00
40) L8	AR-1262-5	4.035					4.035 E4	0.00
41) L9	AR-1268-1	1.698	1.641	1.663	1.700	1.528	1.646 E5	4.28
42) L9	AR-1268-2	1.496	1.440	1.476	1.563	1.508	1.497 E5	3.02
43) L9	AR-1268-3	1.412	1.367	1.345	1.427	1.248	1.360 E5	5.20
44) L9	AR-1268-4	4.601	4.424	4.445	4.573	4.135	4.436 E4	4.17
45) L9	AR-1268-5	3.782	3.761	3.815	3.843	3.255	3.691 E5	6.66

Signal #2 Calibration Files

500 =PP056623.D 1000 =PP056621.D 750 =PP056622.D
 250 =PP056624.D 50 =PP056625.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.675	1.586	1.618	1.717	1.623	1.644 E6	3.15
2) SA	Decachlorobiphenyl	2.210	2.086	2.136	2.327	2.357	2.223 E6	5.29
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP040323.M
 Title : GC EXTRACTABLES
 Last Update : Tue Apr 04 03:10:21 2023
 Response Via : Initial Calibration

Calibration Files

500 =PP056623.D 1000 =PP056621.D 750 =PP056622.D
 250 =PP056624.D 50 =PP056625.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1	2.072					2.072 E4	0.00
9) L2	AR-1221-2	1.536					1.536 E4	0.00
10) L2	AR-1221-3	4.570					4.570 E4	0.00
11) L3	AR-1232-1	3.874					3.874 E4	0.00
12) L3	AR-1232-2	3.140					3.140 E4	0.00
13) L3	AR-1232-3	1.661					1.661 E4	0.00
14) L3	AR-1232-4	1.708					1.708 E4	0.00
15) L3	AR-1232-5	1.875					1.875 E4	0.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1	4.239					4.239 E4	0.00
37) L8	AR-1262-2	8.888					8.888 E4	0.00
38) L8	AR-1262-3	6.589					6.589 E4	0.00
39) L8	AR-1262-4	1.142					1.142 E5	0.00
40) L8	AR-1262-5	4.998					4.998 E4	0.00
41) L9	AR-1268-1	1.923	1.841	1.869	2.012	1.972	1.923 E5	3.68
42) L9	AR-1268-2	1.762	1.694	1.757	1.828	1.773	1.763 E5	2.70
43) L9	AR-1268-3	1.653	1.584	1.606	1.715	1.728	1.657 E5	3.84
44) L9	AR-1268-4	5.674	5.397	5.545	6.035	5.760	5.682 E4	4.22
45) L9	AR-1268-5	4.363	4.260	4.327	4.474	4.312	4.347 E5	1.84

(#) = Out of Range