

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040323\
 Data File : PP056624.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Apr 2023 23:43
 Operator : YP\AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 04 03:05:19 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP040323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 04 03:05:08 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.435	3.664	52108705	42917340	25.347	26.027
2) SA Decachlor...	10.271	8.754	47126792	58167980	25.934	26.564
Target Compounds						
41) L9 AR-1268-1	8.718	7.620	42492726	50306545	253.613	263.208
42) L9 AR-1268-2	8.814	7.685	39074410	45692389	261.565	259.595
43) L9 AR-1268-3	9.051	7.896	35669433	42863484	257.034	261.439
44) L9 AR-1268-4	9.486	8.184	11432566	15086370	253.442	266.423
45) L9 AR-1268-5	9.919	8.487	96082526	111.8E6	252.830	256.757



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040323\
Data File : PP056624.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Apr 2023 23:43
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 04 03:05:19 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP040323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 04 03:05:08 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.435 min
Delta R.T.: 0.000 min
Response: 52108705
Conc: 25.35 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.664 min
Delta R.T.: 0.000 min
Response: 42917340
Conc: 26.03 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.271 min
Delta R.T.: 0.000 min
Response: 47126792
Conc: 25.93 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.754 min
Delta R.T.: 0.000 min
Response: 58167980
Conc: 26.56 ng/ml



#41 AR-1268-1


R.T.: 8.718 min
Delta R.T.: 0.000 min
Response: 42492726
Conc: 253.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#41 AR-1268-1

R.T.: 7.620 min
Delta R.T.: 0.000 min
Response: 50306545
Conc: 263.21 ng/ml




#42 AR-1268-2

R.T.: 8.814 min
Delta R.T.: 0.000 min
Response: 39074410
Conc: 261.57 ng/ml



#42 AR-1268-2


R.T.: 7.685 min
Delta R.T.: 0.000 min
Response: 45692389
Conc: 259.59 ng/ml



#43 AR-1268-3


R.T.: 9.051 min
Delta R.T.: 0.000 min
Response: 35669433
Conc: 257.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#43 AR-1268-3

R.T.: 7.896 min
Delta R.T.: 0.000 min
Response: 42863484
Conc: 261.44 ng/ml




#44 AR-1268-4

R.T.: 9.486 min
Delta R.T.: 0.000 min
Response: 11432566
Conc: 253.44 ng/ml



#44 AR-1268-4


R.T.: 8.184 min
Delta R.T.: 0.000 min
Response: 15086370
Conc: 266.42 ng/ml



#45 AR-1268-5

R.T.: 9.919 min
Delta R.T.: 0.000 min
Response: 96082526
Conc: 252.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250



#45 AR-1268-5

R.T.: 8.487 min
Delta R.T.: 0.000 min
Response: 111842325
Conc: 256.76 ng/ml