

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040323\
 Data File : PP056629.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Apr 2023 01:05
 Operator : YP\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP040323AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 04 02:34:05 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP040323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 04 02:33:08 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.435	3.664	105.4E6	84501278	52.429	51.953
2) SA Decachlor...	10.272	8.755	53601027	66176031	54.069	52.665
Target Compounds						
26) L6 AR-1254-1	6.472	5.580	33807704	43001848	513.618	516.000
27) L6 AR-1254-2	6.692	5.729	48201394	37673421	516.572	521.938
28) L6 AR-1254-3	7.061	6.138	46105985	57563588	521.985	525.102
29) L6 AR-1254-4	7.347	6.368	30274615	30805929	527.575	526.528
30) L6 AR-1254-5	7.769	6.791	33021509	47553063	531.471	527.215



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040323\
Data File : PP056629.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Apr 2023 01:05
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP040323AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 04 02:34:05 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP040323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 04 02:33:08 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.435 min
Delta R.T.: 0.000 min
Response: 105370202
Conc: 52.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP040323AR1254




#1 Tetrachloro-m-xylene

R.T.: 3.664 min
Delta R.T.: 0.000 min
Response: 84501278
Conc: 51.95 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.272 min
Delta R.T.: 0.000 min
Response: 53601027
Conc: 54.07 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.755 min
Delta R.T.: 0.000 min
Response: 66176031
Conc: 52.67 ng/ml



#26 AR-1254-1


R.T.: 6.472 min
Delta R.T.: 0.000 min
Response: 33807704
Conc: 513.62 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP040323AR1254




#26 AR-1254-1

R.T.: 5.580 min
Delta R.T.: 0.000 min
Response: 43001848
Conc: 516.00 ng/ml




#27 AR-1254-2

R.T.: 6.692 min
Delta R.T.: 0.000 min
Response: 48201394
Conc: 516.57 ng/ml



#27 AR-1254-2


R.T.: 5.729 min
Delta R.T.: 0.000 min
Response: 37673421
Conc: 521.94 ng/ml



#28 AR-1254-3


R.T.: 7.061 min
Delta R.T.: 0.000 min
Response: 46105985
Conc: 521.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP040323AR1254




#28 AR-1254-3

R.T.: 6.138 min
Delta R.T.: 0.000 min
Response: 57563588
Conc: 525.10 ng/ml




#29 AR-1254-4

R.T.: 7.347 min
Delta R.T.: 0.000 min
Response: 30274615
Conc: 527.58 ng/ml



#29 AR-1254-4


R.T.: 6.368 min
Delta R.T.: 0.000 min
Response: 30805929
Conc: 526.53 ng/ml



#30 AR-1254-5

R.T.: 7.769 min
Delta R.T.: 0.000 min
Response: 33021509
Conc: 531.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP040323AR1254



#30 AR-1254-5

R.T.: 6.791 min
Delta R.T.: 0.000 min
Response: 47553063
Conc: 527.22 ng/ml