

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040422\
 Data File : PP045503.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Apr 2022 03:36
 Operator : YP\AJ
 Sample : N2234-02 10X
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 05 06:57:51 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 31 07:54:52 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.887	3.145	3760298	3140567	1.558	1.985 #
2)	SA Decachlor...	9.998	8.492	3119003	3286148	2.177	2.167
Target Compounds							
3)	L1 AR-1016-1	5.140	4.297	90674	12181	1.261	0.205 #
4)	L1 AR-1016-2	5.155	4.309	166528	23987	1.551	0.285 #
5)	L1 AR-1016-3	5.227	4.490	188992	18389	2.873	0.406 #
6)	L1 AR-1016-4	5.328	4.539	91592	5621	1.672	0.152 #
7)	L1 AR-1016-5	5.657	4.761	73613	11664	1.421	0.244 #
8)	L2 AR-1221-1	4.122	3.373	71810	14912	2.860	0.717 #
9)	L2 AR-1221-2	4.199	3.463	101510	26808	5.294	1.693 #
10)	L2 AR-1221-3	4.277	3.544	72183	17881	1.201	0.382 #
11)	L3 AR-1232-1	4.277	3.544	72183	17881	1.464	0.487 #
12)	L3 AR-1232-2	4.850	4.309	37605	23987	1.707	0.703 #
13)	L3 AR-1232-3	5.155	4.490	166528	18389	3.868	1.005 #
14)	L3 AR-1232-4	5.340	4.582	66678	3455	3.051	0.211 #
15)	L3 AR-1232-5	5.441	4.761	36288	11664	2.278	0.624 #
16)	L4 AR-1242-1	5.140	4.297	90674	12181	1.558	0.250 #
17)	L4 AR-1242-2	5.155	4.309	166528	23987	1.919	0.358 #
18)	L4 AR-1242-3	5.227	4.490	188992	18389	3.493	0.499 #
19)	L4 AR-1242-4	5.328	4.582	91592	3455	2.031	0.096 #
20)	L4 AR-1242-5	6.138	5.136	50910	51483	1.098	1.083
21)	L5 AR-1248-1	5.140	4.278	90674	19069	1.981	0.500 #
22)	L5 AR-1248-2	5.441	4.539	36288	5621	0.590	0.109 #
23)	L5 AR-1248-3	5.657	4.582	73613	3455	1.029	0.064 #
24)	L5 AR-1248-4	6.107f	4.761	191951	11664	2.553	0.181 #
25)	L5 AR-1248-5	6.138	5.198f	50910	19392	0.678	0.298 #
26)	L6 AR-1254-1	6.065	5.136	268159	51483	3.766	0.529 #
27)	L6 AR-1254-2	6.298	5.297	137448	21756	1.272	0.267 #
28)	L6 AR-1254-3	6.700	5.729	472466	37367	4.261	0.302 #
29)	L6 AR-1254-4	6.996f	5.976	156446	60421	1.946	0.754 #
30)	L6 AR-1254-5	7.475	6.429	89471	53793	1.119	0.492 #
31)	L7 AR-1260-1	6.877	5.870	82274	60070	1.032	0.725 #
32)	L7 AR-1260-2	7.165	6.089	90838	1069699	1.008	10.979 #
33)	L7 AR-1260-3	7.560	6.242	67922	55620	0.980	0.601 #
34)	L7 AR-1260-4	7.819	6.749	193589	19246	2.509	0.251 #
35)	L7 AR-1260-5	8.154	7.013	155777	95918	1.103	0.542 #
36)	L8 AR-1262-1	7.560	6.473	67922	17159	0.673	0.153 #
37)	L8 AR-1262-2	8.154	6.749	155777	19246	0.926	0.191 #
38)	L8 AR-1262-3	8.475	7.320	60263	13037	0.521	0.161 #
39)	L8 AR-1262-4	8.577	7.396	249823	16536	4.086	0.112 #
40)	L8 AR-1262-5	9.266f	7.931	47606	35841	0.779	0.492 #
41)	L9 AR-1268-1	8.475	7.320	60263	13037	0.298	0.058 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040422\
 Data File : PP045503.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Apr 2022 03:36
 Operator : YP\AJ
 Sample : N2234-02 10X
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 05 06:57:51 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 31 07:54:52 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.577	7.396	249823	16536	1.337	0.082 #
43) L9	AR-1268-3	8.800	7.612	55663	44173	0.337	0.258
44) L9	AR-1268-4	9.229f	7.931	73807	35841	1.103	0.453 #
45) L9	AR-1268-5	9.665	8.240	134777	12301	0.260	0.023 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040422\
Data File : PP045503.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Apr 2022 03:36
Operator : YP\AJ
Sample : N2234-02 10X
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 05 06:57:51 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033022.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 31 07:54:52 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.887 min
Delta R.T.: 0.000 min
Response: 3760298
Conc: 1.56 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.145 min
Delta R.T.: -0.001 min
Response: 3140567
Conc: 1.98 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.998 min
Delta R.T.: -0.007 min
Response: 3119003
Conc: 2.18 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.492 min
Delta R.T.: -0.005 min
Response: 3286148
Conc: 2.17 ng/ml



#3 AR-1016-1


R.T.: 5.140 min
Delta R.T.: 0.002 min
Response: 90674
Conc: 1.26 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.297 min
Delta R.T.: 0.009 min
Response: 12181
Conc: 0.20 ng/ml




#4 AR-1016-2

R.T.: 5.155 min
Delta R.T.: -0.006 min
Response: 166528
Conc: 1.55 ng/ml



#4 AR-1016-2


R.T.: 4.309 min
Delta R.T.: 0.003 min
Response: 23987
Conc: 0.29 ng/ml



#5 AR-1016-3


R.T.: 5.227 min
Delta R.T.: 0.000 min
Response: 188992
Conc: 2.87 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.490 min
Delta R.T.: -0.001 min
Response: 18389
Conc: 0.41 ng/ml




#6 AR-1016-4

R.T.: 5.328 min
Delta R.T.: -0.006 min
Response: 91592
Conc: 1.67 ng/ml



#6 AR-1016-4


R.T.: 4.539 min
Delta R.T.: 0.000 min
Response: 5621
Conc: 0.15 ng/ml



#7 AR-1016-5


R.T.: 5.657 min
Delta R.T.: 0.004 min
Response: 73613
Conc: 1.42 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 4.761 min
Delta R.T.: -0.003 min
Response: 11664
Conc: 0.24 ng/ml




#8 AR-1221-1

R.T.: 4.122 min
Delta R.T.: 0.008 min
Response: 71810
Conc: 2.86 ng/ml



#8 AR-1221-1


R.T.: 3.373 min
Delta R.T.: 0.000 min
Response: 14912
Conc: 0.72 ng/ml



#9 AR-1221-2


R.T.: 4.199 min
Delta R.T.: -0.007 min
Response: 101510
Conc: 5.29 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.463 min
Delta R.T.: 0.000 min
Response: 26808
Conc: 1.69 ng/ml




#10 AR-1221-3

R.T.: 4.277 min
Delta R.T.: -0.009 min
Response: 72183
Conc: 1.20 ng/ml



#10 AR-1221-3


R.T.: 3.544 min
Delta R.T.: 0.002 min
Response: 17881
Conc: 0.38 ng/ml



#11 AR-1232-1


R.T.: 4.277 min
Delta R.T.: -0.010 min
Response: 72183
Conc: 1.46 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.544 min
Delta R.T.: 0.003 min
Response: 17881
Conc: 0.49 ng/ml




#12 AR-1232-2

R.T.: 4.850 min
Delta R.T.: 0.001 min
Response: 37605
Conc: 1.71 ng/ml



#12 AR-1232-2


R.T.: 4.309 min
Delta R.T.: 0.003 min
Response: 23987
Conc: 0.70 ng/ml



#13 AR-1232-3


R.T.: 5.155 min
Delta R.T.: -0.007 min
Response: 166528
Conc: 3.87 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.490 min
Delta R.T.: 0.000 min
Response: 18389
Conc: 1.01 ng/ml




#14 AR-1232-4

R.T.: 5.340 min
Delta R.T.: 0.006 min
Response: 66678
Conc: 3.05 ng/ml



#14 AR-1232-4


R.T.: 4.582 min
Delta R.T.: 0.000 min
Response: 3455
Conc: 0.21 ng/ml



#15 AR-1232-5


R.T.: 5.441 min
Delta R.T.: 0.006 min
Response: 36288
Conc: 2.28 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 4.761 min
Delta R.T.: -0.002 min
Response: 11664
Conc: 0.62 ng/ml




#16 AR-1242-1

R.T.: 5.140 min
Delta R.T.: 0.002 min
Response: 90674
Conc: 1.56 ng/ml



#16 AR-1242-1


R.T.: 4.297 min
Delta R.T.: 0.009 min
Response: 12181
Conc: 0.25 ng/ml



#17 AR-1242-2


R.T.: 5.155 min
Delta R.T.: -0.005 min
Response: 166528
Conc: 1.92 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.309 min
Delta R.T.: 0.003 min
Response: 23987
Conc: 0.36 ng/ml




#18 AR-1242-3

R.T.: 5.227 min
Delta R.T.: 0.000 min
Response: 188992
Conc: 3.49 ng/ml



#18 AR-1242-3


R.T.: 4.490 min
Delta R.T.: 0.000 min
Response: 18389
Conc: 0.50 ng/ml



#19 AR-1242-4


R.T.: 5.328 min
Delta R.T.: -0.005 min
Response: 91592
Conc: 2.03 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.582 min
Delta R.T.: 0.000 min
Response: 3455
Conc: 0.10 ng/ml




#20 AR-1242-5

R.T.: 6.138 min
Delta R.T.: 0.006 min
Response: 50910
Conc: 1.10 ng/ml



#20 AR-1242-5


R.T.: 5.136 min
Delta R.T.: -0.003 min
Response: 51483
Conc: 1.08 ng/ml



#21 AR-1248-1


R.T.: 5.140 min
Delta R.T.: 0.002 min
Response: 90674
Conc: 1.98 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.278 min
Delta R.T.: -0.009 min
Response: 19069
Conc: 0.50 ng/ml




#22 AR-1248-2

R.T.: 5.441 min
Delta R.T.: 0.007 min
Response: 36288
Conc: 0.59 ng/ml



#22 AR-1248-2


R.T.: 4.539 min
Delta R.T.: 0.000 min
Response: 5621
Conc: 0.11 ng/ml



#23 AR-1248-3


R.T.: 5.657 min
Delta R.T.: 0.004 min
Response: 73613
Conc: 1.03 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.582 min
Delta R.T.: 0.000 min
Response: 3455
Conc: 0.06 ng/ml




#24 AR-1248-4

R.T.: 6.107 min
Delta R.T.: 0.016 min
Response: 191951
Conc: 2.55 ng/ml



#24 AR-1248-4


R.T.: 4.761 min
Delta R.T.: -0.002 min
Response: 11664
Conc: 0.18 ng/ml



#25 AR-1248-5


R.T.: 6.138 min
Delta R.T.: 0.006 min
Response: 50910
Conc: 0.68 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.198 min
Delta R.T.: 0.016 min
Response: 19392
Conc: 0.30 ng/ml




#26 AR-1254-1

R.T.: 6.065 min
Delta R.T.: 0.003 min
Response: 268159
Conc: 3.77 ng/ml



#26 AR-1254-1


R.T.: 5.136 min
Delta R.T.: -0.003 min
Response: 51483
Conc: 0.53 ng/ml



#27 AR-1254-2


R.T.: 6.298 min
Delta R.T.: -0.002 min
Response: 137448
Conc: 1.27 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.297 min
Delta R.T.: -0.003 min
Response: 21756
Conc: 0.27 ng/ml




#28 AR-1254-3

R.T.: 6.700 min
Delta R.T.: -0.001 min
Response: 472466
Conc: 4.26 ng/ml



#28 AR-1254-3


R.T.: 5.729 min
Delta R.T.: 0.000 min
Response: 37367
Conc: 0.30 ng/ml



#29 AR-1254-4


R.T.: 6.996 min
Delta R.T.: -0.018 min
Response: 156446
Conc: 1.95 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 5.976 min
Delta R.T.: -0.003 min
Response: 60421
Conc: 0.75 ng/ml




#30 AR-1254-5

R.T.: 7.475 min
Delta R.T.: 0.000 min
Response: 89471
Conc: 1.12 ng/ml



#30 AR-1254-5


R.T.: 6.429 min
Delta R.T.: 0.000 min
Response: 53793
Conc: 0.49 ng/ml



#31 AR-1260-1


R.T.: 6.877 min
Delta R.T.: -0.004 min
Response: 82274
Conc: 1.03 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 5.870 min
Delta R.T.: 0.000 min
Response: 60070
Conc: 0.72 ng/ml




#32 AR-1260-2

R.T.: 7.165 min
Delta R.T.: 0.000 min
Response: 90838
Conc: 1.01 ng/ml



#32 AR-1260-2


R.T.: 6.089 min
Delta R.T.: 0.011 min
Response: 1069699
Conc: 10.98 ng/ml



#33 AR-1260-3


R.T.: 7.560 min
Delta R.T.: 0.000 min
Response: 67922
Conc: 0.98 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.242 min
Delta R.T.: 0.003 min
Response: 55620
Conc: 0.60 ng/ml




#34 AR-1260-4

R.T.: 7.819 min
Delta R.T.: 0.013 min
Response: 193589
Conc: 2.51 ng/ml



#34 AR-1260-4


R.T.: 6.749 min
Delta R.T.: -0.001 min
Response: 19246
Conc: 0.25 ng/ml



#35 AR-1260-5


R.T.: 8.154 min
Delta R.T.: 0.007 min
Response: 155777
Conc: 1.10 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.013 min
Delta R.T.: -0.006 min
Response: 95918
Conc: 0.54 ng/ml




#36 AR-1262-1

R.T.: 7.560 min
Delta R.T.: 0.000 min
Response: 67922
Conc: 0.67 ng/ml



#36 AR-1262-1


R.T.: 6.473 min
Delta R.T.: 0.002 min
Response: 17159
Conc: 0.15 ng/ml



#37 AR-1262-2


R.T.: 8.154 min
Delta R.T.: 0.006 min
Response: 155777
Conc: 0.93 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.749 min
Delta R.T.: 0.000 min
Response: 19246
Conc: 0.19 ng/ml




#38 AR-1262-3

R.T.: 8.475 min
Delta R.T.: -0.002 min
Response: 60263
Conc: 0.52 ng/ml



#38 AR-1262-3


R.T.: 7.320 min
Delta R.T.: -0.004 min
Response: 13037
Conc: 0.16 ng/ml



#39 AR-1262-4


R.T.: 8.577 min
Delta R.T.: 0.007 min
Response: 249823
Conc: 4.09 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.396 min
Delta R.T.: 0.002 min
Response: 16536
Conc: 0.11 ng/ml




#40 AR-1262-5

R.T.: 9.266 min
Delta R.T.: 0.019 min
Response: 47606
Conc: 0.78 ng/ml



#40 AR-1262-5


R.T.: 7.931 min
Delta R.T.: -0.006 min
Response: 35841
Conc: 0.49 ng/ml



#41 AR-1268-1


R.T.: 8.475 min
Delta R.T.: 0.002 min
Response: 60263
Conc: 0.30 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.320 min
Delta R.T.: -0.003 min
Response: 13037
Conc: 0.06 ng/ml




#42 AR-1268-2

R.T.: 8.577 min
Delta R.T.: 0.005 min
Response: 249823
Conc: 1.34 ng/ml



#42 AR-1268-2


R.T.: 7.396 min
Delta R.T.: 0.000 min
Response: 16536
Conc: 0.08 ng/ml



#43 AR-1268-3


R.T.: 8.800 min
Delta R.T.: -0.006 min
Response: 55663
Conc: 0.34 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.612 min
Delta R.T.: -0.003 min
Response: 44173
Conc: 0.26 ng/ml




#44 AR-1268-4

R.T.: 9.229 min
Delta R.T.: -0.017 min
Response: 73807
Conc: 1.10 ng/ml



#44 AR-1268-4


R.T.: 7.931 min
Delta R.T.: -0.005 min
Response: 35841
Conc: 0.45 ng/ml



#45 AR-1268-5

R.T.: 9.665 min
Delta R.T.: -0.001 min
Response: 134777
Conc: 0.26 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.240 min
Delta R.T.: 0.000 min
Response: 12301
Conc: 0.02 ng/ml