

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040422\
 Data File : PP045513.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Apr 2022 07:17
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 05 09:03:24 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 31 07:54:52 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.888	3.147	129.9E6	85666345	53.856	54.139
2)	SA Decachlor...	9.998	8.492	77565784	76446858	54.135	50.412
Target Compounds							
16)	L4 AR-1242-1	5.138	4.285	29681828	25107054	510.022	515.890
17)	L4 AR-1242-2	5.160	4.304	43853073	34564258	505.349	515.836
18)	L4 AR-1242-3	5.226	4.488	27208263	19326065	502.872	524.681
19)	L4 AR-1242-4	5.332	4.580	22441978	19054154	497.584	527.752
20)	L4 AR-1242-5	6.131	5.137	21846730	25075542	471.188	527.422



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040422\
Data File : PP045513.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Apr 2022 07:17
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 05 09:03:24 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033022.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 31 07:54:52 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.888 min
Delta R.T.: 0.001 min
Response: 129943221
Conc: 53.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.147 min
Delta R.T.: 0.000 min
Response: 85666345
Conc: 54.14 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.998 min
Delta R.T.: -0.006 min
Response: 77565784
Conc: 54.13 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.492 min
Delta R.T.: -0.006 min
Response: 76446858
Conc: 50.41 ng/ml



#16 AR-1242-1


R.T.: 5.138 min
Delta R.T.: 0.000 min
Response: 29681828
Conc: 510.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.285 min
Delta R.T.: -0.002 min
Response: 25107054
Conc: 515.89 ng/ml




#17 AR-1242-2

R.T.: 5.160 min
Delta R.T.: 0.000 min
Response: 43853073
Conc: 505.35 ng/ml



#17 AR-1242-2


R.T.: 4.304 min
Delta R.T.: -0.002 min
Response: 34564258
Conc: 515.84 ng/ml



#18 AR-1242-3


R.T.: 5.226 min
Delta R.T.: 0.000 min
Response: 27208263
Conc: 502.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.488 min
Delta R.T.: -0.003 min
Response: 19326065
Conc: 524.68 ng/ml




#19 AR-1242-4

R.T.: 5.332 min
Delta R.T.: -0.001 min
Response: 22441978
Conc: 497.58 ng/ml



#19 AR-1242-4


R.T.: 4.580 min
Delta R.T.: -0.003 min
Response: 19054154
Conc: 527.75 ng/ml



#20 AR-1242-5

R.T.: 6.131 min
Delta R.T.: -0.001 min
Response: 21846730
Conc: 471.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.137 min
Delta R.T.: -0.002 min
Response: 25075542
Conc: 527.42 ng/ml