

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040521\
 Data File : PP034671.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 05 Apr 2021 13:02
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 06 03:02:13 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 31 06:02:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.846	4.257	4045711	2467282	53.035	54.923
2) SA Decachlor...	10.767	9.559	3993956	1527142	51.883	54.085
Target Compounds						
3) L1 AR-1016-1	6.158	5.508	1201030	619292	510.472	522.025
4) L1 AR-1016-2	6.181	5.528	1825370	923893	513.941	520.318
5) L1 AR-1016-3	6.247	5.722	1175578	500387	550.546	535.530
6) L1 AR-1016-4	6.353	5.769	944433	410179	530.332	564.759
7) L1 AR-1016-5	6.666	5.999	898187	511718	520.587	559.465
31) L7 AR-1260-1	7.838	7.087	1599830	820554	515.362	534.333
32) L7 AR-1260-2	8.104	7.281	1876349	977823	499.638	541.312
33) L7 AR-1260-3	8.469	7.437	1536373	899473	508.274	518.464
34) L7 AR-1260-4	8.699	7.916	1778173	752611	509.326	524.920
35) L7 AR-1260-5	9.017	8.160	3449250	1720084	481.631	503.545



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040521\
Data File : PP034671.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Apr 2021 13:02
Operator : DD\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 06 03:02:13 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033021.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 31 06:02:19 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.846 min
Delta R.T.: 0.000 min
Response: 4045711
Conc: 53.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.257 min
Delta R.T.: -0.003 min
Response: 2467282
Conc: 54.92 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.767 min
Delta R.T.: 0.002 min
Response: 3993956
Conc: 51.88 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.559 min
Delta R.T.: -0.008 min
Response: 1527142
Conc: 54.08 ng/ml



#3 AR-1016-1


R.T.: 6.158 min
Delta R.T.: 0.000 min
Response: 1201030
Conc: 510.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 5.508 min
Delta R.T.: -0.004 min
Response: 619292
Conc: 522.03 ng/ml




#4 AR-1016-2

R.T.: 6.181 min
Delta R.T.: 0.000 min
Response: 1825370
Conc: 513.94 ng/ml



#4 AR-1016-2


R.T.: 5.528 min
Delta R.T.: -0.004 min
Response: 923893
Conc: 520.32 ng/ml



#5 AR-1016-3


R.T.: 6.247 min
Delta R.T.: 0.001 min
Response: 1175578
Conc: 550.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.722 min
Delta R.T.: -0.004 min
Response: 500387
Conc: 535.53 ng/ml




#6 AR-1016-4

R.T.: 6.353 min
Delta R.T.: 0.000 min
Response: 944433
Conc: 530.33 ng/ml



#6 AR-1016-4


R.T.: 5.769 min
Delta R.T.: -0.004 min
Response: 410179
Conc: 564.76 ng/ml



#7 AR-1016-5


R.T.: 6.666 min
Delta R.T.: 0.000 min
Response: 898187
Conc: 520.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.999 min
Delta R.T.: -0.004 min
Response: 511718
Conc: 559.47 ng/ml




#31 AR-1260-1

R.T.: 7.838 min
Delta R.T.: 0.000 min
Response: 1599830
Conc: 515.36 ng/ml



#31 AR-1260-1


R.T.: 7.087 min
Delta R.T.: -0.005 min
Response: 820554
Conc: 534.33 ng/ml



#32 AR-1260-2


R.T.: 8.104 min
Delta R.T.: 0.002 min
Response: 1876349
Conc: 499.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 7.281 min
Delta R.T.: -0.005 min
Response: 977823
Conc: 541.31 ng/ml




#33 AR-1260-3

R.T.: 8.469 min
Delta R.T.: 0.002 min
Response: 1536373
Conc: 508.27 ng/ml



#33 AR-1260-3


R.T.: 7.437 min
Delta R.T.: -0.005 min
Response: 899473
Conc: 518.46 ng/ml



#34 AR-1260-4


R.T.: 8.699 min
Delta R.T.: 0.002 min
Response: 1778173
Conc: 509.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.916 min
Delta R.T.: -0.006 min
Response: 752611
Conc: 524.92 ng/ml



#35 AR-1260-5

R.T.: 9.017 min
Delta R.T.: 0.001 min
Response: 3449250
Conc: 481.63 ng/ml



#35 AR-1260-5

R.T.: 8.160 min
Delta R.T.: -0.005 min
Response: 1720084
Conc: 503.54 ng/ml