

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040524\
 Data File : PP064141.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Apr 2024 02:24
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 06 02:52:22 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 19 06:23:02 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.484	3.635	110.4E6	114.1E6	46.252	47.771
2) SA Decachlor...	10.354	8.666	124.4E6	122.7E6	48.997	45.384
Target Compounds						
21) L5 AR-1248-1	5.666	4.725	23801376	22025409	509.467	494.486
22) L5 AR-1248-2	5.943	4.962	37466359	33634466	516.222	473.900
23) L5 AR-1248-3	6.149	5.005	40410150	36040578	500.471	487.033
24) L5 AR-1248-4	6.557	5.176	42830824	42610614	523.766	492.084
25) L5 AR-1248-5	6.596	5.571	44365124	40041984	543.980	533.340



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040524\
Data File : PP064141.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Apr 2024 02:24
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 06 02:52:22 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 19 06:23:02 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.484 min
Delta R.T.: -0.005 min
Response: 110365872
Conc: 46.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.635 min
Delta R.T.: -0.005 min
Response: 114055240
Conc: 47.77 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.354 min
Delta R.T.: -0.011 min
Response: 124404321
Conc: 49.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.666 min
Delta R.T.: -0.017 min
Response: 122651112
Conc: 45.38 ng/ml



#21 AR-1248-1


R.T.: 5.666 min
Delta R.T.: -0.007 min
Response: 23801376
Conc: 509.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.725 min
Delta R.T.: -0.007 min
Response: 22025409
Conc: 494.49 ng/ml




#22 AR-1248-2

R.T.: 5.943 min
Delta R.T.: -0.007 min
Response: 37466359
Conc: 516.22 ng/ml



#22 AR-1248-2


R.T.: 4.962 min
Delta R.T.: -0.009 min
Response: 33634466
Conc: 473.90 ng/ml



#23 AR-1248-3


R.T.: 6.149 min
Delta R.T.: -0.007 min
Response: 40410150
Conc: 500.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.005 min
Delta R.T.: -0.008 min
Response: 36040578
Conc: 487.03 ng/ml




#24 AR-1248-4

R.T.: 6.557 min
Delta R.T.: -0.007 min
Response: 42830824
Conc: 523.77 ng/ml



#24 AR-1248-4


R.T.: 5.176 min
Delta R.T.: -0.009 min
Response: 42610614
Conc: 492.08 ng/ml



#25 AR-1248-5

R.T.: 6.596 min
Delta R.T.: -0.006 min
Response: 44365124
Conc: 543.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.571 min
Delta R.T.: -0.009 min
Response: 40041984
Conc: 533.34 ng/ml