

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040622\
 Data File : PP045576.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Apr 2022 17:29
 Operator : YP\AJ
 Sample : N2255-01 10X
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 31496

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/07/2022
 Supervised By :mohammad ahmed 04/07/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 07 04:52:04 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 31 07:54:52 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.884	3.145	4435384	3428666	1.838m	2.167m
2) SA Decachlor...	9.994	8.493	2554805	3674054	1.783m	2.423m#

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040622\
Data File : PP045576.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Apr 2022 17:29
Operator : YP\AJ
Sample : N2255-01 10X
Misc :
ALS Vial : 19 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
31496


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/07/2022
Supervised By :mohammad ahmed 04/07/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 04:52:04 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033022.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 31 07:54:52 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.884 min
Delta R.T.: -0.003 min
Response: 4435384
Conc: 1.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
31496


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/07/2022
Supervised By :mohammad ahmed 04/07/2022




#1 Tetrachloro-m-xylene

R.T.: 3.145 min
Delta R.T.: -0.002 min
Response: 3428666
Conc: 2.17 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.994 min
Delta R.T.: -0.011 min
Response: 2554805
Conc: 1.78 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.493 min
Delta R.T.: -0.004 min
Response: 3674054
Conc: 2.42 ng/ml m