

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040622\
 Data File : PP045601.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Apr 2022 07:43
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 07 09:00:49 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 31 07:54:52 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.887	3.146	129.7E6	89460058	53.743	56.536
2) SA Decachlor...	9.995	8.489	73375814	69690886	51.210	45.957
Target Compounds						
21) L5 AR-1248-1	5.136	4.285	22042211	18225402	481.594	478.138
22) L5 AR-1248-2	5.432	4.536	28764493	24603801	467.552	477.657
23) L5 AR-1248-3	5.651	4.579	33863724	25925269	473.524	481.526
24) L5 AR-1248-4	6.090	4.760	36496882	31465909	485.466	489.023
25) L5 AR-1248-5	6.130	5.178	35998390	31511559	479.669	484.408


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040622\
Data File : PP045601.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2022 07:43
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 09:00:49 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033022.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 31 07:54:52 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.887 min
Delta R.T.: 0.000 min
Response: 129670794
Conc: 53.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.146 min
Delta R.T.: 0.000 min
Response: 89460058
Conc: 56.54 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.995 min
Delta R.T.: -0.010 min
Response: 73375814
Conc: 51.21 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.489 min
Delta R.T.: -0.008 min
Response: 69690886
Conc: 45.96 ng/ml



#21 AR-1248-1


R.T.: 5.136 min
Delta R.T.: -0.001 min
Response: 22042211
Conc: 481.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.285 min
Delta R.T.: -0.002 min
Response: 18225402
Conc: 478.14 ng/ml




#22 AR-1248-2

R.T.: 5.432 min
Delta R.T.: -0.002 min
Response: 28764493
Conc: 467.55 ng/ml



#22 AR-1248-2


R.T.: 4.536 min
Delta R.T.: -0.003 min
Response: 24603801
Conc: 477.66 ng/ml



#23 AR-1248-3


R.T.: 5.651 min
Delta R.T.: -0.002 min
Response: 33863724
Conc: 473.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.579 min
Delta R.T.: -0.003 min
Response: 25925269
Conc: 481.53 ng/ml




#24 AR-1248-4

R.T.: 6.090 min
Delta R.T.: -0.002 min
Response: 36496882
Conc: 485.47 ng/ml



#24 AR-1248-4


R.T.: 4.760 min
Delta R.T.: -0.004 min
Response: 31465909
Conc: 489.02 ng/ml



#25 AR-1248-5

R.T.: 6.130 min
Delta R.T.: -0.002 min
Response: 35998390
Conc: 479.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.178 min
Delta R.T.: -0.004 min
Response: 31511559
Conc: 484.41 ng/ml