

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040622\
 Data File : PP045616.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Apr 2022 12:34
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 07 13:03:02 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 31 07:54:52 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.885	3.146	131.9E6	90748707	54.650	57.351
2) SA Decachlor...	9.993	8.488	73125694	70108011	51.036	46.232
Target Compounds						
21) L5 AR-1248-1	5.134	4.284	22643423	18696291	494.730	490.491
22) L5 AR-1248-2	5.430	4.535	29388678	25004294	477.698	485.433
23) L5 AR-1248-3	5.648	4.578	34564732	26252724	483.326	487.608
24) L5 AR-1248-4	6.087	4.759	36464599	32036234	485.036	497.887
25) L5 AR-1248-5	6.128	5.177	35472707	32741997	472.665	503.323



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040622\
Data File : PP045616.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2022 12:34
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 13:03:02 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033022.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 31 07:54:52 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.885 min
Delta R.T.: -0.002 min
Response: 131860322
Conc: 54.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.146 min
Delta R.T.: 0.000 min
Response: 90748707
Conc: 57.35 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.993 min
Delta R.T.: -0.012 min
Response: 73125694
Conc: 51.04 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.488 min
Delta R.T.: -0.009 min
Response: 70108011
Conc: 46.23 ng/ml



#21 AR-1248-1


R.T.: 5.134 min
Delta R.T.: -0.003 min
Response: 22643423
Conc: 494.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.284 min
Delta R.T.: -0.004 min
Response: 18696291
Conc: 490.49 ng/ml




#22 AR-1248-2

R.T.: 5.430 min
Delta R.T.: -0.004 min
Response: 29388678
Conc: 477.70 ng/ml



#22 AR-1248-2


R.T.: 4.535 min
Delta R.T.: -0.004 min
Response: 25004294
Conc: 485.43 ng/ml



#23 AR-1248-3


R.T.: 5.648 min
Delta R.T.: -0.004 min
Response: 34564732
Conc: 483.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.578 min
Delta R.T.: -0.004 min
Response: 26252724
Conc: 487.61 ng/ml




#24 AR-1248-4

R.T.: 6.087 min
Delta R.T.: -0.004 min
Response: 36464599
Conc: 485.04 ng/ml



#24 AR-1248-4


R.T.: 4.759 min
Delta R.T.: -0.005 min
Response: 32036234
Conc: 497.89 ng/ml



#25 AR-1248-5

R.T.: 6.128 min
Delta R.T.: -0.004 min
Response: 35472707
Conc: 472.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.177 min
Delta R.T.: -0.005 min
Response: 32741997
Conc: 503.32 ng/ml