

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040622\
 Data File : PP045661.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Apr 2022 06:52
 Operator : YP\AJ
 Sample : N2292-01 10X
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 2291

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/08/2022
 Supervised By :mohammad ahmed 04/11/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 08 08:55:26 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 31 07:54:52 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.887	3.146	3316531	2191777	1.375	1.385
2) SA Decachlor...	9.998	8.487	2552527	2252739	1.781	1.486m
Target Compounds						
31) L7 AR-1260-1	6.877	5.866	1556013	1495154	19.511	18.037m
32) L7 AR-1260-2	7.163	6.072	4565949	2382979	50.672m	24.459 #
33) L7 AR-1260-3	7.553	6.235	1372073	2536618	19.806m	27.423 #
34) L7 AR-1260-4	7.806	6.743	3565549	2418690	46.218m	31.568 #
35) L7 AR-1260-5	8.142	7.011	3796050	3772829	26.886m	21.328m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040622\
Data File : PP045661.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Apr 2022 06:52
Operator : YP\AJ
Sample : N2292-01 10X
Misc :
ALS Vial : 25 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
2291


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/08/2022
Supervised By :mohammad ahmed 04/11/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 08 08:55:26 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033022.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 31 07:54:52 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.887 min
Delta R.T.: 0.000 min
Response: 3316531
Conc: 1.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
2291


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/08/2022
Supervised By :mohammad ahmed 04/11/2022




#1 Tetrachloro-m-xylene

R.T.: 3.146 min
Delta R.T.: 0.000 min
Response: 2191777
Conc: 1.39 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.998 min
Delta R.T.: -0.007 min
Response: 2552527
Conc: 1.78 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.487 min
Delta R.T.: -0.010 min
Response: 2252739
Conc: 1.49 ng/ml m




#31 AR-1260-1

R.T.: 6.877 min
Delta R.T.: -0.004 min
Response: 1556013
Conc: 19.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
2291


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/08/2022
Supervised By :mohammad ahmed 04/11/2022




#31 AR-1260-1

R.T.: 5.866 min
Delta R.T.: -0.005 min
Response: 1495154
Conc: 18.04 ng/ml m




#32 AR-1260-2

R.T.: 7.163 min
Delta R.T.: -0.002 min
Response: 4565949
Conc: 50.67 ng/ml m



#32 AR-1260-2

R.T.: 6.072 min
Delta R.T.: -0.006 min
Response: 2382979
Conc: 24.46 ng/ml




#33 AR-1260-3

R.T.: 7.553 min
Delta R.T.: -0.006 min
Response: 1372073
Conc: 19.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
2291


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/08/2022
Supervised By :mohammad ahmed 04/11/2022




#33 AR-1260-3

R.T.: 6.235 min
Delta R.T.: -0.004 min
Response: 2536618
Conc: 27.42 ng/ml




#34 AR-1260-4

R.T.: 7.806 min
Delta R.T.: 0.000 min
Response: 3565549
Conc: 46.22 ng/ml m



#34 AR-1260-4

R.T.: 6.743 min
Delta R.T.: -0.008 min
Response: 2418690
Conc: 31.57 ng/ml




#35 AR-1260-5

R.T.: 8.142 min
Delta R.T.: -0.006 min
Response: 3796050
Conc: 26.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
2291

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/08/2022
Supervised By :mohammad ahmed 04/11/2022



#35 AR-1260-5

R.T.: 7.011 min
Delta R.T.: -0.007 min
Response: 3772829
Conc: 21.33 ng/ml m