

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040623\
 Data File : PP056698.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Apr 2023 02:09
 Operator : YP\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP040623AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 07 03:57:31 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP040623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 07 03:55:59 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.434	3.662	106.4E6	86200496	51.178	51.109
2) SA Decachlor...	10.268	8.747	64338638	76725529	50.225	49.503
Target Compounds						
26) L6 AR-1254-1	6.470	5.576	35177104	45721419	496.110	496.337
27) L6 AR-1254-2	6.690	5.725	50353817	39918990	496.539	496.694
28) L6 AR-1254-3	7.059	6.134	48576764	62027894	501.143	501.777
29) L6 AR-1254-4	7.345	6.363	30643046	32240443	502.057	507.267
30) L6 AR-1254-5	7.767	6.786	34222366	51324551	502.218	488.172



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040623\
Data File : PP056698.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2023 02:09
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP040623AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 03:57:31 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP040623.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 07 03:55:59 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.434 min
Delta R.T.: 0.000 min
Response: 106426916
Conc: 51.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP040623AR1254




#1 Tetrachloro-m-xylene

R.T.: 3.662 min
Delta R.T.: 0.000 min
Response: 86200496
Conc: 51.11 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.268 min
Delta R.T.: -0.002 min
Response: 64338638
Conc: 50.23 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.747 min
Delta R.T.: 0.000 min
Response: 76725529
Conc: 49.50 ng/ml



#26 AR-1254-1


R.T.: 6.470 min
Delta R.T.: -0.001 min
Response: 35177104
Conc: 496.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP040623AR1254




#26 AR-1254-1

R.T.: 5.576 min
Delta R.T.: -0.001 min
Response: 45721419
Conc: 496.34 ng/ml




#27 AR-1254-2

R.T.: 6.690 min
Delta R.T.: -0.002 min
Response: 50353817
Conc: 496.54 ng/ml



#27 AR-1254-2


R.T.: 5.725 min
Delta R.T.: -0.001 min
Response: 39918990
Conc: 496.69 ng/ml



#28 AR-1254-3


R.T.: 7.059 min
Delta R.T.: 0.000 min
Response: 48576764
Conc: 501.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP040623AR1254




#28 AR-1254-3

R.T.: 6.134 min
Delta R.T.: 0.000 min
Response: 62027894
Conc: 501.78 ng/ml




#29 AR-1254-4

R.T.: 7.345 min
Delta R.T.: -0.001 min
Response: 30643046
Conc: 502.06 ng/ml



#29 AR-1254-4


R.T.: 6.363 min
Delta R.T.: -0.001 min
Response: 32240443
Conc: 507.27 ng/ml



#30 AR-1254-5

R.T.: 7.767 min
Delta R.T.: 0.000 min
Response: 34222366
Conc: 502.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP040623AR1254



#30 AR-1254-5

R.T.: 6.786 min
Delta R.T.: 0.000 min
Response: 51324551
Conc: 488.17 ng/ml