

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040822\
 Data File : PP045791.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2022 08:17
 Operator : YP\AJ
 Sample : N2334-03
 Misc :
 ALS Vial : 81 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 MW-20

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/11/2022
 Supervised By :mohammad ahmed 04/11/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 11 02:31:10 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Apr 09 07:55:16 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.882	3.144	47422011	32153276	19.654	20.320
2) SA Decachlor...	9.984	8.481	24041043	23695071	16.779m	15.626

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040822\
Data File : PP045791.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2022 08:17
Operator : YP\AJ
Sample : N2334-03
Misc :
ALS Vial : 81 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
MW-20


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/11/2022
Supervised By :mohammad ahmed 04/11/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 11 02:31:10 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033022.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Apr 09 07:55:16 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.882 min
Delta R.T.: -0.002 min
Response: 47422011
Conc: 19.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
MW-20


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/11/2022
Supervised By :mohammad ahmed 04/11/2022




#1 Tetrachloro-m-xylene

R.T.: 3.144 min
Delta R.T.: 0.000 min
Response: 32153276
Conc: 20.32 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.984 min
Delta R.T.: -0.007 min
Response: 24041043
Conc: 16.78 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.481 min
Delta R.T.: -0.004 min
Response: 23695071
Conc: 15.63 ng/ml