

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_P\Data\PP040921\  
 Data File : PP034762.D  
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
 Acq On : 09 Apr 2021 11:21  
 Operator : DD\AJ  
 Sample : M1946-03  
 Misc :  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 ECD\_P  
 ClientSampleId :  
 FS-SB9

Manual Integrations  
 APPROVED

Ankita  
 4/12/2021 2:24:04 PM

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Apr 09 15:34:37 2021  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_P\methods\PP033021.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Wed Mar 31 06:02:19 2021  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound                    | RT#1   | RT#2  | Resp#1  | Resp#2 | ng/ml   | ng/ml  |
|-----------------------------|--------|-------|---------|--------|---------|--------|
| -----                       |        |       |         |        |         |        |
| System Monitoring Compounds |        |       |         |        |         |        |
| 1) SA Tetrachlo...          | 4.844  | 4.258 | 1528196 | 919217 | 20.033m | 20.462 |
| 2) SA Decachlor...          | 10.757 | 9.553 | 1604917 | 646670 | 20.848  | 22.902 |

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_P\Data\PP040921\  
Data File : PP034762.D  
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
Acq On : 09 Apr 2021 11:21  
Operator : DD\AJ  
Sample : M1946-03  
Misc :  
ALS Vial : 9 Sample Multiplier: 1



Instrument :  
ECD\_P  
ClientSampleId :  
FS-SB9


Manual Integrations  
APPROVED

Ankita  
4/12/2021 2:24:04 PM

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Apr 09 15:34:37 2021  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_P\methods\PP033021.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Wed Mar 31 06:02:19 2021  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.844 min  
Delta R.T.: -0.002 min  
Response: 1528196  
Conc: 20.03 ng/ml m

Instrument :  
ECD\_P  
ClientSampleId :  
FS-SB9


Manual Integrations  
APPROVED

Ankita  
4/12/2021 2:24:04 PM




#1 Tetrachloro-m-xylene

R.T.: 4.258 min  
Delta R.T.: -0.002 min  
Response: 919217  
Conc: 20.46 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.757 min  
Delta R.T.: -0.008 min  
Response: 1604917  
Conc: 20.85 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.553 min  
Delta R.T.: -0.013 min  
Response: 646670  
Conc: 22.90 ng/ml