

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040924\
 Data File : PP064219.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Apr 2024 23:30
 Operator : YP\AJ
 Sample : P2026-02
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 50306

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 00:37:49 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 19 06:23:02 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.485	3.636	46332902	47342732	19.417	19.829
2) SA Decachlor...	10.351	8.663	44009675	40696261	17.333	15.059
Target Compounds						
9) L2 AR-1221-2	0.000	3.936	0	663184	N.D.	32.360 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040924\
Data File : PP064219.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Apr 2024 23:30
Operator : YP\AJ
Sample : P2026-02
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
50306

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 00:37:49 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 19 06:23:02 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.485 min
Delta R.T.: -0.004 min
Response: 46332902
Conc: 19.42 ng/ml

Instrument :
ECD_P
ClientSampleId :
50306




#1 Tetrachloro-m-xylene

R.T.: 3.636 min
Delta R.T.: -0.004 min
Response: 47342732
Conc: 19.83 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.351 min
Delta R.T.: -0.014 min
Response: 44009675
Conc: 17.33 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.663 min
Delta R.T.: -0.021 min
Response: 40696261
Conc: 15.06 ng/ml



#9 AR-1221-2

R.T.: 0.000 min
Exp R.T.: 4.789 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
50306



#9 AR-1221-2

R.T.: 3.936 min
Delta R.T.: -0.005 min
Response: 663184
Conc: 32.36 ng/ml