

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040924\
 Data File : PP064184.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Apr 2024 10:22
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/10/2024
 Supervised By :Ankita Jodhani 04/10/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 09 15:56:10 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 19 06:23:02 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.484	3.635	113.5E6	117.3E6	47.560	49.148
2)	SA Decachlor...	10.349	8.664	135.1E6	127.8E6	53.222	47.273
Target Compounds							
26)	L6 AR-1254-1	6.530	5.529	48766107	66835235	536.378	540.857
27)	L6 AR-1254-2	6.750	5.678	76195874	61926590	555.804	539.523
28)	L6 AR-1254-3	7.119	6.082	76147195	94343805	544.120	524.307
29)	L6 AR-1254-4	7.407	6.312	51538610	52196754	565.687	566.435
30)	L6 AR-1254-5	7.827	6.731	63117654	87412263	536.820m	523.675

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040924\
Data File : PP064184.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Apr 2024 10:22
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/10/2024
Supervised By :Ankita Jodhani 04/10/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 09 15:56:10 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 19 06:23:02 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.484 min
Delta R.T.: -0.005 min
Response: 113484762
Conc: 47.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/10/2024
Supervised By :Ankita Jodhani 04/10/2024




#1 Tetrachloro-m-xylene

R.T.: 3.635 min
Delta R.T.: -0.004 min
Response: 117342536
Conc: 49.15 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.349 min
Delta R.T.: -0.016 min
Response: 135131477
Conc: 53.22 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.664 min
Delta R.T.: -0.020 min
Response: 127756303
Conc: 47.27 ng/ml




#26 AR-1254-1

R.T.: 6.530 min
Delta R.T.: -0.006 min
Response: 48766107
Conc: 536.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/10/2024
Supervised By :Ankita Jodhani 04/10/2024




#26 AR-1254-1

R.T.: 5.529 min
Delta R.T.: -0.009 min
Response: 66835235
Conc: 540.86 ng/ml




#27 AR-1254-2

R.T.: 6.750 min
Delta R.T.: -0.008 min
Response: 76195874
Conc: 555.80 ng/ml



#27 AR-1254-2

R.T.: 5.678 min
Delta R.T.: -0.010 min
Response: 61926590
Conc: 539.52 ng/ml




#28 AR-1254-3

R.T.: 7.119 min
Delta R.T.: -0.007 min
Response: 76147195
Conc: 544.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/10/2024
Supervised By :Ankita Jodhani 04/10/2024




#28 AR-1254-3

R.T.: 6.082 min
Delta R.T.: -0.011 min
Response: 94343805
Conc: 524.31 ng/ml




#29 AR-1254-4

R.T.: 7.407 min
Delta R.T.: -0.007 min
Response: 51538610
Conc: 565.69 ng/ml



#29 AR-1254-4

R.T.: 6.312 min
Delta R.T.: -0.011 min
Response: 52196754
Conc: 566.43 ng/ml




#30 AR-1254-5

R.T.: 7.827 min
Delta R.T.: -0.009 min
Response: 63117654
Conc: 536.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/10/2024
Supervised By :Ankita Jodhani 04/10/2024



#30 AR-1254-5

R.T.: 6.731 min
Delta R.T.: -0.011 min
Response: 87412263
Conc: 523.67 ng/ml