

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041023\
 Data File : PP056819.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2023 15:17
 Operator : YP\AJ
 Sample : 02257-01
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ELI-SOLID-0407

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/11/2023
 Supervised By :Ankita Jodhani 04/11/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 23:10:45 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP040623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 07 04:14:17 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.435	3.662	33655584	26228661	15.665	15.081
2) SA Decachlor...	10.268	8.744	21978928	23775368	16.792	15.121
Target Compounds						
31) L7 AR-1260-1	7.225	6.263	2740278	2592893	33.664	27.110m
32) L7 AR-1260-2	7.482	6.453	3674968	2785089	43.074	26.967 #
33) L7 AR-1260-3	7.843	6.608	2404775	2858720	34.435	27.566
34) L7 AR-1260-4	8.066	7.085	4249107	1661750	55.690	19.456 #
35) L7 AR-1260-5	8.396	7.327	3668582	3018552	28.984	19.073 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041023\
Data File : PP056819.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2023 15:17
Operator : YP\AJ
Sample : 02257-01
Misc :
ALS Vial : 18 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
ELI-SOLID-0407


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/11/2023
Supervised By :Ankita Jodhani 04/11/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 23:10:45 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP040623.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 07 04:14:17 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.435 min
Delta R.T.: -0.002 min
Response: 33655584
Conc: 15.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
ELI-SOLID-0407


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/11/2023
Supervised By :Ankita Jodhani 04/11/2023




#1 Tetrachloro-m-xylene

R.T.: 3.662 min
Delta R.T.: -0.003 min
Response: 26228661
Conc: 15.08 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.268 min
Delta R.T.: -0.007 min
Response: 21978928
Conc: 16.79 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.744 min
Delta R.T.: -0.007 min
Response: 23775368
Conc: 15.12 ng/ml




#31 AR-1260-1

R.T.: 7.225 min
Delta R.T.: -0.003 min
Response: 2740278
Conc: 33.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
ELI-SOLID-0407


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/11/2023
Supervised By :Ankita Jodhani 04/11/2023




#31 AR-1260-1

R.T.: 6.263 min
Delta R.T.: -0.006 min
Response: 2592893
Conc: 27.11 ng/ml m




#32 AR-1260-2

R.T.: 7.482 min
Delta R.T.: -0.004 min
Response: 3674968
Conc: 43.07 ng/ml



#32 AR-1260-2

R.T.: 6.453 min
Delta R.T.: -0.005 min
Response: 2785089
Conc: 26.97 ng/ml




#33 AR-1260-3

R.T.: 7.843 min
Delta R.T.: -0.004 min
Response: 2404775
Conc: 34.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
ELI-SOLID-0407


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/11/2023
Supervised By :Ankita Jodhani 04/11/2023




#33 AR-1260-3

R.T.: 6.608 min
Delta R.T.: -0.006 min
Response: 2858720
Conc: 27.57 ng/ml




#34 AR-1260-4

R.T.: 8.066 min
Delta R.T.: -0.009 min
Response: 4249107
Conc: 55.69 ng/ml



#34 AR-1260-4

R.T.: 7.085 min
Delta R.T.: -0.005 min
Response: 1661750
Conc: 19.46 ng/ml




#35 AR-1260-5

R.T.: 8.396 min
Delta R.T.: -0.005 min
Response: 3668582
Conc: 28.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
ELI-SOLID-0407

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/11/2023
Supervised By :Ankita Jodhani 04/11/2023



#35 AR-1260-5

R.T.: 7.327 min
Delta R.T.: -0.005 min
Response: 3018552
Conc: 19.07 ng/ml