

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041122\
 Data File : PP045827.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Apr 2022 10:38
 Operator : YP\AJ
 Sample : N2350-02 10X
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HZA556D-1-2

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/11/2022
 Supervised By :mohammad ahmed 04/12/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 11 14:42:47 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Apr 11 14:34:35 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.879	3.142	4110309	2759603	1.704m	1.744m
2) SA Decachlor...	9.981	8.480	3368725	2679056	2.351	1.767m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041122\
Data File : PP045827.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Apr 2022 10:38
Operator : YP\AJ
Sample : N2350-02 10X
Misc :
ALS Vial : 18 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
HZA556D-1-2


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/11/2022
Supervised By :mohammad ahmed 04/12/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 11 14:42:47 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033022.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Apr 11 14:34:35 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.879 min
Delta R.T.: -0.008 min
Response: 4110309
Conc: 1.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
HZA556D-1-2


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/11/2022
Supervised By :mohammad ahmed 04/12/2022




#1 Tetrachloro-m-xylene

R.T.: 3.142 min
Delta R.T.: -0.005 min
Response: 2759603
Conc: 1.74 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.981 min
Delta R.T.: -0.009 min
Response: 3368725
Conc: 2.35 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.480 min
Delta R.T.: -0.010 min
Response: 2679056
Conc: 1.77 ng/ml m