

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041123\
 Data File : PP056840.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Apr 2023 10:47
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 11 11:39:48 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP040623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 07 04:14:17 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.435	3.661	109.8E6	81034768	51.102	46.593
2) SA Decachlor...	10.274	8.745	68445061	82467720	52.292	52.450
Target Compounds						
21) L5 AR-1248-1	5.610	4.762	19473930	16217659	519.039	492.675
22) L5 AR-1248-2	5.887	5.002	30170029	25056579	512.274	486.134
23) L5 AR-1248-3	6.093	5.044	32758901	26244831	514.172	488.899
24) L5 AR-1248-4	6.498	5.218	30060501	30578625	517.618	494.199
25) L5 AR-1248-5	6.537	5.616	29288191	25702955	518.498	500.604



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041123\
Data File : PP056840.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Apr 2023 10:47
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 11 11:39:48 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP040623.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 07 04:14:17 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.435 min
Delta R.T.: -0.002 min
Response: 109793358
Conc: 51.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.661 min
Delta R.T.: -0.004 min
Response: 81034768
Conc: 46.59 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.274 min
Delta R.T.: 0.000 min
Response: 68445061
Conc: 52.29 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.745 min
Delta R.T.: -0.006 min
Response: 82467720
Conc: 52.45 ng/ml



#21 AR-1248-1


R.T.: 5.610 min
Delta R.T.: 0.000 min
Response: 19473930
Conc: 519.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.762 min
Delta R.T.: -0.002 min
Response: 16217659
Conc: 492.68 ng/ml




#22 AR-1248-2

R.T.: 5.887 min
Delta R.T.: 0.000 min
Response: 30170029
Conc: 512.27 ng/ml



#22 AR-1248-2


R.T.: 5.002 min
Delta R.T.: -0.002 min
Response: 25056579
Conc: 486.13 ng/ml



#23 AR-1248-3


R.T.: 6.093 min
Delta R.T.: 0.000 min
Response: 32758901
Conc: 514.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.044 min
Delta R.T.: -0.003 min
Response: 26244831
Conc: 488.90 ng/ml




#24 AR-1248-4

R.T.: 6.498 min
Delta R.T.: 0.000 min
Response: 30060501
Conc: 517.62 ng/ml



#24 AR-1248-4


R.T.: 5.218 min
Delta R.T.: -0.002 min
Response: 30578625
Conc: 494.20 ng/ml



#25 AR-1248-5

R.T.: 6.537 min
Delta R.T.: 0.000 min
Response: 29288191
Conc: 518.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.616 min
Delta R.T.: -0.003 min
Response: 25702955
Conc: 500.60 ng/ml