Quantitation Report (Not Reviewed)

Data Path : P:\HPCHEM1\ECD P\Data\PP041218\
Data File : PP025050.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Apr 2018 2:05
Operator : UA/AJ
Sample : J2337-17
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument : ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 13 06:56:11 2018
Quant Method : P:\HPCHEM1\ECD_P\methods\PP032918.M
Quant Title : 8080.M
QLast Update : Mon Apr 02 02:18:04 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

		Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml	
		System Monitoring Compounds							
4)	S	2,4-DCAA	11.479	11.171	149.1E6	271.8E6	281.687	458.100	#
		Target Compounds							
5)	T	DICAMBA	0.000	11.400	0	23394890	N.D.	10.129	#
6)	T	MCPP	12.158	0.000	16743160	0	10.735	N.D.	#
8)	T	DICHLORPROP	0.000	12.521	0	34148473	N.D.	57.985	#
12)	T	2,4,5-T	0.000	14.849	0	10325552	N.D.	3.461	#
14)	T	DINOSEB	0.000	15.900	0	83939648	N.D.	34.329	#
15)	T	Picloram	16.991	0.000	16171312	0	2.763	N.D.	#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (Not Reviewed)

Data Path : P:\HPCHEM1\ECD P\Data\PP041218\
Data File : PP025050.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Apr 2018 2:05
Operator : UA/AJ
Sample : J2337-17
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument : ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 13 06:56:11 2018
Quant Method : P:\HPCHEM1\ECD_P\methods\PP032918.M
Quant Title : 8080.M
QLast Update : Mon Apr 02 02:18:04 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm







#4 2,4-DCAA

R.T.: 11.479 min
Delta R.T.: 0.033 min
Response: 149089464
Conc: 281.69 ng/ml

Instrument :
ECD_P
ClientSampleId :



#4 2,4-DCAA

R.T.: 11.171 min
Delta R.T.: 0.004 min
Response: 271842578
Conc: 458.10 ng/ml



#5 DICAMBA

R.T.: 0.000 min
Exp R.T. : 11.765 min
Response: 0
Conc: N.D.



#5 DICAMBA

R.T.: 11.400 min
Delta R.T.: -0.062 min
Response: 23394890
Conc: 10.13 ng/ml



#6 MCPP

R.T.: 12.158 min
Delta R.T.: 0.053 min
Response: 16743160
Conc: 10.73 ug/ml

Instrument :
ECD_P
ClientSampleId :



#6 MCPP

R.T.: 0.000 min
Exp R.T. : 11.669 min
Response: 0
Conc: N.D.



#8 DICHLORPROP

R.T.: 0.000 min
Exp R.T. : 12.959 min
Response: 0
Conc: N.D.



#8 DICHLORPROP

R.T.: 12.521 min
Delta R.T.: -0.052 min
Response: 34148473
Conc: 57.98 ng/ml



#12 2,4,5-T

R.T.: 0.000 min
Exp R.T. : 14.979 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :



#12 2,4,5-T

R.T.: 14.849 min
Delta R.T.: 0.013 min
Response: 10325552
Conc: 3.46 ng/ml



#14 DINOSEB

R.T.: 0.000 min
Exp R.T. : 17.198 min
Response: 0
Conc: N.D.



#14 DINOSEB

R.T.: 15.900 min
Delta R.T.: 0.069 min
Response: 83939648
Conc: 34.33 ng/ml



#15 Picloram

R.T.: 16.991 min
Delta R.T.: -0.014 min
Response: 16171312
Conc: 2.76 ng/ml

Instrument :
ECD_P
ClientSampleId :



#15 Picloram

R.T.: 0.000 min
Exp R.T. : 17.167 min
Response: 0
Conc: N.D.