

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041221\
 Data File : PP034824.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 Apr 2021 18:47
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 13 02:07:41 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 31 06:02:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.844	4.257	3906601	2401865	51.211	53.467
2) SA Decachlor...	10.759	9.553	3413013	1352018	44.336	47.883
Target Compounds						
3) L1 AR-1016-1	6.155	5.506	1093349	580984	464.705	489.734
4) L1 AR-1016-2	6.178	5.527	1696141	880090	477.556	495.649
5) L1 AR-1016-3	6.244	5.720	1088549	471294	509.788	504.394
6) L1 AR-1016-4	6.350	5.767	877366	396911	492.671	546.491
7) L1 AR-1016-5	6.663	5.998	877956	487548	508.861	533.039
31) L7 AR-1260-1	7.834	7.085	1472074	792797	474.207	516.258
32) L7 AR-1260-2	8.100	7.279	1672627	930871	445.390	515.320
33) L7 AR-1260-3	8.464	7.435	1371049	862520	453.581	497.164
34) L7 AR-1260-4	8.694	7.913	1583793	712719	453.649	497.097
35) L7 AR-1260-5	9.013	8.157	3067411	1555350	428.314	455.320



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041221\
Data File : PP034824.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 12 Apr 2021 18:47
Operator : DD\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 13 02:07:41 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033021.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 31 06:02:19 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.844 min
Delta R.T.: -0.002 min
Response: 3906601
Conc: 51.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.257 min
Delta R.T.: -0.003 min
Response: 2401865
Conc: 53.47 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.759 min
Delta R.T.: -0.006 min
Response: 3413013
Conc: 44.34 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.553 min
Delta R.T.: -0.013 min
Response: 1352018
Conc: 47.88 ng/ml



#3 AR-1016-1


R.T.: 6.155 min
Delta R.T.: -0.003 min
Response: 1093349
Conc: 464.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 5.506 min
Delta R.T.: -0.006 min
Response: 580984
Conc: 489.73 ng/ml




#4 AR-1016-2

R.T.: 6.178 min
Delta R.T.: -0.003 min
Response: 1696141
Conc: 477.56 ng/ml



#4 AR-1016-2


R.T.: 5.527 min
Delta R.T.: -0.006 min
Response: 880090
Conc: 495.65 ng/ml



#5 AR-1016-3


R.T.: 6.244 min
Delta R.T.: -0.002 min
Response: 1088549
Conc: 509.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.720 min
Delta R.T.: -0.005 min
Response: 471294
Conc: 504.39 ng/ml




#6 AR-1016-4

R.T.: 6.350 min
Delta R.T.: -0.003 min
Response: 877366
Conc: 492.67 ng/ml



#6 AR-1016-4


R.T.: 5.767 min
Delta R.T.: -0.006 min
Response: 396911
Conc: 546.49 ng/ml



#7 AR-1016-5


R.T.: 6.663 min
Delta R.T.: -0.004 min
Response: 877956
Conc: 508.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.998 min
Delta R.T.: -0.006 min
Response: 487548
Conc: 533.04 ng/ml




#31 AR-1260-1

R.T.: 7.834 min
Delta R.T.: -0.003 min
Response: 1472074
Conc: 474.21 ng/ml



#31 AR-1260-1


R.T.: 7.085 min
Delta R.T.: -0.008 min
Response: 792797
Conc: 516.26 ng/ml



#32 AR-1260-2


R.T.: 8.100 min
Delta R.T.: -0.003 min
Response: 1672627
Conc: 445.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 7.279 min
Delta R.T.: -0.008 min
Response: 930871
Conc: 515.32 ng/ml




#33 AR-1260-3

R.T.: 8.464 min
Delta R.T.: -0.003 min
Response: 1371049
Conc: 453.58 ng/ml



#33 AR-1260-3


R.T.: 7.435 min
Delta R.T.: -0.008 min
Response: 862520
Conc: 497.16 ng/ml



#34 AR-1260-4


R.T.: 8.694 min
Delta R.T.: -0.003 min
Response: 1583793
Conc: 453.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.913 min
Delta R.T.: -0.008 min
Response: 712719
Conc: 497.10 ng/ml



#35 AR-1260-5

R.T.: 9.013 min
Delta R.T.: -0.003 min
Response: 3067411
Conc: 428.31 ng/ml



#35 AR-1260-5

R.T.: 8.157 min
Delta R.T.: -0.008 min
Response: 1555350
Conc: 455.32 ng/ml