

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041222\
 Data File : PP045844.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Apr 2022 21:10
 Operator : YP\AJ
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 12 06:50:29 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 12 05:31:12 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.877	3.138	129.5E6	79281574	52.669	49.895
2)	SA Decachlor...	9.983	8.478	92832472	72449811	56.551	50.831
Target Compounds							
3)	L1 AR-1016-1	5.126	4.276	2675548	2042261	34.897	34.776
4)	L1 AR-1016-2	5.149	4.294	4400525	3142953	38.055	37.620
5)	L1 AR-1016-3	5.215	4.478	3089704	1369580	44.844	31.004 #
6)	L1 AR-1016-4	5.321	4.527	2143335	468725	36.424	13.135 #
7)	L1 AR-1016-5	5.640	4.751	501280	629582	9.090	13.253 #
8)	L2 AR-1221-1	4.102	3.363	13025735	10157426	500.000	500.000
9)	L2 AR-1221-2	4.194	3.453	9508790	7599663	500.000	500.000
10)	L2 AR-1221-3	4.274	3.532	30734908	23075191	500.000	500.000
11)	L3 AR-1232-1	4.274	3.532	30734908	23075191	541.559	543.884
12)	L3 AR-1232-2	4.836	4.294	2481331	3142953	93.049	78.418
13)	L3 AR-1232-3	5.149	4.478	4400525	1369580	81.751	63.738
14)	L3 AR-1232-4	5.321	4.570	2143335	524275	81.023	27.672 #
15)	L3 AR-1232-5	5.424	4.751	834561	629582	45.275	29.442 #
16)	L4 AR-1242-1	5.126	4.276	2675548	2042261	44.768	44.140
17)	L4 AR-1242-2	5.149	4.294	4400525	3142953	49.021	48.468
18)	L4 AR-1242-3	5.215	4.478	3089704	1369580	55.697	39.580 #
19)	L4 AR-1242-4	5.321	4.570	2143335	524275	47.230	15.597 #
20)	L4 AR-1242-5	6.123	5.128	2164271	718951	46.596	16.370 #
21)	L5 AR-1248-1	5.126	4.276	2675548	2042261	62.022	59.553
22)	L5 AR-1248-2	5.424	4.527	834561	468725	13.960	10.005 #
23)	L5 AR-1248-3	5.640	4.570	501280	524275	7.166	10.726 #
24)	L5 AR-1248-4	6.077	4.751	1482318	629582	18.748	10.836 #
25)	L5 AR-1248-5	6.123	5.168	2164271	788768	27.643	13.438 #
26)	L6 AR-1254-1	6.047	5.128	1049089	718951	12.693	8.151 #
27)	L6 AR-1254-2	6.288	5.288	504919	170097	4.100	2.254 #
28)	L6 AR-1254-3	6.697	5.719	365400	347656	3.040	2.956
29)	L6 AR-1254-4	7.009	5.969	132167	77332	1.499	1.025 #
30)	L6 AR-1254-5	7.460	6.404	617159	65617	6.588	0.640 #
31)	L7 AR-1260-1	6.864	5.859	175823	86997	1.992	1.099 #
32)	L7 AR-1260-2	7.152	6.071	89343	19683	0.872	0.210 #
33)	L7 AR-1260-3	7.532	6.223	149604	299109	1.816	3.377 #
34)	L7 AR-1260-4	7.790	6.735	199354	42714	2.353	0.595 #
35)	L7 AR-1260-5	8.136	6.997	36993	150604	0.225	0.887 #
36)	L8 AR-1262-1	7.532	6.454	149604	56353	1.300	0.533 #
37)	L8 AR-1262-2	8.136	6.735	36993	42714	0.207	0.454 #
38)	L8 AR-1262-3	8.460	7.314	173218	130817	1.329	1.709 #
39)	L8 AR-1262-4	8.544	7.378	206659	34720	3.281	0.244 #
40)	L8 AR-1262-5	9.220	7.927	149586	186188	2.191	2.658
41)	L9 AR-1268-1	8.460	7.314	173218	130817	0.637	0.590

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041222\
 Data File : PP045844.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Apr 2022 21:10
 Operator : YP\AJ
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 12 06:50:29 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 12 05:31:12 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.559	7.378	257291	34720	1.017	0.171 #
43) L9	AR-1268-3	8.787	7.602	203036	288494	0.893	1.694 #
44) L9	AR-1268-4	9.235	7.927	225335	186188	2.497	2.359
45) L9	AR-1268-5	9.646	8.219	429379	269003	0.632	0.485



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041222\
Data File : PP045844.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Apr 2022 21:10
Operator : YP\AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 12 06:50:29 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041222.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 12 05:31:12 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.877 min
Delta R.T.: 0.000 min
Response: 129548983
Conc: 52.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.138 min
Delta R.T.: 0.000 min
Response: 79281574
Conc: 49.89 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.983 min
Delta R.T.: 0.005 min
Response: 92832472
Conc: 56.55 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.478 min
Delta R.T.: 0.000 min
Response: 72449811
Conc: 50.83 ng/ml



#3 AR-1016-1


R.T.: 5.126 min
Delta R.T.: 0.002 min
Response: 2675548
Conc: 34.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#3 AR-1016-1

R.T.: 4.276 min
Delta R.T.: 0.000 min
Response: 2042261
Conc: 34.78 ng/ml




#4 AR-1016-2

R.T.: 5.149 min
Delta R.T.: 0.001 min
Response: 4400525
Conc: 38.05 ng/ml



#4 AR-1016-2


R.T.: 4.294 min
Delta R.T.: 0.000 min
Response: 3142953
Conc: 37.62 ng/ml



#5 AR-1016-3


R.T.: 5.215 min
Delta R.T.: 0.001 min
Response: 3089704
Conc: 44.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#5 AR-1016-3

R.T.: 4.478 min
Delta R.T.: 0.000 min
Response: 1369580
Conc: 31.00 ng/ml




#6 AR-1016-4

R.T.: 5.321 min
Delta R.T.: 0.002 min
Response: 2143335
Conc: 36.42 ng/ml



#6 AR-1016-4


R.T.: 4.527 min
Delta R.T.: 0.000 min
Response: 468725
Conc: 13.14 ng/ml



#7 AR-1016-5


R.T.: 5.640 min
Delta R.T.: 0.002 min
Response: 501280
Conc: 9.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#7 AR-1016-5

R.T.: 4.751 min
Delta R.T.: 0.000 min
Response: 629582
Conc: 13.25 ng/ml




#8 AR-1221-1

R.T.: 4.102 min
Delta R.T.: 0.000 min
Response: 13025735
Conc: 500.00 ng/ml



#8 AR-1221-1


R.T.: 3.363 min
Delta R.T.: 0.000 min
Response: 10157426
Conc: 500.00 ng/ml



#9 AR-1221-2


R.T.: 4.194 min
Delta R.T.: 0.000 min
Response: 9508790
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#9 AR-1221-2

R.T.: 3.453 min
Delta R.T.: 0.000 min
Response: 7599663
Conc: 500.00 ng/ml




#10 AR-1221-3

R.T.: 4.274 min
Delta R.T.: 0.000 min
Response: 30734908
Conc: 500.00 ng/ml



#10 AR-1221-3


R.T.: 3.532 min
Delta R.T.: 0.000 min
Response: 23075191
Conc: 500.00 ng/ml



#11 AR-1232-1


R.T.: 4.274 min
Delta R.T.: 0.001 min
Response: 30734908
Conc: 541.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#11 AR-1232-1

R.T.: 3.532 min
Delta R.T.: 0.000 min
Response: 23075191
Conc: 543.88 ng/ml




#12 AR-1232-2

R.T.: 4.836 min
Delta R.T.: 0.001 min
Response: 2481331
Conc: 93.05 ng/ml



#12 AR-1232-2


R.T.: 4.294 min
Delta R.T.: 0.000 min
Response: 3142953
Conc: 78.42 ng/ml



#13 AR-1232-3


R.T.: 5.149 min
Delta R.T.: 0.002 min
Response: 4400525
Conc: 81.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#13 AR-1232-3

R.T.: 4.478 min
Delta R.T.: 0.000 min
Response: 1369580
Conc: 63.74 ng/ml




#14 AR-1232-4

R.T.: 5.321 min
Delta R.T.: 0.003 min
Response: 2143335
Conc: 81.02 ng/ml



#14 AR-1232-4


R.T.: 4.570 min
Delta R.T.: 0.000 min
Response: 524275
Conc: 27.67 ng/ml



#15 AR-1232-5


R.T.: 5.424 min
Delta R.T.: 0.004 min
Response: 834561
Conc: 45.28 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#15 AR-1232-5

R.T.: 4.751 min
Delta R.T.: 0.000 min
Response: 629582
Conc: 29.44 ng/ml




#16 AR-1242-1

R.T.: 5.126 min
Delta R.T.: 0.000 min
Response: 2675548
Conc: 44.77 ng/ml



#16 AR-1242-1


R.T.: 4.276 min
Delta R.T.: 0.000 min
Response: 2042261
Conc: 44.14 ng/ml



#17 AR-1242-2


R.T.: 5.149 min
Delta R.T.: 0.000 min
Response: 4400525
Conc: 49.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#17 AR-1242-2

R.T.: 4.294 min
Delta R.T.: 0.000 min
Response: 3142953
Conc: 48.47 ng/ml




#18 AR-1242-3

R.T.: 5.215 min
Delta R.T.: 0.000 min
Response: 3089704
Conc: 55.70 ng/ml



#18 AR-1242-3


R.T.: 4.478 min
Delta R.T.: 0.000 min
Response: 1369580
Conc: 39.58 ng/ml



#19 AR-1242-4


R.T.: 5.321 min
Delta R.T.: 0.001 min
Response: 2143335
Conc: 47.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#19 AR-1242-4

R.T.: 4.570 min
Delta R.T.: 0.000 min
Response: 524275
Conc: 15.60 ng/ml




#20 AR-1242-5

R.T.: 6.123 min
Delta R.T.: 0.004 min
Response: 2164271
Conc: 46.60 ng/ml



#20 AR-1242-5


R.T.: 5.128 min
Delta R.T.: 0.001 min
Response: 718951
Conc: 16.37 ng/ml



#21 AR-1248-1


R.T.: 5.126 min
Delta R.T.: 0.002 min
Response: 2675548
Conc: 62.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#21 AR-1248-1

R.T.: 4.276 min
Delta R.T.: 0.000 min
Response: 2042261
Conc: 59.55 ng/ml




#22 AR-1248-2

R.T.: 5.424 min
Delta R.T.: 0.004 min
Response: 834561
Conc: 13.96 ng/ml



#22 AR-1248-2


R.T.: 4.527 min
Delta R.T.: 0.000 min
Response: 468725
Conc: 10.01 ng/ml



#23 AR-1248-3


R.T.: 5.640 min
Delta R.T.: 0.002 min
Response: 501280
Conc: 7.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#23 AR-1248-3

R.T.: 4.570 min
Delta R.T.: 0.000 min
Response: 524275
Conc: 10.73 ng/ml




#24 AR-1248-4

R.T.: 6.077 min
Delta R.T.: 0.000 min
Response: 1482318
Conc: 18.75 ng/ml



#24 AR-1248-4


R.T.: 4.751 min
Delta R.T.: 0.000 min
Response: 629582
Conc: 10.84 ng/ml



#25 AR-1248-5


R.T.: 6.123 min
Delta R.T.: 0.005 min
Response: 2164271
Conc: 27.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#25 AR-1248-5

R.T.: 5.168 min
Delta R.T.: 0.000 min
Response: 788768
Conc: 13.44 ng/ml




#26 AR-1254-1

R.T.: 6.047 min
Delta R.T.: -0.002 min
Response: 1049089
Conc: 12.69 ng/ml



#26 AR-1254-1


R.T.: 5.128 min
Delta R.T.: 0.003 min
Response: 718951
Conc: 8.15 ng/ml



#27 AR-1254-2


R.T.: 6.288 min
Delta R.T.: 0.000 min
Response: 504919
Conc: 4.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#27 AR-1254-2

R.T.: 5.288 min
Delta R.T.: 0.003 min
Response: 170097
Conc: 2.25 ng/ml




#28 AR-1254-3

R.T.: 6.697 min
Delta R.T.: 0.008 min
Response: 365400
Conc: 3.04 ng/ml



#28 AR-1254-3


R.T.: 5.719 min
Delta R.T.: 0.003 min
Response: 347656
Conc: 2.96 ng/ml



#29 AR-1254-4


R.T.: 7.009 min
Delta R.T.: 0.007 min
Response: 132167
Conc: 1.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#29 AR-1254-4

R.T.: 5.969 min
Delta R.T.: 0.004 min
Response: 77332
Conc: 1.02 ng/ml




#30 AR-1254-5

R.T.: 7.460 min
Delta R.T.: -0.002 min
Response: 617159
Conc: 6.59 ng/ml



#30 AR-1254-5


R.T.: 6.404 min
Delta R.T.: -0.011 min
Response: 65617
Conc: 0.64 ng/ml



#31 AR-1260-1


R.T.: 6.864 min
Delta R.T.: 0.000 min
Response: 175823
Conc: 1.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#31 AR-1260-1

R.T.: 5.859 min
Delta R.T.: 0.003 min
Response: 86997
Conc: 1.10 ng/ml




#32 AR-1260-2

R.T.: 7.152 min
Delta R.T.: 0.003 min
Response: 89343
Conc: 0.87 ng/ml



#32 AR-1260-2


R.T.: 6.071 min
Delta R.T.: 0.008 min
Response: 19683
Conc: 0.21 ng/ml



#33 AR-1260-3


R.T.: 7.532 min
Delta R.T.: -0.012 min
Response: 149604
Conc: 1.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#33 AR-1260-3

R.T.: 6.223 min
Delta R.T.: 0.000 min
Response: 299109
Conc: 3.38 ng/ml




#34 AR-1260-4

R.T.: 7.790 min
Delta R.T.: 0.000 min
Response: 199354
Conc: 2.35 ng/ml



#34 AR-1260-4


R.T.: 6.735 min
Delta R.T.: 0.002 min
Response: 42714
Conc: 0.59 ng/ml



#35 AR-1260-5


R.T.: 8.136 min
Delta R.T.: 0.004 min
Response: 36993
Conc: 0.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#35 AR-1260-5

R.T.: 6.997 min
Delta R.T.: -0.006 min
Response: 150604
Conc: 0.89 ng/ml




#36 AR-1262-1

R.T.: 7.532 min
Delta R.T.: -0.010 min
Response: 149604
Conc: 1.30 ng/ml



#36 AR-1262-1


R.T.: 6.454 min
Delta R.T.: 0.000 min
Response: 56353
Conc: 0.53 ng/ml



#37 AR-1262-2


R.T.: 8.136 min
Delta R.T.: 0.007 min
Response: 36993
Conc: 0.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#37 AR-1262-2

R.T.: 6.735 min
Delta R.T.: 0.001 min
Response: 42714
Conc: 0.45 ng/ml




#38 AR-1262-3

R.T.: 8.460 min
Delta R.T.: 0.004 min
Response: 173218
Conc: 1.33 ng/ml



#38 AR-1262-3


R.T.: 7.314 min
Delta R.T.: 0.007 min
Response: 130817
Conc: 1.71 ng/ml



#39 AR-1262-4


R.T.: 8.544 min
Delta R.T.: -0.007 min
Response: 206659
Conc: 3.28 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#39 AR-1262-4

R.T.: 7.378 min
Delta R.T.: 0.000 min
Response: 34720
Conc: 0.24 ng/ml




#40 AR-1262-5

R.T.: 9.220 min
Delta R.T.: -0.004 min
Response: 149586
Conc: 2.19 ng/ml



#40 AR-1262-5


R.T.: 7.927 min
Delta R.T.: 0.007 min
Response: 186188
Conc: 2.66 ng/ml



#41 AR-1268-1


R.T.: 8.460 min
Delta R.T.: 0.001 min
Response: 173218
Conc: 0.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#41 AR-1268-1

R.T.: 7.314 min
Delta R.T.: 0.006 min
Response: 130817
Conc: 0.59 ng/ml




#42 AR-1268-2

R.T.: 8.559 min
Delta R.T.: 0.001 min
Response: 257291
Conc: 1.02 ng/ml



#42 AR-1268-2


R.T.: 7.378 min
Delta R.T.: 0.000 min
Response: 34720
Conc: 0.17 ng/ml



#43 AR-1268-3


R.T.: 8.787 min
Delta R.T.: -0.002 min
Response: 203036
Conc: 0.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#43 AR-1268-3

R.T.: 7.602 min
Delta R.T.: 0.005 min
Response: 288494
Conc: 1.69 ng/ml




#44 AR-1268-4

R.T.: 9.235 min
Delta R.T.: 0.005 min
Response: 225335
Conc: 2.50 ng/ml



#44 AR-1268-4


R.T.: 7.927 min
Delta R.T.: 0.007 min
Response: 186188
Conc: 2.36 ng/ml



#45 AR-1268-5

R.T.: 9.646 min
Delta R.T.: 0.000 min
Response: 429379
Conc: 0.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500



#45 AR-1268-5

R.T.: 8.219 min
Delta R.T.: -0.002 min
Response: 269003
Conc: 0.49 ng/ml