

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041223\
 Data File : PP056896.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Apr 2023 20:11
 Operator : YP\AJ
 Sample : 02291-02 10X
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 2344

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/13/2023
 Supervised By :Ankita Jodhani 04/13/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 12 22:23:45 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP040623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 07 04:14:17 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.436	3.660	2683176	3686706	1.249	2.120m#
2) SA Decachlor...	10.273	8.750	1894855	13127963	1.448m	8.350m#

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041223\
Data File : PP056896.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Apr 2023 20:11
Operator : YP\AJ
Sample : 02291-02 10X
Misc :
ALS Vial : 22 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
2344


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/13/2023
Supervised By :Ankita Jodhani 04/13/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 12 22:23:45 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP040623.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 07 04:14:17 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.436 min
Delta R.T.: -0.001 min
Response: 2683176
Conc: 1.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
2344


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/13/2023
Supervised By :Ankita Jodhani 04/13/2023




#1 Tetrachloro-m-xylene

R.T.: 3.660 min
Delta R.T.: -0.005 min
Response: 3686706
Conc: 2.12 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.273 min
Delta R.T.: -0.002 min
Response: 1894855
Conc: 1.45 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.750 min
Delta R.T.: 0.000 min
Response: 13127963
Conc: 8.35 ng/ml