

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041321\
 Data File : PP034841.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Apr 2021 10:28
 Operator : DD\AJ
 Sample : M1997-04
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 801-675

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 13 15:39:18 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 31 06:02:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.841	4.255	1383225	843471	18.132	18.776
2) SA Decachlor...	10.758	9.550	1686884	427779	21.913	15.150 #

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041321\
Data File : PP034841.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Apr 2021 10:28
Operator : DD\AJ
Sample : M1997-04
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
801-675

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 13 15:39:18 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033021.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 31 06:02:19 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.841 min
Delta R.T.: -0.005 min
Response: 1383225
Conc: 18.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
801-675




#1 Tetrachloro-m-xylene

R.T.: 4.255 min
Delta R.T.: -0.005 min
Response: 843471
Conc: 18.78 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.758 min
Delta R.T.: -0.008 min
Response: 1686884
Conc: 21.91 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.550 min
Delta R.T.: -0.010 min
Response: 427779
Conc: 15.15 ng/ml