

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041322\
 Data File : PP045875.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Apr 2022 11:17
 Operator : YP\AJ
 Sample : PB144019BL
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 13 02:43:04 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 12 05:31:12 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.875	3.137	49627918	33296964	20.176	20.955
2)	SA Decachlor...	9.983	8.475	38796632	31527199	23.634	22.119
Target Compounds							
3)	L1 AR-1016-1	5.121	4.280	127276	6959	1.660	0.118 #
4)	L1 AR-1016-2	5.139	4.296	150045	19129	1.298	0.229 #
5)	L1 AR-1016-3	5.225	4.480	44112	15543	0.640	0.352 #
6)	L1 AR-1016-4	5.319	4.523	75895	35418	1.290	0.993
7)	L1 AR-1016-5	5.631	4.752	23405	80746	0.424	1.700 #
8)	L2 AR-1221-1	4.096	3.367	12079234	14816	463.668	0.729 #
9)	L2 AR-1221-2	4.201	3.430f	600677	10047868	31.585	661.073 #
10)	L2 AR-1221-3	4.278	3.530	230814	37260	3.755	0.807 #
11)	L3 AR-1232-1	4.278	3.530	230814	37260	4.067	0.878 #
12)	L3 AR-1232-2	4.839	4.296	101298	19129	3.799	0.477 #
13)	L3 AR-1232-3	5.139	4.480	150045	15543	2.787	0.723 #
14)	L3 AR-1232-4	5.319	4.570	75895	8474	2.869	0.447 #
15)	L3 AR-1232-5	5.431	4.752	273632	80746	14.845	3.776 #
16)	L4 AR-1242-1	5.121	4.271	127276	23760	2.130	0.514 #
17)	L4 AR-1242-2	5.139	4.296	150045	19129	1.671	0.295 #
18)	L4 AR-1242-3	5.225	4.480	44112	15543	0.795	0.449 #
19)	L4 AR-1242-4	5.319	4.570	75895	8474	1.672	0.252 #
20)	L4 AR-1242-5	6.120	5.124	1712834	10220	36.876	0.233 #
21)	L5 AR-1248-1	5.121	4.280	127276	6959	2.950	0.203 #
22)	L5 AR-1248-2	5.431	4.523	273632	35418	4.577	0.756 #
23)	L5 AR-1248-3	5.631	4.570	23405	8474	0.335	0.173 #
24)	L5 AR-1248-4	6.077	4.752	871309	80746	11.020	1.390 #
25)	L5 AR-1248-5	6.120	5.179	1712834	47278	21.877	0.805 #
26)	L6 AR-1254-1	6.043	5.124	692194	10220	8.375	0.116 #
27)	L6 AR-1254-2	6.294	5.284	225460	45484	1.831	0.603 #
28)	L6 AR-1254-3	6.695	5.716	447281	34957	3.721	0.297 #
29)	L6 AR-1254-4	0.000	5.967	0	41593	N.D.	0.551 #
30)	L6 AR-1254-5	7.489f	0.000	1901829	0	20.302	N.D. #
31)	L7 AR-1260-1	6.859	5.854	169255	65250	1.917	0.824 #
32)	L7 AR-1260-2	7.142	6.067	1191354	52616	11.629	0.562 #
33)	L7 AR-1260-3	7.552	6.226	159399	157369	1.935	1.777
34)	L7 AR-1260-4	7.786	0.000	628050	0	7.412	N.D. #
35)	L7 AR-1260-5	8.129	6.999	51960	113022	0.316	0.665 #
36)	L8 AR-1262-1	7.552	0.000	159399	0	1.385	N.D. #
37)	L8 AR-1262-2	8.129	0.000	51960	0	0.290	N.D. #
38)	L8 AR-1262-3	8.456	7.299	188720	258640	1.448	3.378 #
39)	L8 AR-1262-4	8.574f	7.374	1027609	30370	16.316	0.213 #
40)	L8 AR-1262-5	9.221	7.905	345539	421494	5.061	6.017
41)	L9 AR-1268-1	8.456	7.299	188720	258640	0.694	1.166 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041322\
 Data File : PP045875.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Apr 2022 11:17
 Operator : YP\AJ
 Sample : PB144019BL
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 13 02:43:04 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 12 05:31:12 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.574f	7.374	1027609	30370	4.062	0.150 #
43) L9	AR-1268-3	8.798	7.609	110162	309980	0.485	1.820 #
44) L9	AR-1268-4	9.231	7.905	171800	421494	1.904	5.340 #
45) L9	AR-1268-5	9.650	8.230	260268	253872	0.383	0.458


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041322\
Data File : PP045875.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Apr 2022 11:17
Operator : YP\AJ
Sample : PB144019BL
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 13 02:43:04 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041222.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 12 05:31:12 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.875 min
Delta R.T.: 0.000 min
Response: 49627918
Conc: 20.18 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.137 min
Delta R.T.: 0.000 min
Response: 33296964
Conc: 20.95 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.983 min
Delta R.T.: 0.005 min
Response: 38796632
Conc: 23.63 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.475 min
Delta R.T.: -0.002 min
Response: 31527199
Conc: 22.12 ng/ml



#3 AR-1016-1


R.T.: 5.121 min
Delta R.T.: -0.003 min
Response: 127276
Conc: 1.66 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.280 min
Delta R.T.: 0.004 min
Response: 6959
Conc: 0.12 ng/ml




#4 AR-1016-2

R.T.: 5.139 min
Delta R.T.: -0.009 min
Response: 150045
Conc: 1.30 ng/ml



#4 AR-1016-2


R.T.: 4.296 min
Delta R.T.: 0.002 min
Response: 19129
Conc: 0.23 ng/ml



#5 AR-1016-3


R.T.: 5.225 min
Delta R.T.: 0.011 min
Response: 44112
Conc: 0.64 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.480 min
Delta R.T.: 0.002 min
Response: 15543
Conc: 0.35 ng/ml




#6 AR-1016-4

R.T.: 5.319 min
Delta R.T.: 0.000 min
Response: 75895
Conc: 1.29 ng/ml



#6 AR-1016-4


R.T.: 4.523 min
Delta R.T.: -0.004 min
Response: 35418
Conc: 0.99 ng/ml



#7 AR-1016-5


R.T.: 5.631 min
Delta R.T.: -0.008 min
Response: 23405
Conc: 0.42 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 4.752 min
Delta R.T.: 0.002 min
Response: 80746
Conc: 1.70 ng/ml




#8 AR-1221-1

R.T.: 4.096 min
Delta R.T.: -0.006 min
Response: 12079234
Conc: 463.67 ng/ml



#8 AR-1221-1


R.T.: 3.367 min
Delta R.T.: 0.004 min
Response: 14816
Conc: 0.73 ng/ml



#9 AR-1221-2


R.T.: 4.201 min
Delta R.T.: 0.007 min
Response: 600677
Conc: 31.59 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.430 min
Delta R.T.: -0.022 min
Response: 10047868
Conc: 661.07 ng/ml




#10 AR-1221-3

R.T.: 4.278 min
Delta R.T.: 0.004 min
Response: 230814
Conc: 3.75 ng/ml



#10 AR-1221-3


R.T.: 3.530 min
Delta R.T.: -0.002 min
Response: 37260
Conc: 0.81 ng/ml



#11 AR-1232-1


R.T.: 4.278 min
Delta R.T.: 0.006 min
Response: 230814
Conc: 4.07 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.530 min
Delta R.T.: 0.000 min
Response: 37260
Conc: 0.88 ng/ml




#12 AR-1232-2

R.T.: 4.839 min
Delta R.T.: 0.004 min
Response: 101298
Conc: 3.80 ng/ml



#12 AR-1232-2


R.T.: 4.296 min
Delta R.T.: 0.003 min
Response: 19129
Conc: 0.48 ng/ml



#13 AR-1232-3


R.T.: 5.139 min
Delta R.T.: -0.008 min
Response: 150045
Conc: 2.79 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.480 min
Delta R.T.: 0.003 min
Response: 15543
Conc: 0.72 ng/ml




#14 AR-1232-4

R.T.: 5.319 min
Delta R.T.: 0.000 min
Response: 75895
Conc: 2.87 ng/ml



#14 AR-1232-4


R.T.: 4.570 min
Delta R.T.: 0.000 min
Response: 8474
Conc: 0.45 ng/ml



#15 AR-1232-5


R.T.: 5.431 min
Delta R.T.: 0.011 min
Response: 273632
Conc: 14.84 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 4.752 min
Delta R.T.: 0.002 min
Response: 80746
Conc: 3.78 ng/ml




#16 AR-1242-1

R.T.: 5.121 min
Delta R.T.: -0.005 min
Response: 127276
Conc: 2.13 ng/ml



#16 AR-1242-1


R.T.: 4.271 min
Delta R.T.: -0.005 min
Response: 23760
Conc: 0.51 ng/ml



#17 AR-1242-2


R.T.: 5.139 min
Delta R.T.: -0.010 min
Response: 150045
Conc: 1.67 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.296 min
Delta R.T.: 0.003 min
Response: 19129
Conc: 0.29 ng/ml




#18 AR-1242-3

R.T.: 5.225 min
Delta R.T.: 0.010 min
Response: 44112
Conc: 0.80 ng/ml



#18 AR-1242-3


R.T.: 4.480 min
Delta R.T.: 0.002 min
Response: 15543
Conc: 0.45 ng/ml



#19 AR-1242-4


R.T.: 5.319 min
Delta R.T.: -0.001 min
Response: 75895
Conc: 1.67 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.570 min
Delta R.T.: 0.000 min
Response: 8474
Conc: 0.25 ng/ml




#20 AR-1242-5

R.T.: 6.120 min
Delta R.T.: 0.001 min
Response: 1712834
Conc: 36.88 ng/ml



#20 AR-1242-5


R.T.: 5.124 min
Delta R.T.: -0.002 min
Response: 10220
Conc: 0.23 ng/ml



#21 AR-1248-1


R.T.: 5.121 min
Delta R.T.: -0.004 min
Response: 127276
Conc: 2.95 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.280 min
Delta R.T.: 0.004 min
Response: 6959
Conc: 0.20 ng/ml




#22 AR-1248-2

R.T.: 5.431 min
Delta R.T.: 0.010 min
Response: 273632
Conc: 4.58 ng/ml



#22 AR-1248-2


R.T.: 4.523 min
Delta R.T.: -0.004 min
Response: 35418
Conc: 0.76 ng/ml



#23 AR-1248-3


R.T.: 5.631 min
Delta R.T.: -0.007 min
Response: 23405
Conc: 0.33 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.570 min
Delta R.T.: 0.000 min
Response: 8474
Conc: 0.17 ng/ml




#24 AR-1248-4

R.T.: 6.077 min
Delta R.T.: 0.000 min
Response: 871309
Conc: 11.02 ng/ml



#24 AR-1248-4


R.T.: 4.752 min
Delta R.T.: 0.001 min
Response: 80746
Conc: 1.39 ng/ml



#25 AR-1248-5


R.T.: 6.120 min
Delta R.T.: 0.003 min
Response: 1712834
Conc: 21.88 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.179 min
Delta R.T.: 0.009 min
Response: 47278
Conc: 0.81 ng/ml




#26 AR-1254-1

R.T.: 6.043 min
Delta R.T.: -0.007 min
Response: 692194
Conc: 8.38 ng/ml



#26 AR-1254-1


R.T.: 5.124 min
Delta R.T.: 0.000 min
Response: 10220
Conc: 0.12 ng/ml



#27 AR-1254-2


R.T.: 6.294 min
Delta R.T.: 0.006 min
Response: 225460
Conc: 1.83 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.284 min
Delta R.T.: -0.001 min
Response: 45484
Conc: 0.60 ng/ml




#28 AR-1254-3

R.T.: 6.695 min
Delta R.T.: 0.006 min
Response: 447281
Conc: 3.72 ng/ml



#28 AR-1254-3


R.T.: 5.716 min
Delta R.T.: 0.000 min
Response: 34957
Conc: 0.30 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T.: 7.002 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 5.967 min
Delta R.T.: 0.002 min
Response: 41593
Conc: 0.55 ng/ml




#30 AR-1254-5

R.T.: 7.489 min
Delta R.T.: 0.027 min
Response: 1901829
Conc: 20.30 ng/ml



#30 AR-1254-5


R.T.: 0.000 min
Exp R.T.: 6.415 min
Response: 0
Conc: N.D.



#31 AR-1260-1


R.T.: 6.859 min
Delta R.T.: -0.006 min
Response: 169255
Conc: 1.92 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 5.854 min
Delta R.T.: -0.002 min
Response: 65250
Conc: 0.82 ng/ml




#32 AR-1260-2

R.T.: 7.142 min
Delta R.T.: -0.007 min
Response: 1191354
Conc: 11.63 ng/ml



#32 AR-1260-2


R.T.: 6.067 min
Delta R.T.: 0.005 min
Response: 52616
Conc: 0.56 ng/ml



#33 AR-1260-3


R.T.: 7.552 min
Delta R.T.: 0.008 min
Response: 159399
Conc: 1.93 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.226 min
Delta R.T.: 0.002 min
Response: 157369
Conc: 1.78 ng/ml




#34 AR-1260-4

R.T.: 7.786 min
Delta R.T.: -0.004 min
Response: 628050
Conc: 7.41 ng/ml



#34 AR-1260-4


R.T.: 0.000 min
Exp R.T. : 6.733 min
Response: 0
Conc: N.D.



#35 AR-1260-5


R.T.: 8.129 min
Delta R.T.: -0.003 min
Response: 51960
Conc: 0.32 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 6.999 min
Delta R.T.: -0.004 min
Response: 113022
Conc: 0.67 ng/ml




#36 AR-1262-1

R.T.: 7.552 min
Delta R.T.: 0.010 min
Response: 159399
Conc: 1.39 ng/ml



#36 AR-1262-1


R.T.: 0.000 min
Exp R.T.: 6.455 min
Response: 0
Conc: N.D.



#37 AR-1262-2


R.T.: 8.129 min
Delta R.T.: 0.000 min
Response: 51960
Conc: 0.29 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 0.000 min
Exp R.T. : 6.734 min
Response: 0
Conc: N.D.




#38 AR-1262-3

R.T.: 8.456 min
Delta R.T.: 0.000 min
Response: 188720
Conc: 1.45 ng/ml



#38 AR-1262-3


R.T.: 7.299 min
Delta R.T.: -0.007 min
Response: 258640
Conc: 3.38 ng/ml



#39 AR-1262-4


R.T.: 8.574 min
Delta R.T.: 0.023 min
Response: 1027609
Conc: 16.32 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.374 min
Delta R.T.: -0.004 min
Response: 30370
Conc: 0.21 ng/ml




#40 AR-1262-5

R.T.: 9.221 min
Delta R.T.: -0.003 min
Response: 345539
Conc: 5.06 ng/ml



#40 AR-1262-5


R.T.: 7.905 min
Delta R.T.: -0.014 min
Response: 421494
Conc: 6.02 ng/ml



#41 AR-1268-1


R.T.: 8.456 min
Delta R.T.: -0.002 min
Response: 188720
Conc: 0.69 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.299 min
Delta R.T.: -0.008 min
Response: 258640
Conc: 1.17 ng/ml




#42 AR-1268-2

R.T.: 8.574 min
Delta R.T.: 0.016 min
Response: 1027609
Conc: 4.06 ng/ml



#42 AR-1268-2


R.T.: 7.374 min
Delta R.T.: -0.005 min
Response: 30370
Conc: 0.15 ng/ml



#43 AR-1268-3


R.T.: 8.798 min
Delta R.T.: 0.008 min
Response: 110162
Conc: 0.48 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.609 min
Delta R.T.: 0.012 min
Response: 309980
Conc: 1.82 ng/ml




#44 AR-1268-4

R.T.: 9.231 min
Delta R.T.: 0.002 min
Response: 171800
Conc: 1.90 ng/ml



#44 AR-1268-4


R.T.: 7.905 min
Delta R.T.: -0.015 min
Response: 421494
Conc: 5.34 ng/ml



#45 AR-1268-5

R.T.: 9.650 min
Delta R.T.: 0.004 min
Response: 260268
Conc: 0.38 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.230 min
Delta R.T.: 0.009 min
Response: 253872
Conc: 0.46 ng/ml