

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041322\
 Data File : PP045926.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Apr 2022 02:10
 Operator : YP\AJ
 Sample : N2348-14 10X
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 13 03:08:04 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 12 05:31:12 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.876	3.139	3857901	2933597	1.568	1.846
2)	SA Decachlor...	9.973	8.476	3540777	3179644	2.157	2.231
Target Compounds							
3)	L1 AR-1016-1	5.122	4.275	89714	20458	1.170	0.348 #
4)	L1 AR-1016-2	5.152	4.297	168340	106373	1.456	1.273
5)	L1 AR-1016-3	5.222	4.475	425185	11632	6.171	0.263 #
6)	L1 AR-1016-4	5.314	4.524	359588	10800	6.111	0.303 #
7)	L1 AR-1016-5	5.633	4.749	68617	14457	1.244	0.304 #
8)	L2 AR-1221-1	4.108	3.366	131289	17738	5.040	0.873 #
9)	L2 AR-1221-2	4.194	3.446	102630	37725	5.397	2.482 #
10)	L2 AR-1221-3	4.270	3.528	160095	33910	2.604	0.735 #
11)	L3 AR-1232-1	4.270	3.528	160095	33910	2.821	0.799 #
12)	L3 AR-1232-2	4.831	4.297	142795	106373	5.355	2.654 #
13)	L3 AR-1232-3	5.152	4.475	168340	11632	3.127	0.541 #
14)	L3 AR-1232-4	5.314	4.569	359588	23967	13.593	1.265 #
15)	L3 AR-1232-5	5.422	4.749	171426	14457	9.300	0.676 #
16)	L4 AR-1242-1	5.130	4.275	63529	20458	1.063	0.442 #
17)	L4 AR-1242-2	5.152	4.297	168340	106373	1.875	1.640
18)	L4 AR-1242-3	5.222	4.475	425185	11632	7.665	0.336 #
19)	L4 AR-1242-4	5.314	4.569	359588	23967	7.924	0.713 #
20)	L4 AR-1242-5	6.108	5.120	57387	14529	1.236	0.331 #
21)	L5 AR-1248-1	5.122	4.275	89714	20458	2.080	0.597 #
22)	L5 AR-1248-2	5.422	4.524	171426	10800	2.868	0.231 #
23)	L5 AR-1248-3	5.633	4.569	68617	23967	0.981	0.490 #
24)	L5 AR-1248-4	6.077	4.749	77895	14457	0.985	0.249 #
25)	L5 AR-1248-5	6.108	5.170	57387	14972	0.733	0.255 #
26)	L6 AR-1254-1	6.045	5.120	172362	14529	2.085	0.165 #
27)	L6 AR-1254-2	6.287	5.288	64895	32985	0.527	0.437
28)	L6 AR-1254-3	6.698	5.718	298976	16565	2.487	0.141 #
29)	L6 AR-1254-4	7.007	5.959	183819	12147	2.084	0.161 #
30)	L6 AR-1254-5	7.451	6.411	658817	46385	7.033	0.453 #
31)	L7 AR-1260-1	6.871	5.854	98512	24447	1.116	0.309 #
32)	L7 AR-1260-2	7.149	6.054	165303	31884	1.614	0.341 #
33)	L7 AR-1260-3	7.541	6.218	107387	157907	1.303	1.783 #
34)	L7 AR-1260-4	7.789	6.727	93389	36171	1.102	0.504 #
35)	L7 AR-1260-5	8.131	6.999	103588	41649	0.631	0.245 #
36)	L8 AR-1262-1	7.541	6.456	107387	34847	0.933	0.329 #
37)	L8 AR-1262-2	8.131	6.727	103588	36171	0.578	0.384 #
38)	L8 AR-1262-3	8.459	7.302	135307	129890	1.038	1.697 #
39)	L8 AR-1262-4	8.545	7.377	248339	113955	3.943	0.801 #
40)	L8 AR-1262-5	9.226	7.922	274270	81853	4.017	1.168 #
41)	L9 AR-1268-1	8.459	7.302	135307	129890	0.498	0.585

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041322\
 Data File : PP045926.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Apr 2022 02:10
 Operator : YP\AJ
 Sample : N2348-14 10X
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 13 03:08:04 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 12 05:31:12 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.545	7.377	248339	113955	0.982	0.562 #
43) L9	AR-1268-3	8.785	7.591	82361	29629	0.362	0.174 #
44) L9	AR-1268-4	9.226	7.922	274270	81853	3.039	1.037 #
45) L9	AR-1268-5	9.634	8.220	94049	35055	0.138	0.063 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041322\
Data File : PP045926.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Apr 2022 02:10
Operator : YP\AJ
Sample : N2348-14 10X
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 13 03:08:04 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041222.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 12 05:31:12 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.876 min
Delta R.T.: 0.000 min
Response: 3857901
Conc: 1.57 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.139 min
Delta R.T.: 0.000 min
Response: 2933597
Conc: 1.85 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.973 min
Delta R.T.: -0.005 min
Response: 3540777
Conc: 2.16 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.476 min
Delta R.T.: 0.000 min
Response: 3179644
Conc: 2.23 ng/ml



#3 AR-1016-1


R.T.: 5.122 min
Delta R.T.: -0.002 min
Response: 89714
Conc: 1.17 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.275 min
Delta R.T.: 0.000 min
Response: 20458
Conc: 0.35 ng/ml




#4 AR-1016-2

R.T.: 5.152 min
Delta R.T.: 0.004 min
Response: 168340
Conc: 1.46 ng/ml



#4 AR-1016-2


R.T.: 4.297 min
Delta R.T.: 0.003 min
Response: 106373
Conc: 1.27 ng/ml



#5 AR-1016-3


R.T.: 5.222 min
Delta R.T.: 0.008 min
Response: 425185
Conc: 6.17 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.475 min
Delta R.T.: -0.003 min
Response: 11632
Conc: 0.26 ng/ml




#6 AR-1016-4

R.T.: 5.314 min
Delta R.T.: -0.005 min
Response: 359588
Conc: 6.11 ng/ml



#6 AR-1016-4


R.T.: 4.524 min
Delta R.T.: -0.003 min
Response: 10800
Conc: 0.30 ng/ml



#7 AR-1016-5


R.T.: 5.633 min
Delta R.T.: -0.005 min
Response: 68617
Conc: 1.24 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 4.749 min
Delta R.T.: -0.001 min
Response: 14457
Conc: 0.30 ng/ml




#8 AR-1221-1

R.T.: 4.108 min
Delta R.T.: 0.006 min
Response: 131289
Conc: 5.04 ng/ml



#8 AR-1221-1


R.T.: 3.366 min
Delta R.T.: 0.003 min
Response: 17738
Conc: 0.87 ng/ml



#9 AR-1221-2


R.T.: 4.194 min
Delta R.T.: 0.000 min
Response: 102630
Conc: 5.40 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.446 min
Delta R.T.: -0.006 min
Response: 37725
Conc: 2.48 ng/ml




#10 AR-1221-3

R.T.: 4.270 min
Delta R.T.: -0.003 min
Response: 160095
Conc: 2.60 ng/ml



#10 AR-1221-3


R.T.: 3.528 min
Delta R.T.: -0.003 min
Response: 33910
Conc: 0.73 ng/ml



#11 AR-1232-1


R.T.: 4.270 min
Delta R.T.: -0.002 min
Response: 160095
Conc: 2.82 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.528 min
Delta R.T.: -0.002 min
Response: 33910
Conc: 0.80 ng/ml




#12 AR-1232-2

R.T.: 4.831 min
Delta R.T.: -0.004 min
Response: 142795
Conc: 5.35 ng/ml



#12 AR-1232-2


R.T.: 4.297 min
Delta R.T.: 0.003 min
Response: 106373
Conc: 2.65 ng/ml



#13 AR-1232-3


R.T.: 5.152 min
Delta R.T.: 0.005 min
Response: 168340
Conc: 3.13 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.475 min
Delta R.T.: -0.003 min
Response: 11632
Conc: 0.54 ng/ml




#14 AR-1232-4

R.T.: 5.314 min
Delta R.T.: -0.004 min
Response: 359588
Conc: 13.59 ng/ml



#14 AR-1232-4


R.T.: 4.569 min
Delta R.T.: 0.000 min
Response: 23967
Conc: 1.27 ng/ml



#15 AR-1232-5


R.T.: 5.422 min
Delta R.T.: 0.002 min
Response: 171426
Conc: 9.30 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 4.749 min
Delta R.T.: 0.000 min
Response: 14457
Conc: 0.68 ng/ml




#16 AR-1242-1

R.T.: 5.130 min
Delta R.T.: 0.004 min
Response: 63529
Conc: 1.06 ng/ml



#16 AR-1242-1


R.T.: 4.275 min
Delta R.T.: 0.000 min
Response: 20458
Conc: 0.44 ng/ml



#17 AR-1242-2


R.T.: 5.152 min
Delta R.T.: 0.003 min
Response: 168340
Conc: 1.88 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.297 min
Delta R.T.: 0.003 min
Response: 106373
Conc: 1.64 ng/ml




#18 AR-1242-3

R.T.: 5.222 min
Delta R.T.: 0.008 min
Response: 425185
Conc: 7.66 ng/ml



#18 AR-1242-3


R.T.: 4.475 min
Delta R.T.: -0.003 min
Response: 11632
Conc: 0.34 ng/ml



#19 AR-1242-4


R.T.: 5.314 min
Delta R.T.: -0.006 min
Response: 359588
Conc: 7.92 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.569 min
Delta R.T.: -0.001 min
Response: 23967
Conc: 0.71 ng/ml




#20 AR-1242-5

R.T.: 6.108 min
Delta R.T.: -0.011 min
Response: 57387
Conc: 1.24 ng/ml



#20 AR-1242-5


R.T.: 5.120 min
Delta R.T.: -0.006 min
Response: 14529
Conc: 0.33 ng/ml



#21 AR-1248-1


R.T.: 5.122 min
Delta R.T.: -0.003 min
Response: 89714
Conc: 2.08 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.275 min
Delta R.T.: 0.000 min
Response: 20458
Conc: 0.60 ng/ml




#22 AR-1248-2

R.T.: 5.422 min
Delta R.T.: 0.001 min
Response: 171426
Conc: 2.87 ng/ml



#22 AR-1248-2


R.T.: 4.524 min
Delta R.T.: -0.003 min
Response: 10800
Conc: 0.23 ng/ml



#23 AR-1248-3


R.T.: 5.633 min
Delta R.T.: -0.005 min
Response: 68617
Conc: 0.98 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.569 min
Delta R.T.: 0.000 min
Response: 23967
Conc: 0.49 ng/ml




#24 AR-1248-4

R.T.: 6.077 min
Delta R.T.: 0.000 min
Response: 77895
Conc: 0.99 ng/ml



#24 AR-1248-4


R.T.: 4.749 min
Delta R.T.: -0.001 min
Response: 14457
Conc: 0.25 ng/ml



#25 AR-1248-5


R.T.: 6.108 min
Delta R.T.: -0.009 min
Response: 57387
Conc: 0.73 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.170 min
Delta R.T.: 0.000 min
Response: 14972
Conc: 0.26 ng/ml




#26 AR-1254-1

R.T.: 6.045 min
Delta R.T.: -0.005 min
Response: 172362
Conc: 2.09 ng/ml



#26 AR-1254-1


R.T.: 5.120 min
Delta R.T.: -0.005 min
Response: 14529
Conc: 0.16 ng/ml



#27 AR-1254-2


R.T.: 6.287 min
Delta R.T.: -0.001 min
Response: 64895
Conc: 0.53 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.288 min
Delta R.T.: 0.002 min
Response: 32985
Conc: 0.44 ng/ml




#28 AR-1254-3

R.T.: 6.698 min
Delta R.T.: 0.010 min
Response: 298976
Conc: 2.49 ng/ml



#28 AR-1254-3


R.T.: 5.718 min
Delta R.T.: 0.003 min
Response: 16565
Conc: 0.14 ng/ml



#29 AR-1254-4


R.T.: 7.007 min
Delta R.T.: 0.005 min
Response: 183819
Conc: 2.08 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 5.959 min
Delta R.T.: -0.006 min
Response: 12147
Conc: 0.16 ng/ml




#30 AR-1254-5

R.T.: 7.451 min
Delta R.T.: -0.011 min
Response: 658817
Conc: 7.03 ng/ml



#30 AR-1254-5


R.T.: 6.411 min
Delta R.T.: -0.004 min
Response: 46385
Conc: 0.45 ng/ml



#31 AR-1260-1


R.T.: 6.871 min
Delta R.T.: 0.006 min
Response: 98512
Conc: 1.12 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 5.854 min
Delta R.T.: -0.002 min
Response: 24447
Conc: 0.31 ng/ml




#32 AR-1260-2

R.T.: 7.149 min
Delta R.T.: 0.001 min
Response: 165303
Conc: 1.61 ng/ml



#32 AR-1260-2


R.T.: 6.054 min
Delta R.T.: -0.009 min
Response: 31884
Conc: 0.34 ng/ml



#33 AR-1260-3


R.T.: 7.541 min
Delta R.T.: -0.002 min
Response: 107387
Conc: 1.30 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.218 min
Delta R.T.: -0.006 min
Response: 157907
Conc: 1.78 ng/ml




#34 AR-1260-4

R.T.: 7.789 min
Delta R.T.: -0.001 min
Response: 93389
Conc: 1.10 ng/ml



#34 AR-1260-4


R.T.: 6.727 min
Delta R.T.: -0.006 min
Response: 36171
Conc: 0.50 ng/ml



#35 AR-1260-5


R.T.: 8.131 min
Delta R.T.: 0.000 min
Response: 103588
Conc: 0.63 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 6.999 min
Delta R.T.: -0.004 min
Response: 41649
Conc: 0.25 ng/ml




#36 AR-1262-1

R.T.: 7.541 min
Delta R.T.: 0.000 min
Response: 107387
Conc: 0.93 ng/ml



#36 AR-1262-1


R.T.: 6.456 min
Delta R.T.: 0.002 min
Response: 34847
Conc: 0.33 ng/ml



#37 AR-1262-2


R.T.: 8.131 min
Delta R.T.: 0.002 min
Response: 103588
Conc: 0.58 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.727 min
Delta R.T.: -0.007 min
Response: 36171
Conc: 0.38 ng/ml




#38 AR-1262-3

R.T.: 8.459 min
Delta R.T.: 0.003 min
Response: 135307
Conc: 1.04 ng/ml



#38 AR-1262-3


R.T.: 7.302 min
Delta R.T.: -0.005 min
Response: 129890
Conc: 1.70 ng/ml



#39 AR-1262-4


R.T.: 8.545 min
Delta R.T.: -0.006 min
Response: 248339
Conc: 3.94 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.377 min
Delta R.T.: 0.000 min
Response: 113955
Conc: 0.80 ng/ml




#40 AR-1262-5

R.T.: 9.226 min
Delta R.T.: 0.002 min
Response: 274270
Conc: 4.02 ng/ml



#40 AR-1262-5


R.T.: 7.922 min
Delta R.T.: 0.002 min
Response: 81853
Conc: 1.17 ng/ml



#41 AR-1268-1


R.T.: 8.459 min
Delta R.T.: 0.000 min
Response: 135307
Conc: 0.50 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.302 min
Delta R.T.: -0.005 min
Response: 129890
Conc: 0.59 ng/ml




#42 AR-1268-2

R.T.: 8.545 min
Delta R.T.: -0.013 min
Response: 248339
Conc: 0.98 ng/ml



#42 AR-1268-2


R.T.: 7.377 min
Delta R.T.: -0.001 min
Response: 113955
Conc: 0.56 ng/ml



#43 AR-1268-3


R.T.: 8.785 min
Delta R.T.: -0.004 min
Response: 82361
Conc: 0.36 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.591 min
Delta R.T.: -0.006 min
Response: 29629
Conc: 0.17 ng/ml




#44 AR-1268-4

R.T.: 9.226 min
Delta R.T.: -0.004 min
Response: 274270
Conc: 3.04 ng/ml



#44 AR-1268-4


R.T.: 7.922 min
Delta R.T.: 0.002 min
Response: 81853
Conc: 1.04 ng/ml



#45 AR-1268-5

R.T.: 9.634 min
Delta R.T.: -0.012 min
Response: 94049
Conc: 0.14 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.220 min
Delta R.T.: -0.001 min
Response: 35055
Conc: 0.06 ng/ml