

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041322\
 Data File : PP045871.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Apr 2022 10:16
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 13 02:40:48 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 12 05:31:12 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.881	3.138	133.4E6	87533444	54.217	55.088
2) SA Decachlor...	9.992	8.479	93496424	77381984	56.955	54.291
Target Compounds						
21) L5 AR-1248-1	5.131	4.276	23471516	17905380	544.097	522.125
22) L5 AR-1248-2	5.428	4.528	31736968	24207842	530.881	516.744
23) L5 AR-1248-3	5.645	4.570	37772727	25317269	539.980	517.963
24) L5 AR-1248-4	6.084	4.752	43301600	29997441	547.681	516.294
25) L5 AR-1248-5	6.125	5.170	42612348	30075807	544.266	512.407



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041322\
Data File : PP045871.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Apr 2022 10:16
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 13 02:40:48 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041222.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 12 05:31:12 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.881 min
Delta R.T.: 0.005 min
Response: 133357195
Conc: 54.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.138 min
Delta R.T.: 0.000 min
Response: 87533444
Conc: 55.09 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.992 min
Delta R.T.: 0.014 min
Response: 93496424
Conc: 56.96 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.479 min
Delta R.T.: 0.002 min
Response: 77381984
Conc: 54.29 ng/ml



#21 AR-1248-1


R.T.: 5.131 min
Delta R.T.: 0.007 min
Response: 23471516
Conc: 544.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.276 min
Delta R.T.: 0.000 min
Response: 17905380
Conc: 522.13 ng/ml




#22 AR-1248-2

R.T.: 5.428 min
Delta R.T.: 0.008 min
Response: 31736968
Conc: 530.88 ng/ml



#22 AR-1248-2


R.T.: 4.528 min
Delta R.T.: 0.001 min
Response: 24207842
Conc: 516.74 ng/ml



#23 AR-1248-3


R.T.: 5.645 min
Delta R.T.: 0.007 min
Response: 37772727
Conc: 539.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.570 min
Delta R.T.: 0.000 min
Response: 25317269
Conc: 517.96 ng/ml




#24 AR-1248-4

R.T.: 6.084 min
Delta R.T.: 0.008 min
Response: 43301600
Conc: 547.68 ng/ml



#24 AR-1248-4


R.T.: 4.752 min
Delta R.T.: 0.001 min
Response: 29997441
Conc: 516.29 ng/ml



#25 AR-1248-5

R.T.: 6.125 min
Delta R.T.: 0.008 min
Response: 42612348
Conc: 544.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.170 min
Delta R.T.: 0.000 min
Response: 30075807
Conc: 512.41 ng/ml