

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041322\
 Data File : PP045887.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Apr 2022 14:49
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 13 02:48:50 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 12 05:31:12 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.874	3.137	137.0E6	89661938	55.693	56.428
2) SA Decachlor...	9.974	8.475	79617254	78796136	48.501	55.283
Target Compounds						
26) L6 AR-1254-1	6.046	5.124	40305395	45592542	487.670	516.872
27) L6 AR-1254-2	6.285	5.284	59882313	38908740	486.272	515.538
28) L6 AR-1254-3	6.684	5.714	59453639	61171573	494.597	520.121
29) L6 AR-1254-4	6.998	5.963	43964498	39681685	498.486	525.829
30) L6 AR-1254-5	7.458	6.413	45465205	46153860	485.338	450.491



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041322\
Data File : PP045887.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Apr 2022 14:49
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 13 02:48:50 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041222.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 12 05:31:12 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.874 min
Delta R.T.: -0.002 min
Response: 136987184
Conc: 55.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.137 min
Delta R.T.: 0.000 min
Response: 89661938
Conc: 56.43 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.974 min
Delta R.T.: -0.004 min
Response: 79617254
Conc: 48.50 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.475 min
Delta R.T.: -0.002 min
Response: 78796136
Conc: 55.28 ng/ml



#26 AR-1254-1


R.T.: 6.046 min
Delta R.T.: -0.004 min
Response: 40305395
Conc: 487.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.124 min
Delta R.T.: 0.000 min
Response: 45592542
Conc: 516.87 ng/ml




#27 AR-1254-2

R.T.: 6.285 min
Delta R.T.: -0.003 min
Response: 59882313
Conc: 486.27 ng/ml



#27 AR-1254-2


R.T.: 5.284 min
Delta R.T.: -0.001 min
Response: 38908740
Conc: 515.54 ng/ml



#28 AR-1254-3


R.T.: 6.684 min
Delta R.T.: -0.004 min
Response: 59453639
Conc: 494.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.714 min
Delta R.T.: -0.001 min
Response: 61171573
Conc: 520.12 ng/ml




#29 AR-1254-4

R.T.: 6.998 min
Delta R.T.: -0.004 min
Response: 43964498
Conc: 498.49 ng/ml



#29 AR-1254-4


R.T.: 5.963 min
Delta R.T.: -0.002 min
Response: 39681685
Conc: 525.83 ng/ml



#30 AR-1254-5

R.T.: 7.458 min
Delta R.T.: -0.004 min
Response: 45465205
Conc: 485.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.413 min
Delta R.T.: -0.002 min
Response: 46153860
Conc: 450.49 ng/ml