

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041421\
 Data File : PP034891.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 Apr 2021 18:34
 Operator : DD\AJ
 Sample : M2025-07
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 WC-2

Manual Integrations
 APPROVED

Ankita
 4/15/2021 2:07:07 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 15 02:33:20 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 31 06:02:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.843	4.259	1756423	1102512	23.025	24.543
2) SA Decachlor...	10.759	9.555	1689073	658583	21.942	23.324
Target Compounds						
26) L6 AR-1254-1	7.058	6.376	53137	39802	19.326	21.574
27) L6 AR-1254-2	7.289	6.533	110974	32536	26.098	20.337
28) L6 AR-1254-3	7.669	6.953	94934	55571	20.800m	22.266
29) L6 AR-1254-4	7.959	7.188	86267	25633	27.901m	19.500 #
30) L6 AR-1254-5	8.397	7.612	400747	55829	112.765	27.469 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041421\
Data File : PP034891.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 Apr 2021 18:34
Operator : DD\AJ
Sample : M2025-07
Misc :
ALS Vial : 24 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampled :
WC-2


Manual Integrations
APPROVED

Ankita
4/15/2021 2:07:07 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 15 02:33:20 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033021.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 31 06:02:19 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.843 min
Delta R.T.: -0.003 min
Response: 1756423
Conc: 23.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
WC-2


Manual Integrations
APPROVED

Ankita
4/15/2021 2:07:07 PM




#1 Tetrachloro-m-xylene

R.T.: 4.259 min
Delta R.T.: -0.001 min
Response: 1102512
Conc: 24.54 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.759 min
Delta R.T.: -0.006 min
Response: 1689073
Conc: 21.94 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.555 min
Delta R.T.: -0.011 min
Response: 658583
Conc: 23.32 ng/ml




#26 AR-1254-1

R.T.: 7.058 min
Delta R.T.: -0.005 min
Response: 53137
Conc: 19.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
WC-2


Manual Integrations
APPROVED

Ankita
4/15/2021 2:07:07 PM




#26 AR-1254-1

R.T.: 6.376 min
Delta R.T.: -0.006 min
Response: 39802
Conc: 21.57 ng/ml




#27 AR-1254-2

R.T.: 7.289 min
Delta R.T.: -0.004 min
Response: 110974
Conc: 26.10 ng/ml



#27 AR-1254-2

R.T.: 6.533 min
Delta R.T.: -0.005 min
Response: 32536
Conc: 20.34 ng/ml




#28 AR-1254-3

R.T.: 7.669 min
Delta R.T.: -0.004 min
Response: 94934
Conc: 20.80 ng/ml m

Instrument :
ECD_P
ClientSampleId :
WC-2


Manual Integrations
APPROVED

Ankita
4/15/2021 2:07:07 PM




#28 AR-1254-3

R.T.: 6.953 min
Delta R.T.: -0.006 min
Response: 55571
Conc: 22.27 ng/ml




#29 AR-1254-4

R.T.: 7.959 min
Delta R.T.: -0.009 min
Response: 86267
Conc: 27.90 ng/ml m



#29 AR-1254-4

R.T.: 7.188 min
Delta R.T.: -0.006 min
Response: 25633
Conc: 19.50 ng/ml




#30 AR-1254-5

R.T.: 8.397 min
Delta R.T.: 0.002 min
Response: 400747
Conc: 112.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
WC-2

Manual Integrations
APPROVED

Ankita
4/15/2021 2:07:07 PM



#30 AR-1254-5

R.T.: 7.612 min
Delta R.T.: -0.009 min
Response: 55829
Conc: 27.47 ng/ml