

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041423\
 Data File : PP056934.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Apr 2023 18:57
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 14 22:11:17 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP040623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 07 04:14:17 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.435	3.660	114.6E6	79746058	53.350	45.852
2) SA Decachlor...	10.275	8.742	72349067	71086190	55.275	45.212
Target Compounds						
21) L5 AR-1248-1	5.610	4.760	19800303	15419243	527.738	468.420
22) L5 AR-1248-2	5.886	5.000	31799469	24556998	539.942	476.442
23) L5 AR-1248-3	6.092	5.043	34261921	25643876	537.762	477.704
24) L5 AR-1248-4	6.498	5.217	30545876	29739905	525.975	480.644
25) L5 AR-1248-5	6.536	5.614	30107645	24350125	533.005	474.256



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041423\
Data File : PP056934.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Apr 2023 18:57
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 14 22:11:17 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP040623.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 07 04:14:17 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.435 min
Delta R.T.: -0.002 min
Response: 114623773
Conc: 53.35 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.660 min
Delta R.T.: -0.005 min
Response: 79746058
Conc: 45.85 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.275 min
Delta R.T.: 0.000 min
Response: 72349067
Conc: 55.27 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.742 min
Delta R.T.: -0.009 min
Response: 71086190
Conc: 45.21 ng/ml



#21 AR-1248-1


R.T.: 5.610 min
Delta R.T.: 0.000 min
Response: 19800303
Conc: 527.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.760 min
Delta R.T.: -0.004 min
Response: 15419243
Conc: 468.42 ng/ml




#22 AR-1248-2

R.T.: 5.886 min
Delta R.T.: 0.000 min
Response: 31799469
Conc: 539.94 ng/ml



#22 AR-1248-2


R.T.: 5.000 min
Delta R.T.: -0.004 min
Response: 24556998
Conc: 476.44 ng/ml



#23 AR-1248-3


R.T.: 6.092 min
Delta R.T.: 0.000 min
Response: 34261921
Conc: 537.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.043 min
Delta R.T.: -0.004 min
Response: 25643876
Conc: 477.70 ng/ml




#24 AR-1248-4

R.T.: 6.498 min
Delta R.T.: 0.000 min
Response: 30545876
Conc: 525.98 ng/ml



#24 AR-1248-4


R.T.: 5.217 min
Delta R.T.: -0.004 min
Response: 29739905
Conc: 480.64 ng/ml



#25 AR-1248-5

R.T.: 6.536 min
Delta R.T.: -0.001 min
Response: 30107645
Conc: 533.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.614 min
Delta R.T.: -0.005 min
Response: 24350125
Conc: 474.26 ng/ml