

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041425\
 Data File : PP071260.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Apr 2025 18:22
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 14 22:57:02 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Mar 28 03:13:46 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.519	3.816	73180047	48135705	50.069	46.434
2) SA Decachlor...	10.243	8.857	50101759	31395143	47.597	39.832
Target Compounds						
16) L4 AR-1242-1	5.672	4.901	21725222	17074393	494.012	499.312
17) L4 AR-1242-2	5.694	4.920	32428911	23739160	504.210	481.938
18) L4 AR-1242-3	5.756	5.097	19372279	13587204	472.040	509.948
19) L4 AR-1242-4	5.854	5.181	16025939	13059735	525.039	495.095
20) L4 AR-1242-5	6.585	5.705	19038736	16402189	521.461	529.816



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041425\
Data File : PP071260.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Apr 2025 18:22
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 14 22:57:02 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.519 min
Delta R.T.: 0.007 min
Response: 73180047
Conc: 50.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 48135705
Conc: 46.43 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.243 min
Delta R.T.: 0.017 min
Response: 50101759
Conc: 47.60 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.857 min
Delta R.T.: -0.002 min
Response: 31395143
Conc: 39.83 ng/ml



#16 AR-1242-1


R.T.: 5.672 min
Delta R.T.: 0.008 min
Response: 21725222
Conc: 494.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.901 min
Delta R.T.: 0.000 min
Response: 17074393
Conc: 499.31 ng/ml




#17 AR-1242-2

R.T.: 5.694 min
Delta R.T.: 0.009 min
Response: 32428911
Conc: 504.21 ng/ml



#17 AR-1242-2


R.T.: 4.920 min
Delta R.T.: -0.001 min
Response: 23739160
Conc: 481.94 ng/ml



#18 AR-1242-3


R.T.: 5.756 min
Delta R.T.: 0.008 min
Response: 19372279
Conc: 472.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.097 min
Delta R.T.: -0.002 min
Response: 13587204
Conc: 509.95 ng/ml




#19 AR-1242-4

R.T.: 5.854 min
Delta R.T.: 0.009 min
Response: 16025939
Conc: 525.04 ng/ml



#19 AR-1242-4


R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 13059735
Conc: 495.10 ng/ml



#20 AR-1242-5

R.T.: 6.585 min
Delta R.T.: 0.010 min
Response: 19038736
Conc: 521.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.705 min
Delta R.T.: -0.002 min
Response: 16402189
Conc: 529.82 ng/ml