

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041522\
 Data File : PP046092.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Apr 2022 06:02
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 15 07:10:03 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 12 05:31:12 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.876	3.138	137.6E6	92876853	55.958	58.451
2) SA Decachlor...	9.973	8.472	80617454	78738414	49.110	55.243
Target Compounds						
21) L5 AR-1248-1	5.123	4.273	23999594	19175962	556.338	559.176
22) L5 AR-1248-2	5.419	4.525	31820844	25568747	532.284	545.794
23) L5 AR-1248-3	5.637	4.567	37335290	27053142	533.727	553.477
24) L5 AR-1248-4	6.075	4.748	40266168	32729410	509.288	563.314
25) L5 AR-1248-5	6.116	5.165	38481529	34202887	491.505	582.721


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041522\
Data File : PP046092.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Apr 2022 06:02
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 15 07:10:03 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041222.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 12 05:31:12 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.876 min
Delta R.T.: 0.000 min
Response: 137639852
Conc: 55.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.138 min
Delta R.T.: 0.000 min
Response: 92876853
Conc: 58.45 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.973 min
Delta R.T.: -0.005 min
Response: 80617454
Conc: 49.11 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.472 min
Delta R.T.: -0.005 min
Response: 78738414
Conc: 55.24 ng/ml



#21 AR-1248-1


R.T.: 5.123 min
Delta R.T.: -0.001 min
Response: 23999594
Conc: 556.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.273 min
Delta R.T.: -0.003 min
Response: 19175962
Conc: 559.18 ng/ml




#22 AR-1248-2

R.T.: 5.419 min
Delta R.T.: 0.000 min
Response: 31820844
Conc: 532.28 ng/ml



#22 AR-1248-2


R.T.: 4.525 min
Delta R.T.: -0.003 min
Response: 25568747
Conc: 545.79 ng/ml



#23 AR-1248-3


R.T.: 5.637 min
Delta R.T.: -0.001 min
Response: 37335290
Conc: 533.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.567 min
Delta R.T.: -0.003 min
Response: 27053142
Conc: 553.48 ng/ml




#24 AR-1248-4

R.T.: 6.075 min
Delta R.T.: -0.001 min
Response: 40266168
Conc: 509.29 ng/ml



#24 AR-1248-4


R.T.: 4.748 min
Delta R.T.: -0.003 min
Response: 32729410
Conc: 563.31 ng/ml



#25 AR-1248-5

R.T.: 6.116 min
Delta R.T.: -0.001 min
Response: 38481529
Conc: 491.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.165 min
Delta R.T.: -0.004 min
Response: 34202887
Conc: 582.72 ng/ml