

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_P\Data\PP041524\  
 Data File : PP064310.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 15 Apr 2024 11:56  
 Operator : YP\AJ  
 Sample : P2110-02  
 Misc :  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 ECD\_P  
 ClientSampleId :  
 BASEMENT-CEILING

Manual Integrations  
 APPROVED

Reviewed By :Yogesh Patel 04/16/2024  
 Supervised By :Ankita Jodhani 04/16/2024

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Apr 15 12:51:55 2024  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_P\methods\PP031924.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Tue Mar 19 06:23:02 2024  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound                    | RT#1   | RT#2  | Resp#1   | Resp#2   | ng/ml   | ng/ml   |
|-----------------------------|--------|-------|----------|----------|---------|---------|
| -----                       |        |       |          |          |         |         |
| System Monitoring Compounds |        |       |          |          |         |         |
| 1) SA Tetrachlo...          | 4.485  | 3.636 | 43212244 | 47629152 | 18.110  | 19.949  |
| 2) SA Decachlor...          | 10.348 | 8.665 | 47275786 | 44402755 | 18.620  | 16.430  |
| Target Compounds            |        |       |          |          |         |         |
| 26) L6 AR-1254-1            | 6.529  | 5.530 | 7865143  | 12013587 | 86.509m | 97.219  |
| 27) L6 AR-1254-2            | 6.751  | 5.679 | 13014998 | 10629938 | 94.937  | 92.611  |
| 28) L6 AR-1254-3            | 7.120  | 6.083 | 14012369 | 16104917 | 100.127 | 89.502  |
| 29) L6 AR-1254-4            | 7.408  | 6.313 | 9497071  | 9795687  | 104.240 | 106.302 |
| 30) L6 AR-1254-5            | 7.829  | 6.732 | 8069556  | 11513479 | 68.632  | 68.976m |
| -----                       |        |       |          |          |         |         |

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_P\Data\PP041524\  
Data File : PP064310.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 15 Apr 2024 11:56  
Operator : YP\AJ  
Sample : P2110-02  
Misc :  
ALS Vial : 11 Sample Multiplier: 1



Instrument :  
ECD\_P  
ClientSampleId :  
BASEMENT-CEILING


Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 04/16/2024  
Supervised By :Ankita Jodhani 04/16/2024

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Apr 15 12:51:55 2024  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_P\methods\PP031924.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Tue Mar 19 06:23:02 2024  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.485 min  
Delta R.T.: -0.004 min  
Response: 43212244  
Conc: 18.11 ng/ml

Instrument :  
ECD\_P  
ClientSampleId :  
BASEMENT-CEILING


Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 04/16/2024  
Supervised By :Ankita Jodhani 04/16/2024




#1 Tetrachloro-m-xylene

R.T.: 3.636 min  
Delta R.T.: -0.003 min  
Response: 47629152  
Conc: 19.95 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.348 min  
Delta R.T.: -0.012 min  
Response: 47275786  
Conc: 18.62 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.665 min  
Delta R.T.: -0.019 min  
Response: 44402755  
Conc: 16.43 ng/ml




#26 AR-1254-1

R.T.: 6.529 min  
Delta R.T.: -0.007 min  
Response: 7865143  
Conc: 86.51 ng/ml

Instrument :  
ECD\_P  
ClientSampleId :  
BASEMENT-CEILING


Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 04/16/2024  
Supervised By :Ankita Jodhani 04/16/2024




#26 AR-1254-1

R.T.: 5.530 min  
Delta R.T.: -0.009 min  
Response: 12013587  
Conc: 97.22 ng/ml




#27 AR-1254-2

R.T.: 6.751 min  
Delta R.T.: -0.007 min  
Response: 13014998  
Conc: 94.94 ng/ml



#27 AR-1254-2

R.T.: 5.679 min  
Delta R.T.: -0.009 min  
Response: 10629938  
Conc: 92.61 ng/ml




#28 AR-1254-3

R.T.: 7.120 min  
Delta R.T.: -0.006 min  
Response: 14012369  
Conc: 100.13 ng/ml

Instrument :  
ECD\_P  
ClientSampleId :  
BASEMENT-CEILING


Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 04/16/2024  
Supervised By :Ankita Jodhani 04/16/2024




#28 AR-1254-3

R.T.: 6.083 min  
Delta R.T.: -0.010 min  
Response: 16104917  
Conc: 89.50 ng/ml




#29 AR-1254-4

R.T.: 7.408 min  
Delta R.T.: -0.006 min  
Response: 9497071  
Conc: 104.24 ng/ml



#29 AR-1254-4

R.T.: 6.313 min  
Delta R.T.: -0.010 min  
Response: 9795687  
Conc: 106.30 ng/ml




#30 AR-1254-5

R.T.: 7.829 min  
Delta R.T.: -0.007 min  
Response: 8069556  
Conc: 68.63 ng/ml

Instrument :  
ECD\_P  
ClientSampleId :  
BASEMENT-CEILING

Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 04/16/2024  
Supervised By :Ankita Jodhani 04/16/2024



#30 AR-1254-5

R.T.: 6.732 min  
Delta R.T.: -0.011 min  
Response: 11513479  
Conc: 68.98 ng/ml m