

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041621\
 Data File : PP034942.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Apr 2021 9:13
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 16 17:27:17 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 31 06:02:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.844	4.259	3987826	2457914	52.276	54.715
2)	SA Decachlor...	10.763	9.554	3740964	1470484	48.597	52.078
Target Compounds							
3)	L1 AR-1016-1	6.156	5.510	1156459	602570	491.528	507.929
4)	L1 AR-1016-2	6.180	5.529	1810061	929809	509.631	523.650
5)	L1 AR-1016-3	6.245	5.722	1174881	502414	550.219	537.700
6)	L1 AR-1016-4	6.352	5.769	945524	419171	530.945	577.139
7)	L1 AR-1016-5	6.664	6.000	934430	474122	541.593	518.361
8)	L2 AR-1221-1	5.090	4.512	127048	81724	173.796	180.077
9)	L2 AR-1221-2	5.192	4.614	184562	120786	324.001	346.217
10)	L2 AR-1221-3	5.277	4.702	669203	425675	367.709	405.781
11)	L3 AR-1232-1	5.277	4.702	669203	425675	450.579	503.911
12)	L3 AR-1232-2	5.862	5.529	955883	929809	1259.641	1235.531
13)	L3 AR-1232-3	6.180	5.722	1810061	502414	1234.558	1265.784
14)	L3 AR-1232-4	6.352	5.814	945524	509055	1316.313	1479.975
15)	L3 AR-1232-5	6.449	6.000	757966	474122	1610.823	1333.828
16)	L4 AR-1242-1	6.156	5.510	1156459	602570	617.862	643.133
17)	L4 AR-1242-2	6.180	5.529	1810061	929809	641.500	654.053
18)	L4 AR-1242-3	6.245	5.722	1174881	502414	689.822	672.898
19)	L4 AR-1242-4	6.352	5.814	945524	509055	676.124	691.855
20)	L4 AR-1242-5	7.132	6.375	173848	413031	118.060	503.721 #
21)	L5 AR-1248-1	6.156	5.510	1156459	602570	813.014	825.978
22)	L5 AR-1248-2	6.449	5.769	757966	419171	383.024	408.347
23)	L5 AR-1248-3	6.664	5.814	934430	509055	393.097	464.470
24)	L5 AR-1248-4	7.092	6.000	223970	474122	87.766	375.653 #
25)	L5 AR-1248-5	7.132	6.420	173848	98157	68.224	86.854 #
26)	L6 AR-1254-1	7.061	6.375	664194	413031	241.566	223.876
27)	L6 AR-1254-2	7.289	6.532	681406	408225	160.250	255.159 #
28)	L6 AR-1254-3	7.670	6.967	504683	681739	110.575	273.159 #
29)	L6 AR-1254-4	7.964	7.189	395067	102993	127.776	78.349 #
30)	L6 AR-1254-5	8.393	7.614	2204259	1044513	620.248	513.923
31)	L7 AR-1260-1	7.837	7.086	1706463	891403	549.712	580.469
32)	L7 AR-1260-2	8.102	7.280	1889683	1056448	503.188	584.838
33)	L7 AR-1260-3	8.467	7.436	1501562	945258	496.758	544.855
34)	L7 AR-1260-4	8.696	7.914	1724762	763758	494.028	532.695
35)	L7 AR-1260-5	9.015	8.159	3355830	1724548	468.587	504.851
36)	L8 AR-1262-1	8.467	7.614	1501562	1044513	328.156	972.545 #
37)	L8 AR-1262-2	9.015	8.159	3355830	1724548	430.107	502.636
38)	L8 AR-1262-3	9.336f	8.442	2187960	411611	406.571	274.705 #
39)	L8 AR-1262-4	9.405	8.506	600650	1300375	135.255	472.155 #
40)	L8 AR-1262-5	10.048	9.003	1070783	493639	371.850	434.999
41)	L9 AR-1268-1	9.336f	8.442	2187960	411611	218.546	95.265 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041621\
 Data File : PP034942.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Apr 2021 9:13
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 16 17:27:17 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 31 06:02:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.405	8.506	600650	1300375	62.795	314.178 #
43) L9	AR-1268-3	9.625	8.712	76617	39799	8.725	10.940 #
44) L9	AR-1268-4	10.048	9.003	1070783	493639	335.209	383.178
45) L9	AR-1268-5	10.445	9.297	325237	151202	11.390	14.558 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041621\
Data File : PP034942.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Apr 2021 9:13
Operator : DD\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 16 17:27:17 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033021.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 31 06:02:19 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.844 min
Delta R.T.: -0.002 min
Response: 3987826
Conc: 52.28 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.259 min
Delta R.T.: 0.000 min
Response: 2457914
Conc: 54.71 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.763 min
Delta R.T.: -0.002 min
Response: 3740964
Conc: 48.60 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.554 min
Delta R.T.: -0.012 min
Response: 1470484
Conc: 52.08 ng/ml



#3 AR-1016-1


R.T.: 6.156 min
Delta R.T.: -0.001 min
Response: 1156459
Conc: 491.53 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 5.510 min
Delta R.T.: -0.003 min
Response: 602570
Conc: 507.93 ng/ml




#4 AR-1016-2

R.T.: 6.180 min
Delta R.T.: -0.001 min
Response: 1810061
Conc: 509.63 ng/ml



#4 AR-1016-2


R.T.: 5.529 min
Delta R.T.: -0.003 min
Response: 929809
Conc: 523.65 ng/ml



#5 AR-1016-3


R.T.: 6.245 min
Delta R.T.: 0.000 min
Response: 1174881
Conc: 550.22 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 5.722 min
Delta R.T.: -0.004 min
Response: 502414
Conc: 537.70 ng/ml




#6 AR-1016-4

R.T.: 6.352 min
Delta R.T.: 0.000 min
Response: 945524
Conc: 530.94 ng/ml



#6 AR-1016-4


R.T.: 5.769 min
Delta R.T.: -0.004 min
Response: 419171
Conc: 577.14 ng/ml



#7 AR-1016-5


R.T.: 6.664 min
Delta R.T.: -0.002 min
Response: 934430
Conc: 541.59 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 6.000 min
Delta R.T.: -0.004 min
Response: 474122
Conc: 518.36 ng/ml




#8 AR-1221-1

R.T.: 5.090 min
Delta R.T.: -0.006 min
Response: 127048
Conc: 173.80 ng/ml



#8 AR-1221-1


R.T.: 4.512 min
Delta R.T.: -0.002 min
Response: 81724
Conc: 180.08 ng/ml



#9 AR-1221-2


R.T.: 5.192 min
Delta R.T.: 0.000 min
Response: 184562
Conc: 324.00 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.614 min
Delta R.T.: 0.000 min
Response: 120786
Conc: 346.22 ng/ml




#10 AR-1221-3

R.T.: 5.277 min
Delta R.T.: -0.002 min
Response: 669203
Conc: 367.71 ng/ml



#10 AR-1221-3


R.T.: 4.702 min
Delta R.T.: -0.002 min
Response: 425675
Conc: 405.78 ng/ml



#11 AR-1232-1


R.T.: 5.277 min
Delta R.T.: -0.001 min
Response: 669203
Conc: 450.58 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.702 min
Delta R.T.: -0.002 min
Response: 425675
Conc: 503.91 ng/ml




#12 AR-1232-2

R.T.: 5.862 min
Delta R.T.: -0.002 min
Response: 955883
Conc: 1259.64 ng/ml



#12 AR-1232-2


R.T.: 5.529 min
Delta R.T.: -0.003 min
Response: 929809
Conc: 1235.53 ng/ml



#13 AR-1232-3


R.T.: 6.180 min
Delta R.T.: 0.000 min
Response: 1810061
Conc: 1234.56 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 5.722 min
Delta R.T.: -0.003 min
Response: 502414
Conc: 1265.78 ng/ml




#14 AR-1232-4

R.T.: 6.352 min
Delta R.T.: 0.000 min
Response: 945524
Conc: 1316.31 ng/ml



#14 AR-1232-4


R.T.: 5.814 min
Delta R.T.: -0.004 min
Response: 509055
Conc: 1479.98 ng/ml



#15 AR-1232-5


R.T.: 6.449 min
Delta R.T.: -0.002 min
Response: 757966
Conc: 1610.82 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 6.000 min
Delta R.T.: -0.003 min
Response: 474122
Conc: 1333.83 ng/ml




#16 AR-1242-1

R.T.: 6.156 min
Delta R.T.: -0.002 min
Response: 1156459
Conc: 617.86 ng/ml



#16 AR-1242-1


R.T.: 5.510 min
Delta R.T.: -0.003 min
Response: 602570
Conc: 643.13 ng/ml



#17 AR-1242-2


R.T.: 6.180 min
Delta R.T.: -0.002 min
Response: 1810061
Conc: 641.50 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 5.529 min
Delta R.T.: -0.003 min
Response: 929809
Conc: 654.05 ng/ml




#18 AR-1242-3

R.T.: 6.245 min
Delta R.T.: -0.002 min
Response: 1174881
Conc: 689.82 ng/ml



#18 AR-1242-3


R.T.: 5.722 min
Delta R.T.: -0.004 min
Response: 502414
Conc: 672.90 ng/ml



#19 AR-1242-4


R.T.: 6.352 min
Delta R.T.: -0.002 min
Response: 945524
Conc: 676.12 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.814 min
Delta R.T.: -0.006 min
Response: 509055
Conc: 691.85 ng/ml




#20 AR-1242-5

R.T.: 7.132 min
Delta R.T.: 0.000 min
Response: 173848
Conc: 118.06 ng/ml



#20 AR-1242-5


R.T.: 6.375 min
Delta R.T.: -0.006 min
Response: 413031
Conc: 503.72 ng/ml



#21 AR-1248-1


R.T.: 6.156 min
Delta R.T.: -0.001 min
Response: 1156459
Conc: 813.01 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.510 min
Delta R.T.: -0.003 min
Response: 602570
Conc: 825.98 ng/ml




#22 AR-1248-2

R.T.: 6.449 min
Delta R.T.: -0.002 min
Response: 757966
Conc: 383.02 ng/ml



#22 AR-1248-2


R.T.: 5.769 min
Delta R.T.: -0.004 min
Response: 419171
Conc: 408.35 ng/ml



#23 AR-1248-3


R.T.: 6.664 min
Delta R.T.: -0.002 min
Response: 934430
Conc: 393.10 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.814 min
Delta R.T.: -0.005 min
Response: 509055
Conc: 464.47 ng/ml




#24 AR-1248-4

R.T.: 7.092 min
Delta R.T.: -0.002 min
Response: 223970
Conc: 87.77 ng/ml



#24 AR-1248-4


R.T.: 6.000 min
Delta R.T.: -0.004 min
Response: 474122
Conc: 375.65 ng/ml



#25 AR-1248-5


R.T.: 7.132 min
Delta R.T.: 0.000 min
Response: 173848
Conc: 68.22 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 6.420 min
Delta R.T.: -0.005 min
Response: 98157
Conc: 86.85 ng/ml




#26 AR-1254-1

R.T.: 7.061 min
Delta R.T.: -0.002 min
Response: 664194
Conc: 241.57 ng/ml



#26 AR-1254-1


R.T.: 6.375 min
Delta R.T.: -0.006 min
Response: 413031
Conc: 223.88 ng/ml



#27 AR-1254-2


R.T.: 7.289 min
Delta R.T.: -0.004 min
Response: 681406
Conc: 160.25 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 6.532 min
Delta R.T.: -0.006 min
Response: 408225
Conc: 255.16 ng/ml




#28 AR-1254-3

R.T.: 7.670 min
Delta R.T.: -0.002 min
Response: 504683
Conc: 110.57 ng/ml



#28 AR-1254-3


R.T.: 6.967 min
Delta R.T.: 0.009 min
Response: 681739
Conc: 273.16 ng/ml



#29 AR-1254-4


R.T.: 7.964 min
Delta R.T.: -0.004 min
Response: 395067
Conc: 127.78 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 7.189 min
Delta R.T.: -0.006 min
Response: 102993
Conc: 78.35 ng/ml




#30 AR-1254-5

R.T.: 8.393 min
Delta R.T.: -0.002 min
Response: 2204259
Conc: 620.25 ng/ml



#30 AR-1254-5


R.T.: 7.614 min
Delta R.T.: -0.008 min
Response: 1044513
Conc: 513.92 ng/ml



#31 AR-1260-1


R.T.: 7.837 min
Delta R.T.: -0.001 min
Response: 1706463
Conc: 549.71 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 7.086 min
Delta R.T.: -0.007 min
Response: 891403
Conc: 580.47 ng/ml




#32 AR-1260-2

R.T.: 8.102 min
Delta R.T.: 0.000 min
Response: 1889683
Conc: 503.19 ng/ml



#32 AR-1260-2


R.T.: 7.280 min
Delta R.T.: -0.006 min
Response: 1056448
Conc: 584.84 ng/ml



#33 AR-1260-3


R.T.: 8.467 min
Delta R.T.: 0.000 min
Response: 1501562
Conc: 496.76 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 7.436 min
Delta R.T.: -0.007 min
Response: 945258
Conc: 544.85 ng/ml




#34 AR-1260-4

R.T.: 8.696 min
Delta R.T.: -0.001 min
Response: 1724762
Conc: 494.03 ng/ml



#34 AR-1260-4


R.T.: 7.914 min
Delta R.T.: -0.007 min
Response: 763758
Conc: 532.69 ng/ml



#35 AR-1260-5


R.T.: 9.015 min
Delta R.T.: 0.000 min
Response: 3355830
Conc: 468.59 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 8.159 min
Delta R.T.: -0.007 min
Response: 1724548
Conc: 504.85 ng/ml




#36 AR-1262-1

R.T.: 8.467 min
Delta R.T.: -0.001 min
Response: 1501562
Conc: 328.16 ng/ml



#36 AR-1262-1


R.T.: 7.614 min
Delta R.T.: -0.007 min
Response: 1044513
Conc: 972.55 ng/ml



#37 AR-1262-2


R.T.: 9.015 min
Delta R.T.: -0.002 min
Response: 3355830
Conc: 430.11 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 8.159 min
Delta R.T.: -0.007 min
Response: 1724548
Conc: 502.64 ng/ml




#38 AR-1262-3

R.T.: 9.336 min
Delta R.T.: 0.019 min
Response: 2187960
Conc: 406.57 ng/ml



#38 AR-1262-3


R.T.: 8.442 min
Delta R.T.: -0.007 min
Response: 411611
Conc: 274.71 ng/ml



#39 AR-1262-4


R.T.: 9.405 min
Delta R.T.: -0.004 min
Response: 600650
Conc: 135.25 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.506 min
Delta R.T.: -0.008 min
Response: 1300375
Conc: 472.16 ng/ml




#40 AR-1262-5

R.T.: 10.048 min
Delta R.T.: -0.001 min
Response: 1070783
Conc: 371.85 ng/ml



#40 AR-1262-5


R.T.: 9.003 min
Delta R.T.: -0.008 min
Response: 493639
Conc: 435.00 ng/ml



#41 AR-1268-1


R.T.: 9.336 min
Delta R.T.: 0.021 min
Response: 2187960
Conc: 218.55 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 8.442 min
Delta R.T.: -0.007 min
Response: 411611
Conc: 95.27 ng/ml




#42 AR-1268-2

R.T.: 9.405 min
Delta R.T.: -0.004 min
Response: 600650
Conc: 62.80 ng/ml



#42 AR-1268-2


R.T.: 8.506 min
Delta R.T.: -0.008 min
Response: 1300375
Conc: 314.18 ng/ml



#43 AR-1268-3


R.T.: 9.625 min
Delta R.T.: 0.000 min
Response: 76617
Conc: 8.72 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 8.712 min
Delta R.T.: -0.009 min
Response: 39799
Conc: 10.94 ng/ml




#44 AR-1268-4

R.T.: 10.048 min
Delta R.T.: 0.000 min
Response: 1070783
Conc: 335.21 ng/ml



#44 AR-1268-4


R.T.: 9.003 min
Delta R.T.: -0.006 min
Response: 493639
Conc: 383.18 ng/ml



#45 AR-1268-5

R.T.: 10.445 min
Delta R.T.: 0.000 min
Response: 325237
Conc: 11.39 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 9.297 min
Delta R.T.: -0.009 min
Response: 151202
Conc: 14.56 ng/ml