

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041621\
 Data File : PP034957.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Apr 2021 15:14
 Operator : DD\AJ
 Sample : M2034-03
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 16 17:33:38 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 31 06:02:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.843	4.259	1266462	857215	16.602	19.082
2)	SA Decachlor...	10.759	9.554	1559011	554341	20.252	19.632
Target Compounds							
3)	L1 AR-1016-1	6.156	5.529f	14123	50498	6.003	42.567 #
4)	L1 AR-1016-2	6.188	5.529	171116	50498	48.179	28.439 #
5)	L1 AR-1016-3	6.254	0.000	323202	0	151.362	N.D. #
6)	L1 AR-1016-4	0.000	5.769	0	44063	N.D.	60.669 #
7)	L1 AR-1016-5	0.000	5.999	0	31805	N.D.	34.772 #
8)	L2 AR-1221-1	5.092	0.000	37877	0	51.814	N.D. #
9)	L2 AR-1221-2	0.000	4.610	0	27568	N.D.	79.021 #
10)	L2 AR-1221-3	0.000	4.700	0	16543	N.D.	15.769 #
11)	L3 AR-1232-1	0.000	4.700	0	16543	N.D.	19.583 #
12)	L3 AR-1232-2	5.861	5.529	57393	50498	75.631	67.102
13)	L3 AR-1232-3	6.188	0.000	171116	0	116.710	N.D. #
14)	L3 AR-1232-4	0.000	5.812	0	124432	N.D.	361.761 #
15)	L3 AR-1232-5	6.447	5.999	64914	31805	137.955	89.475 #
16)	L4 AR-1242-1	6.156	5.529f	14123	50498	7.546	53.897 #
17)	L4 AR-1242-2	6.188	5.529	171116	50498	60.645	35.522 #
18)	L4 AR-1242-3	6.254	0.000	323202	0	189.766	N.D. #
19)	L4 AR-1242-4	0.000	5.812	0	124432	N.D.	169.115 #
20)	L4 AR-1242-5	7.132	6.376	97384	105679	66.134	128.884 #
21)	L5 AR-1248-1	6.156	5.529f	14123	50498	9.929	69.221 #
22)	L5 AR-1248-2	6.447	5.769	64914	44063	32.803	42.926 #
23)	L5 AR-1248-3	0.000	5.812	0	124432	N.D.	113.534 #
24)	L5 AR-1248-4	7.089	5.999	61135	31805	23.957	25.199
25)	L5 AR-1248-5	7.132	6.418	97384	26322	38.217	23.291 #
26)	L6 AR-1254-1	7.060	6.376	172696	105679	62.809	57.282
27)	L6 AR-1254-2	7.285	6.535	371782	150157	87.434	93.855
28)	L6 AR-1254-3	7.669	6.967	262338	180659	57.477	72.386 #
29)	L6 AR-1254-4	7.961	7.188	111508	30313	36.065	23.060 #
30)	L6 AR-1254-5	8.390	7.612	632125	276460	177.871	136.024
31)	L7 AR-1260-1	7.835	7.086	372622	222295	120.035	144.755
32)	L7 AR-1260-2	8.101	7.280	1497542	258943	398.768	143.348 #
33)	L7 AR-1260-3	8.464	7.436	338796	195319	112.083	112.584
34)	L7 AR-1260-4	8.694	7.914	467980	145766	134.045	101.667
35)	L7 AR-1260-5	9.013	8.159	771590	390858	107.740	114.421
36)	L8 AR-1262-1	8.464	7.612	338796	276460	74.042	257.411 #
37)	L8 AR-1262-2	9.013	8.159	771590	390858	98.892	113.919
38)	L8 AR-1262-3	9.335f	8.443	854364	73916	158.759	49.331 #
39)	L8 AR-1262-4	9.385f	8.506	310741	293241	69.973	106.473 #
40)	L8 AR-1262-5	10.045	9.003	213387	104060	74.103	91.699
41)	L9 AR-1268-1	9.335f	8.443	854364	73916	85.339	17.107 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041621\
 Data File : PP034957.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Apr 2021 15:14
 Operator : DD\AJ
 Sample : M2034-03
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 16 17:33:38 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 31 06:02:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.385f	8.506	310741	293241	32.487	70.849 #
43) L9	AR-1268-3	0.000	8.713	0	17509	N.D.	4.813 #
44) L9	AR-1268-4	10.045	9.003	213387	104060	66.801	80.775
45) L9	AR-1268-5	10.439	9.299	77254	79445	2.706	7.649 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041621\
Data File : PP034957.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Apr 2021 15:14
Operator : DD\AJ
Sample : M2034-03
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 16 17:33:38 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033021.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 31 06:02:19 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.843 min
Delta R.T.: -0.003 min
Response: 1266462
Conc: 16.60 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.259 min
Delta R.T.: -0.001 min
Response: 857215
Conc: 19.08 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.759 min
Delta R.T.: -0.006 min
Response: 1559011
Conc: 20.25 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.554 min
Delta R.T.: -0.012 min
Response: 554341
Conc: 19.63 ng/ml



#3 AR-1016-1


R.T.: 6.156 min
Delta R.T.: -0.001 min
Response: 14123
Conc: 6.00 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 5.529 min
Delta R.T.: 0.017 min
Response: 50498
Conc: 42.57 ng/ml




#4 AR-1016-2

R.T.: 6.188 min
Delta R.T.: 0.007 min
Response: 171116
Conc: 48.18 ng/ml



#4 AR-1016-2


R.T.: 5.529 min
Delta R.T.: -0.004 min
Response: 50498
Conc: 28.44 ng/ml



#5 AR-1016-3


R.T.: 6.254 min
Delta R.T.: 0.008 min
Response: 323202
Conc: 151.36 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 0.000 min
Exp R.T. : 5.725 min
Response: 0
Conc: N.D.




#6 AR-1016-4

R.T.: 0.000 min
Exp R.T. : 6.353 min
Response: 0
Conc: N.D.



#6 AR-1016-4


R.T.: 5.769 min
Delta R.T.: -0.005 min
Response: 44063
Conc: 60.67 ng/ml



#7 AR-1016-5


R.T.: 0.000 min
Exp R.T.: 6.666 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.999 min
Delta R.T.: -0.005 min
Response: 31805
Conc: 34.77 ng/ml




#8 AR-1221-1

R.T.: 5.092 min
Delta R.T.: -0.003 min
Response: 37877
Conc: 51.81 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T.: 4.514 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T.: 5.193 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.610 min
Delta R.T.: -0.005 min
Response: 27568
Conc: 79.02 ng/ml




#10 AR-1221-3

R.T.: 0.000 min
Exp R.T.: 5.279 min
Response: 0
Conc: N.D.



#10 AR-1221-3


R.T.: 4.700 min
Delta R.T.: -0.003 min
Response: 16543
Conc: 15.77 ng/ml



#11 AR-1232-1


R.T.: 0.000 min
Exp R.T.: 5.279 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.700 min
Delta R.T.: -0.003 min
Response: 16543
Conc: 19.58 ng/ml




#12 AR-1232-2

R.T.: 5.861 min
Delta R.T.: -0.002 min
Response: 57393
Conc: 75.63 ng/ml



#12 AR-1232-2


R.T.: 5.529 min
Delta R.T.: -0.004 min
Response: 50498
Conc: 67.10 ng/ml



#13 AR-1232-3


R.T.: 6.188 min
Delta R.T.: 0.007 min
Response: 171116
Conc: 116.71 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 0.000 min
Exp R.T. : 5.725 min
Response: 0
Conc: N.D.




#14 AR-1232-4

R.T.: 0.000 min
Exp R.T. : 6.353 min
Response: 0
Conc: N.D.



#14 AR-1232-4


R.T.: 5.812 min
Delta R.T.: -0.006 min
Response: 124432
Conc: 361.76 ng/ml



#15 AR-1232-5


R.T.: 6.447 min
Delta R.T.: -0.003 min
Response: 64914
Conc: 137.95 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.999 min
Delta R.T.: -0.005 min
Response: 31805
Conc: 89.47 ng/ml




#16 AR-1242-1

R.T.: 6.156 min
Delta R.T.: -0.002 min
Response: 14123
Conc: 7.55 ng/ml



#16 AR-1242-1


R.T.: 5.529 min
Delta R.T.: 0.016 min
Response: 50498
Conc: 53.90 ng/ml



#17 AR-1242-2


R.T.: 6.188 min
Delta R.T.: 0.006 min
Response: 171116
Conc: 60.64 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 5.529 min
Delta R.T.: -0.004 min
Response: 50498
Conc: 35.52 ng/ml




#18 AR-1242-3

R.T.: 6.254 min
Delta R.T.: 0.007 min
Response: 323202
Conc: 189.77 ng/ml



#18 AR-1242-3


R.T.: 0.000 min
Exp R.T.: 5.726 min
Response: 0
Conc: N.D.



#19 AR-1242-4


R.T.: 0.000 min
Exp R.T.: 6.354 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.812 min
Delta R.T.: -0.007 min
Response: 124432
Conc: 169.12 ng/ml




#20 AR-1242-5

R.T.: 7.132 min
Delta R.T.: -0.002 min
Response: 97384
Conc: 66.13 ng/ml



#20 AR-1242-5


R.T.: 6.376 min
Delta R.T.: -0.005 min
Response: 105679
Conc: 128.88 ng/ml



#21 AR-1248-1


R.T.: 6.156 min
Delta R.T.: -0.002 min
Response: 14123
Conc: 9.93 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.529 min
Delta R.T.: 0.016 min
Response: 50498
Conc: 69.22 ng/ml




#22 AR-1248-2

R.T.: 6.447 min
Delta R.T.: -0.004 min
Response: 64914
Conc: 32.80 ng/ml



#22 AR-1248-2


R.T.: 5.769 min
Delta R.T.: -0.005 min
Response: 44063
Conc: 42.93 ng/ml



#23 AR-1248-3


R.T.: 0.000 min
Exp R.T.: 6.666 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.812 min
Delta R.T.: -0.006 min
Response: 124432
Conc: 113.53 ng/ml




#24 AR-1248-4

R.T.: 7.089 min
Delta R.T.: -0.005 min
Response: 61135
Conc: 23.96 ng/ml



#24 AR-1248-4


R.T.: 5.999 min
Delta R.T.: -0.005 min
Response: 31805
Conc: 25.20 ng/ml



#25 AR-1248-5


R.T.: 7.132 min
Delta R.T.: 0.000 min
Response: 97384
Conc: 38.22 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 6.418 min
Delta R.T.: -0.006 min
Response: 26322
Conc: 23.29 ng/ml




#26 AR-1254-1

R.T.: 7.060 min
Delta R.T.: -0.003 min
Response: 172696
Conc: 62.81 ng/ml



#26 AR-1254-1


R.T.: 6.376 min
Delta R.T.: -0.005 min
Response: 105679
Conc: 57.28 ng/ml



#27 AR-1254-2


R.T.: 7.285 min
Delta R.T.: -0.008 min
Response: 371782
Conc: 87.43 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 6.535 min
Delta R.T.: -0.003 min
Response: 150157
Conc: 93.85 ng/ml




#28 AR-1254-3

R.T.: 7.669 min
Delta R.T.: -0.003 min
Response: 262338
Conc: 57.48 ng/ml



#28 AR-1254-3


R.T.: 6.967 min
Delta R.T.: 0.009 min
Response: 180659
Conc: 72.39 ng/ml



#29 AR-1254-4


R.T.: 7.961 min
Delta R.T.: -0.007 min
Response: 111508
Conc: 36.06 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 7.188 min
Delta R.T.: -0.007 min
Response: 30313
Conc: 23.06 ng/ml




#30 AR-1254-5

R.T.: 8.390 min
Delta R.T.: -0.005 min
Response: 632125
Conc: 177.87 ng/ml



#30 AR-1254-5


R.T.: 7.612 min
Delta R.T.: -0.009 min
Response: 276460
Conc: 136.02 ng/ml



#31 AR-1260-1


R.T.: 7.835 min
Delta R.T.: -0.003 min
Response: 372622
Conc: 120.03 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 7.086 min
Delta R.T.: -0.006 min
Response: 222295
Conc: 144.76 ng/ml




#32 AR-1260-2

R.T.: 8.101 min
Delta R.T.: -0.001 min
Response: 1497542
Conc: 398.77 ng/ml



#32 AR-1260-2


R.T.: 7.280 min
Delta R.T.: -0.006 min
Response: 258943
Conc: 143.35 ng/ml



#33 AR-1260-3


R.T.: 8.464 min
Delta R.T.: -0.003 min
Response: 338796
Conc: 112.08 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 7.436 min
Delta R.T.: -0.007 min
Response: 195319
Conc: 112.58 ng/ml




#34 AR-1260-4

R.T.: 8.694 min
Delta R.T.: -0.003 min
Response: 467980
Conc: 134.04 ng/ml



#34 AR-1260-4


R.T.: 7.914 min
Delta R.T.: -0.008 min
Response: 145766
Conc: 101.67 ng/ml



#35 AR-1260-5


R.T.: 9.013 min
Delta R.T.: -0.003 min
Response: 771590
Conc: 107.74 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 8.159 min
Delta R.T.: -0.007 min
Response: 390858
Conc: 114.42 ng/ml




#36 AR-1262-1

R.T.: 8.464 min
Delta R.T.: -0.003 min
Response: 338796
Conc: 74.04 ng/ml



#36 AR-1262-1


R.T.: 7.612 min
Delta R.T.: -0.008 min
Response: 276460
Conc: 257.41 ng/ml



#37 AR-1262-2


R.T.: 9.013 min
Delta R.T.: -0.005 min
Response: 771590
Conc: 98.89 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 8.159 min
Delta R.T.: -0.007 min
Response: 390858
Conc: 113.92 ng/ml




#38 AR-1262-3

R.T.: 9.335 min
Delta R.T.: 0.017 min
Response: 854364
Conc: 158.76 ng/ml



#38 AR-1262-3


R.T.: 8.443 min
Delta R.T.: -0.007 min
Response: 73916
Conc: 49.33 ng/ml



#39 AR-1262-4


R.T.: 9.385 min
Delta R.T.: -0.023 min
Response: 310741
Conc: 69.97 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.506 min
Delta R.T.: -0.008 min
Response: 293241
Conc: 106.47 ng/ml




#40 AR-1262-5

R.T.: 10.045 min
Delta R.T.: -0.005 min
Response: 213387
Conc: 74.10 ng/ml



#40 AR-1262-5


R.T.: 9.003 min
Delta R.T.: -0.008 min
Response: 104060
Conc: 91.70 ng/ml



#41 AR-1268-1


R.T.: 9.335 min
Delta R.T.: 0.019 min
Response: 854364
Conc: 85.34 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 8.443 min
Delta R.T.: -0.007 min
Response: 73916
Conc: 17.11 ng/ml




#42 AR-1268-2

R.T.: 9.385 min
Delta R.T.: -0.023 min
Response: 310741
Conc: 32.49 ng/ml



#42 AR-1268-2


R.T.: 8.506 min
Delta R.T.: -0.008 min
Response: 293241
Conc: 70.85 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 9.625 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 8.713 min
Delta R.T.: -0.008 min
Response: 17509
Conc: 4.81 ng/ml




#44 AR-1268-4

R.T.: 10.045 min
Delta R.T.: -0.003 min
Response: 213387
Conc: 66.80 ng/ml



#44 AR-1268-4


R.T.: 9.003 min
Delta R.T.: -0.006 min
Response: 104060
Conc: 80.77 ng/ml



#45 AR-1268-5

R.T.: 10.439 min
Delta R.T.: -0.006 min
Response: 77254
Conc: 2.71 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 9.299 min
Delta R.T.: -0.008 min
Response: 79445
Conc: 7.65 ng/ml