

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041621\
 Data File : PP034972.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Apr 2021 19:44
 Operator : DD\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Manual Integrations
 APPROVED

Ankita
 4/20/2021 11:12:52 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 17 06:39:11 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 31 06:02:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.842	4.258	4223131	2671429	55.360	59.468
2)	SA Decachlor...	10.757	9.553	3927238	1585647	51.016	56.157
Target Compounds							
26)	L6 AR-1254-1	7.058	6.374	1412825	1042980	513.842	565.329
27)	L6 AR-1254-2	7.287	6.531	2177501	905816	512.096	566.174
28)	L6 AR-1254-3	7.667	6.951	2312123	1406641	506.580	563.612
29)	L6 AR-1254-4	7.962	7.187	1549057	775837	501.011	590.195m
30)	L6 AR-1254-5	8.390	7.613	1788129	1203945	503.155	592.367

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041621\
Data File : PP034972.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Apr 2021 19:44
Operator : DD\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampled :
AR1254CCC500


Manual Integrations
APPROVED

Ankita
4/20/2021 11:12:52 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 17 06:39:11 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033021.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 31 06:02:19 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.842 min
Delta R.T.: -0.004 min
Response: 4223131
Conc: 55.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Ankita
4/20/2021 11:12:52 AM




#1 Tetrachloro-m-xylene

R.T.: 4.258 min
Delta R.T.: -0.002 min
Response: 2671429
Conc: 59.47 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.757 min
Delta R.T.: -0.008 min
Response: 3927238
Conc: 51.02 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.553 min
Delta R.T.: -0.013 min
Response: 1585647
Conc: 56.16 ng/ml




#26 AR-1254-1

R.T.: 7.058 min
Delta R.T.: -0.005 min
Response: 1412825
Conc: 513.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Ankita
4/20/2021 11:12:52 AM




#26 AR-1254-1

R.T.: 6.374 min
Delta R.T.: -0.007 min
Response: 1042980
Conc: 565.33 ng/ml




#27 AR-1254-2

R.T.: 7.287 min
Delta R.T.: -0.006 min
Response: 2177501
Conc: 512.10 ng/ml



#27 AR-1254-2

R.T.: 6.531 min
Delta R.T.: -0.007 min
Response: 905816
Conc: 566.17 ng/ml




#28 AR-1254-3

R.T.: 7.667 min
Delta R.T.: -0.006 min
Response: 2312123
Conc: 506.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Ankita
4/20/2021 11:12:52 AM




#28 AR-1254-3

R.T.: 6.951 min
Delta R.T.: -0.008 min
Response: 1406641
Conc: 563.61 ng/ml




#29 AR-1254-4

R.T.: 7.962 min
Delta R.T.: -0.006 min
Response: 1549057
Conc: 501.01 ng/ml



#29 AR-1254-4

R.T.: 7.187 min
Delta R.T.: -0.008 min
Response: 775837
Conc: 590.20 ng/ml m




#30 AR-1254-5

R.T.: 8.390 min
Delta R.T.: -0.005 min
Response: 1788129
Conc: 503.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Manual Integrations
APPROVED

Ankita
4/20/2021 11:12:52 AM



#30 AR-1254-5

R.T.: 7.613 min
Delta R.T.: -0.008 min
Response: 1203945
Conc: 592.37 ng/ml