

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041622\
 Data File : PP046152.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Apr 2022 02:16
 Operator : YP\AJ
 Sample : N2394-11MS
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 CS-6-(DUMPSTER-LEFT)MS

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/18/2022
 Supervised By :mohammad ahmed 04/18/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 18 01:06:06 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 12 05:31:12 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.872	3.137	50535124	37295827	20.545m	23.472
2) SA Decachlor...	9.973	8.469	36443000	27703154	22.200	19.436
Target Compounds						
3) L1 AR-1016-1	5.121	4.271	39087762	31136356	509.818	530.189
4) L1 AR-1016-2	5.144	4.289	56869031	44830400	491.788	536.602
5) L1 AR-1016-3	5.210	4.474	34601159	23897613	502.198	540.991
6) L1 AR-1016-4	5.315	4.522	29646220	18592363	503.807	521.018
7) L1 AR-1016-5	5.635	4.746	24643296	22865914	446.880	481.330
31) L7 AR-1260-1	6.862	5.849	44203021	45049677	500.746	569.187
32) L7 AR-1260-2	7.145	6.057	57804175	51840090	564.243	553.792
33) L7 AR-1260-3	7.540	6.217	37989545	47502135	461.122	536.242
34) L7 AR-1260-4	7.788	6.728	43990876	34303610	519.129	477.725
35) L7 AR-1260-5	8.129	6.997	81241560	80207287	494.675	472.183

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041622\
Data File : PP046152.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Apr 2022 02:16
Operator : YP\AJ
Sample : N2394-11MS
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
CS-6-(DUMPSTER-LEFT)MS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/18/2022
Supervised By :mohammad ahmed 04/18/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 18 01:06:06 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041222.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 12 05:31:12 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

