

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041624\
 Data File : PP064377.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Apr 2024 22:57
 Operator : YP\AJ
 Sample : AR1268ICC050
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/17/2024
 Supervised By :Ankita Jodhani 04/17/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 17 04:47:19 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 17 04:22:26 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.479	3.629	16525038	10571172	4.629	4.262
2) SA Decachlor...	10.337	8.659	19858826	19186002	4.919	5.000
Target Compounds						
41) L9 AR-1268-1	8.775	7.553	14187179	14463877	47.443m	46.340m
42) L9 AR-1268-2	8.873	7.618	12551936	13520301	47.966	46.153
43) L9 AR-1268-3	9.110	7.825	12080292	12473960	48.514	48.025
44) L9 AR-1268-4	9.550	8.114	3313273	3896884	39.485	45.756
45) L9 AR-1268-5	9.983	8.404	34454048	28473802	45.667	44.567

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041624\
Data File : PP064377.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Apr 2024 22:57
Operator : YP\AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1268ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/17/2024
Supervised By :Ankita Jodhani 04/17/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 17 04:47:19 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041624.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 17 04:22:26 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.479 min
Delta R.T.: 0.000 min
Response: 16525038
Conc: 4.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/17/2024
Supervised By :Ankita Jodhani 04/17/2024




#1 Tetrachloro-m-xylene

R.T.: 3.629 min
Delta R.T.: 0.000 min
Response: 10571172
Conc: 4.26 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.337 min
Delta R.T.: 0.000 min
Response: 19858826
Conc: 4.92 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.659 min
Delta R.T.: 0.000 min
Response: 19186002
Conc: 5.00 ng/ml




#41 AR-1268-1

R.T.: 8.775 min
Delta R.T.: 0.000 min
Response: 14187179
Conc: 47.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/17/2024
Supervised By :Ankita Jodhani 04/17/2024




#41 AR-1268-1

R.T.: 7.553 min
Delta R.T.: 0.000 min
Response: 14463877
Conc: 46.34 ng/ml m




#42 AR-1268-2

R.T.: 8.873 min
Delta R.T.: 0.000 min
Response: 12551936
Conc: 47.97 ng/ml



#42 AR-1268-2

R.T.: 7.618 min
Delta R.T.: 0.000 min
Response: 13520301
Conc: 46.15 ng/ml




#43 AR-1268-3

R.T.: 9.110 min
Delta R.T.: 0.000 min
Response: 12080292
Conc: 48.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/17/2024
Supervised By :Ankita Jodhani 04/17/2024




#43 AR-1268-3

R.T.: 7.825 min
Delta R.T.: 0.000 min
Response: 12473960
Conc: 48.02 ng/ml




#44 AR-1268-4

R.T.: 9.550 min
Delta R.T.: 0.000 min
Response: 3313273
Conc: 39.49 ng/ml



#44 AR-1268-4

R.T.: 8.114 min
Delta R.T.: 0.000 min
Response: 3896884
Conc: 45.76 ng/ml




#45 AR-1268-5

R.T.: 9.983 min
Delta R.T.: 0.000 min
Response: 34454048
Conc: 45.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/17/2024
Supervised By :Ankita Jodhani 04/17/2024



#45 AR-1268-5

R.T.: 8.404 min
Delta R.T.: 0.000 min
Response: 28473802
Conc: 44.57 ng/ml