

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041624\
 Data File : PP064380.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Apr 2024 23:46
 Operator : YP\AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP041624AR1248

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 17 04:01:21 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 17 03:47:53 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.479	3.629	167.2E6	117.1E6	47.530	47.740
2)	SA Decachlor...	10.334	8.657	128.3E6	118.8E6	54.307	52.553
Target Compounds							
21)	L5 AR-1248-1	5.661	4.720	30414056	20346139	476.399	473.699
22)	L5 AR-1248-2	5.937	4.958	48595910	34330517	482.915	485.411
23)	L5 AR-1248-3	6.142	4.999	51852521	36008249	484.496	487.535
24)	L5 AR-1248-4	6.549	5.171	48693027	41853927	485.016	484.502
25)	L5 AR-1248-5	6.588	5.565	50315507	35113816	494.600	494.202



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041624\
Data File : PP064380.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Apr 2024 23:46
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP041624AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 17 04:01:21 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041624.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 17 03:47:53 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.479 min
Delta R.T.: -0.001 min
Response: 167171866
Conc: 47.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP041624AR1248




#1 Tetrachloro-m-xylene

R.T.: 3.629 min
Delta R.T.: -0.001 min
Response: 117061576
Conc: 47.74 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.334 min
Delta R.T.: -0.004 min
Response: 128306390
Conc: 54.31 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.657 min
Delta R.T.: -0.004 min
Response: 118819272
Conc: 52.55 ng/ml



#21 AR-1248-1


R.T.: 5.661 min
Delta R.T.: -0.002 min
Response: 30414056
Conc: 476.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP041624AR1248




#21 AR-1248-1

R.T.: 4.720 min
Delta R.T.: -0.002 min
Response: 20346139
Conc: 473.70 ng/ml




#22 AR-1248-2

R.T.: 5.937 min
Delta R.T.: -0.002 min
Response: 48595910
Conc: 482.92 ng/ml



#22 AR-1248-2


R.T.: 4.958 min
Delta R.T.: -0.002 min
Response: 34330517
Conc: 485.41 ng/ml



#23 AR-1248-3


R.T.: 6.142 min
Delta R.T.: -0.002 min
Response: 51852521
Conc: 484.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP041624AR1248




#23 AR-1248-3

R.T.: 4.999 min
Delta R.T.: -0.002 min
Response: 36008249
Conc: 487.53 ng/ml




#24 AR-1248-4

R.T.: 6.549 min
Delta R.T.: -0.002 min
Response: 48693027
Conc: 485.02 ng/ml



#24 AR-1248-4


R.T.: 5.171 min
Delta R.T.: -0.002 min
Response: 41853927
Conc: 484.50 ng/ml



#25 AR-1248-5

R.T.: 6.588 min
Delta R.T.: -0.002 min
Response: 50315507
Conc: 494.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP041624AR1248



#25 AR-1248-5

R.T.: 5.565 min
Delta R.T.: -0.003 min
Response: 35113816
Conc: 494.20 ng/ml