

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041822\
 Data File : PP046253.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Apr 2022 21:46
 Operator : YP\AJ
 Sample : N2446-01
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 339

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/19/2022
 Supervised By :mohammad ahmed 04/19/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 19 04:57:28 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 12 05:31:12 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.870	3.135	45948587	33434258	18.681m	21.041
2) SA Decachlor...	9.963	8.464	20792567	20331384	12.666	14.264

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041822\
Data File : PP046253.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Apr 2022 21:46
Operator : YP\AJ
Sample : N2446-01
Misc :
ALS Vial : 27 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
339


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/19/2022
Supervised By :mohammad ahmed 04/19/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 19 04:57:28 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041222.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 12 05:31:12 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.870 min
Delta R.T.: -0.006 min
Response: 45948587
Conc: 18.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
339


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/19/2022
Supervised By :mohammad ahmed 04/19/2022




#1 Tetrachloro-m-xylene

R.T.: 3.135 min
Delta R.T.: -0.003 min
Response: 33434258
Conc: 21.04 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.963 min
Delta R.T.: -0.015 min
Response: 20792567
Conc: 12.67 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.464 min
Delta R.T.: -0.013 min
Response: 20331384
Conc: 14.26 ng/ml