

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041823\
 Data File : PP056967.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Apr 2023 17:08
 Operator : YP\AJ
 Sample : AR1254ICC1000
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 19 00:43:05 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 19 00:41:26 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.404	3.629	233.1E6	171.2E6	97.509	96.823
2) SA Decachlor...	10.225	8.688	87637737	128.6E6	94.310	93.168
Target Compounds						
26) L6 AR-1254-1	6.442	5.537	61312563	85772392	915.875	927.667
27) L6 AR-1254-2	6.662	5.685	84980620	73128834	926.596	923.246
28) L6 AR-1254-3	7.031	6.093	80357946	116.0E6	944.613	940.148
29) L6 AR-1254-4	7.317	6.323	52846415	67903716	933.215	934.851
30) L6 AR-1254-5	7.739	6.743	54307036	97248689	940.222	926.157



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041823\
Data File : PP056967.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Apr 2023 17:08
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 19 00:43:05 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041823.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 19 00:41:26 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.404 min
Delta R.T.: 0.000 min
Response: 233141470
Conc: 97.51 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.629 min
Delta R.T.: 0.000 min
Response: 171222826
Conc: 96.82 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.225 min
Delta R.T.: 0.003 min
Response: 87637737
Conc: 94.31 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.688 min
Delta R.T.: 0.000 min
Response: 128630753
Conc: 93.17 ng/ml



#26 AR-1254-1


R.T.: 6.442 min
Delta R.T.: 0.000 min
Response: 61312563
Conc: 915.87 ng/ml

Instrument :
ECD_P
ClientSampleId :




#26 AR-1254-1

R.T.: 5.537 min
Delta R.T.: 0.000 min
Response: 85772392
Conc: 927.67 ng/ml




#27 AR-1254-2

R.T.: 6.662 min
Delta R.T.: 0.000 min
Response: 84980620
Conc: 926.60 ng/ml



#27 AR-1254-2


R.T.: 5.685 min
Delta R.T.: 0.000 min
Response: 73128834
Conc: 923.25 ng/ml



#28 AR-1254-3


R.T.: 7.031 min
Delta R.T.: 0.000 min
Response: 80357946
Conc: 944.61 ng/ml

Instrument :
ECD_P
ClientSampleId :




#28 AR-1254-3

R.T.: 6.093 min
Delta R.T.: 0.000 min
Response: 116048465
Conc: 940.15 ng/ml




#29 AR-1254-4

R.T.: 7.317 min
Delta R.T.: 0.000 min
Response: 52846415
Conc: 933.22 ng/ml



#29 AR-1254-4


R.T.: 6.323 min
Delta R.T.: 0.000 min
Response: 67903716
Conc: 934.85 ng/ml



#30 AR-1254-5

R.T.: 7.739 min
Delta R.T.: 0.000 min
Response: 54307036
Conc: 940.22 ng/ml

Instrument :
ECD_P
ClientSampleId :



#30 AR-1254-5

R.T.: 6.743 min
Delta R.T.: 0.000 min
Response: 97248689
Conc: 926.16 ng/ml