

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041823\
 Data File : PP056974.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Apr 2023 19:04
 Operator : YP\AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 19 01:23:47 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 19 01:21:22 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.402	3.628	178.8E6	131.1E6	76.076	75.382
2) SA Decachlor...	10.223	8.688	118.6E6	174.0E6	74.210	75.093
Target Compounds						
41) L9 AR-1268-1	8.682	7.570	109.0E6	162.8E6	755.199	747.302
42) L9 AR-1268-2	8.778	7.634	95280219	147.3E6	750.905	751.863
43) L9 AR-1268-3	9.013	7.843	89140171	128.4E6	750.750	747.849
44) L9 AR-1268-4	9.447	8.130	31471686	49265748	745.887	738.583
45) L9 AR-1268-5	9.874	8.427	239.0E6	364.5E6	757.421	757.170



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041823\
Data File : PP056974.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Apr 2023 19:04
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 19 01:23:47 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041823.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 19 01:21:22 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.402 min
Delta R.T.: 0.000 min
Response: 178840930
Conc: 76.08 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.628 min
Delta R.T.: 0.000 min
Response: 131134335
Conc: 75.38 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.223 min
Delta R.T.: 0.000 min
Response: 118574196
Conc: 74.21 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.688 min
Delta R.T.: 0.000 min
Response: 173977434
Conc: 75.09 ng/ml



#41 AR-1268-1


R.T.: 8.682 min
Delta R.T.: 0.000 min
Response: 108972941
Conc: 755.20 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.570 min
Delta R.T.: 0.000 min
Response: 162799464
Conc: 747.30 ng/ml




#42 AR-1268-2

R.T.: 8.778 min
Delta R.T.: 0.000 min
Response: 95280219
Conc: 750.90 ng/ml



#42 AR-1268-2


R.T.: 7.634 min
Delta R.T.: 0.000 min
Response: 147279963
Conc: 751.86 ng/ml



#43 AR-1268-3


R.T.: 9.013 min
Delta R.T.: 0.002 min
Response: 89140171
Conc: 750.75 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.843 min
Delta R.T.: 0.000 min
Response: 128449449
Conc: 747.85 ng/ml




#44 AR-1268-4

R.T.: 9.447 min
Delta R.T.: 0.000 min
Response: 31471686
Conc: 745.89 ng/ml



#44 AR-1268-4


R.T.: 8.130 min
Delta R.T.: 0.000 min
Response: 49265748
Conc: 738.58 ng/ml



#45 AR-1268-5

R.T.: 9.874 min
Delta R.T.: 0.000 min
Response: 238955309
Conc: 757.42 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.427 min
Delta R.T.: 0.000 min
Response: 364466772
Conc: 757.17 ng/ml