

Data Path : P:\HPCHEM1\ECD_P\Data\PP041918\
 Data File : PP025179.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Apr 2018 3:16
 Operator : UA/AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 20 07:00:40 2018
 Quant Method : P:\HPCHEM1\ECD_P\methods\PP032918.M
 Quant Title : 8080.M
 QLast Update : Mon Apr 02 02:18:04 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.413	11.116	447.8E6	475.3E6	846.149	800.963
Target Compounds							
1) T	Dalapon	2.904	2.451	523.7E6	563.1E6	678.894	665.258
2) T	3,5-DICHL...	9.820	9.242	619.1E6	590.1E6	751.231	731.961
3) T	4-Nitroph...	10.989	10.276	269.7E6	258.7E6	736.716	724.918
5) T	DICAMBA	11.730	11.410	1844.2E6	1701.3E6	747.205	736.578
6) T	MCPP	12.072	11.620	123.7E6	108.8E6	79.298	75.900
7) T	MCPA	12.327	12.011	176.1E6	156.3E6	75.698	74.070
8) T	DICHLORPROP	12.922	12.521	452.6E6	421.3E6	760.791	715.297
9) T	2,4-D	13.282	13.006	569.3E6	560.8E6	777.154	738.755
10) T	Pentachlo...	13.667	13.553	7430.2E6	6537.5E6	791.587	736.564
11) T	2,4,5-TP ...	14.540	14.157	3085.6E6	2726.1E6	731.501	733.542
12) T	2,4,5-T	14.940	14.781	2863.5E6	2240.3E6	785.856	750.954
13) T	2,4-DB	15.687	15.380	356.6E6	309.5E6	785.749	731.213
14) T	DINOSEB	17.155	15.777	2062.5E6	1749.9E6	744.133	715.687
15) T	Picloram	16.961	17.111	4497.0E6	3707.7E6	768.365	770.440
16) T	DCPA	17.524	16.992	3760.8E6	3155.9E6	776.860	729.967



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : P:\HPCHEM1\ECD_P\Data\PP041918\
Data File : PP025179.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Apr 2018 3:16
Operator : UA/AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 20 07:00:40 2018
Quant Method : P:\HPCHEM1\ECD_P\methods\PP032918.M
Quant Title : 8080.M
QLast Update : Mon Apr 02 02:18:04 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.904 min
Delta R.T.: 0.008 min
Response: 523677536
Conc: 678.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750




#1 Dalapon

R.T.: 2.451 min
Delta R.T.: -0.020 min
Response: 563141021
Conc: 665.26 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 9.820 min
Delta R.T.: -0.022 min
Response: 619137062
Conc: 751.23 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 9.242 min
Delta R.T.: -0.049 min
Response: 590061051
Conc: 731.96 ng/ml



#3 4-Nitrophenol


R.T.: 10.989 min
Delta R.T.: -0.031 min
Response: 269715511
Conc: 736.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750




#3 4-Nitrophenol

R.T.: 10.276 min
Delta R.T.: -0.052 min
Response: 258701400
Conc: 724.92 ng/ml




#4 2,4-DCAA

R.T.: 11.413 min
Delta R.T.: -0.033 min
Response: 447843941
Conc: 846.15 ng/ml



#4 2,4-DCAA


R.T.: 11.116 min
Delta R.T.: -0.051 min
Response: 475302070
Conc: 800.96 ng/ml



#5 DICAMBA


R.T.: 11.730 min
Delta R.T.: -0.034 min
Response: 1844238131
Conc: 747.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750




#5 DICAMBA

R.T.: 11.410 min
Delta R.T.: -0.051 min
Response: 1701334977
Conc: 736.58 ng/ml




#6 MCP

R.T.: 12.072 min
Delta R.T.: -0.034 min
Response: 123683605
Conc: 79.30 ug/ml



#6 MCP


R.T.: 11.620 min
Delta R.T.: -0.050 min
Response: 108832940
Conc: 75.90 ug/ml



#7 MCPA


R.T.: 12.327 min
Delta R.T.: -0.035 min
Response: 176063154
Conc: 75.70 ug/ml

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750




#7 MCPA

R.T.: 12.011 min
Delta R.T.: -0.050 min
Response: 156271747
Conc: 74.07 ug/ml




#8 DICHLORPROP

R.T.: 12.922 min
Delta R.T.: -0.037 min
Response: 452635130
Conc: 760.79 ng/ml



#8 DICHLORPROP


R.T.: 12.521 min
Delta R.T.: -0.052 min
Response: 421252184
Conc: 715.30 ng/ml



#9 2,4-D


R.T.: 13.282 min
Delta R.T.: -0.038 min
Response: 569289717
Conc: 777.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750




#9 2,4-D

R.T.: 13.006 min
Delta R.T.: -0.052 min
Response: 560821853
Conc: 738.76 ng/ml




#10 Pentachlorophenol

R.T.: 13.667 min
Delta R.T.: -0.037 min
Response: 7430225052
Conc: 791.59 ng/ml



#10 Pentachlorophenol


R.T.: 13.553 min
Delta R.T.: -0.055 min
Response: 6537546183
Conc: 736.56 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 14.540 min
Delta R.T.: -0.038 min
Response: 3085594866
Conc: 731.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750




#11 2,4,5-TP (SILVEX)

R.T.: 14.157 min
Delta R.T.: -0.053 min
Response: 2726107413
Conc: 733.54 ng/ml




#12 2,4,5-T

R.T.: 14.940 min
Delta R.T.: -0.039 min
Response: 2863514905
Conc: 785.86 ng/ml



#12 2,4,5-T


R.T.: 14.781 min
Delta R.T.: -0.055 min
Response: 2240324073
Conc: 750.95 ng/ml



#13 2,4-DB


R.T.: 15.687 min
Delta R.T.: -0.039 min
Response: 356567135
Conc: 785.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750




#13 2,4-DB

R.T.: 15.380 min
Delta R.T.: -0.054 min
Response: 309539221
Conc: 731.21 ng/ml




#14 DINOSEB

R.T.: 17.155 min
Delta R.T.: -0.043 min
Response: 2062546093
Conc: 744.13 ng/ml



#14 DINOSEB


R.T.: 15.777 min
Delta R.T.: -0.054 min
Response: 1749938542
Conc: 715.69 ng/ml



#15 Picloram


R.T.: 16.961 min
Delta R.T.: -0.043 min
Response: 4497010965
Conc: 768.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750




#15 Picloram

R.T.: 17.111 min
Delta R.T.: -0.056 min
Response: 3707673292
Conc: 770.44 ng/ml



#16 DCPA

R.T.: 17.524 min
Delta R.T.: -0.044 min
Response: 3760830107
Conc: 776.86 ng/ml



#16 DCPA

R.T.: 16.992 min
Delta R.T.: -0.055 min
Response: 3155858350
Conc: 729.97 ng/ml