

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041922\
 Data File : PP046355.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Apr 2022 15:12
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 20 15:25:52 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 19 10:41:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.881	3.136	126.5E6	88242493	51.433	55.534
2) SA Decachlor...	9.975	8.466	72539872	72951979	44.189	51.183
Target Compounds						
21) L5 AR-1248-1	5.128	4.269	22627437	17668910	524.530	515.230
22) L5 AR-1248-2	5.423	4.521	28975314	23756257	484.685	507.104
23) L5 AR-1248-3	5.641	4.563	32827528	24854072	469.286	508.486
24) L5 AR-1248-4	6.080	4.744	35571165	29538016	449.906	508.386
25) L5 AR-1248-5	6.120	5.162	34447307	30002149	439.978	511.152


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041922\
Data File : PP046355.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Apr 2022 15:12
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 20 15:25:52 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041222.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 19 10:41:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.881 min
Delta R.T.: 0.012 min
Response: 126510015
Conc: 51.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.136 min
Delta R.T.: 0.003 min
Response: 88242493
Conc: 55.53 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.975 min
Delta R.T.: 0.003 min
Response: 72539872
Conc: 44.19 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.466 min
Delta R.T.: 0.003 min
Response: 72951979
Conc: 51.18 ng/ml



#21 AR-1248-1


R.T.: 5.128 min
Delta R.T.: 0.003 min
Response: 22627437
Conc: 524.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.269 min
Delta R.T.: -0.006 min
Response: 17668910
Conc: 515.23 ng/ml




#22 AR-1248-2

R.T.: 5.423 min
Delta R.T.: 0.003 min
Response: 28975314
Conc: 484.69 ng/ml



#22 AR-1248-2


R.T.: 4.521 min
Delta R.T.: -0.007 min
Response: 23756257
Conc: 507.10 ng/ml



#23 AR-1248-3


R.T.: 5.641 min
Delta R.T.: 0.003 min
Response: 32827528
Conc: 469.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.563 min
Delta R.T.: -0.007 min
Response: 24854072
Conc: 508.49 ng/ml




#24 AR-1248-4

R.T.: 6.080 min
Delta R.T.: 0.003 min
Response: 35571165
Conc: 449.91 ng/ml



#24 AR-1248-4


R.T.: 4.744 min
Delta R.T.: -0.007 min
Response: 29538016
Conc: 508.39 ng/ml



#25 AR-1248-5

R.T.: 6.120 min
Delta R.T.: 0.002 min
Response: 34447307
Conc: 439.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.162 min
Delta R.T.: -0.007 min
Response: 30002149
Conc: 511.15 ng/ml