

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041923\
 Data File : PP057008.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Apr 2023 15:16
 Operator : YP\AJ
 Sample : 02270-13
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 SB-24-0-2-20230411

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/20/2023
 Supervised By :Ankita Jodhani 04/20/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 19 22:12:08 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 19 04:08:43 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.401	3.628	44927920	32958912	18.792	18.552
2) SA Decachlor...	10.217	8.686	31073567	33926177	33.788	24.535m#
Target Compounds						
31) L7 AR-1260-1	7.193	6.223	16343167	32601533	228.807m	359.764 #
32) L7 AR-1260-2	7.449	6.422	43336501	86307756	598.050	845.939 #
33) L7 AR-1260-3	7.813	6.567	14885019	22591584	273.171	230.839
34) L7 AR-1260-4	8.038	7.041	23904825	16914568	394.605	216.091 #
35) L7 AR-1260-5	8.364	7.283	28097300	43226758	267.958	262.618

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP041923\
Data File : PP057008.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Apr 2023 15:16
Operator : YP\AJ
Sample : 02270-13
Misc :
ALS Vial : 26 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
SB-24-0-2-20230411


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/20/2023
Supervised By :Ankita Jodhani 04/20/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 19 22:12:08 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041823.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 19 04:08:43 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.401 min
Delta R.T.: -0.003 min
Response: 44927920
Conc: 18.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB-24-0-2-20230411


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/20/2023
Supervised By :Ankita Jodhani 04/20/2023




#1 Tetrachloro-m-xylene

R.T.: 3.628 min
Delta R.T.: -0.001 min
Response: 32958912
Conc: 18.55 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.217 min
Delta R.T.: -0.007 min
Response: 31073567
Conc: 33.79 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.686 min
Delta R.T.: -0.002 min
Response: 33926177
Conc: 24.54 ng/ml m




#31 AR-1260-1

R.T.: 7.193 min
Delta R.T.: -0.003 min
Response: 16343167
Conc: 228.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB-24-0-2-20230411


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/20/2023
Supervised By :Ankita Jodhani 04/20/2023




#31 AR-1260-1

R.T.: 6.223 min
Delta R.T.: -0.001 min
Response: 32601533
Conc: 359.76 ng/ml




#32 AR-1260-2

R.T.: 7.449 min
Delta R.T.: -0.006 min
Response: 43336501
Conc: 598.05 ng/ml



#32 AR-1260-2

R.T.: 6.422 min
Delta R.T.: 0.010 min
Response: 86307756
Conc: 845.94 ng/ml




#33 AR-1260-3

R.T.: 7.813 min
Delta R.T.: -0.004 min
Response: 14885019
Conc: 273.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB-24-0-2-20230411


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/20/2023
Supervised By :Ankita Jodhani 04/20/2023




#33 AR-1260-3

R.T.: 6.567 min
Delta R.T.: -0.002 min
Response: 22591584
Conc: 230.84 ng/ml




#34 AR-1260-4

R.T.: 8.038 min
Delta R.T.: -0.005 min
Response: 23904825
Conc: 394.60 ng/ml



#34 AR-1260-4

R.T.: 7.041 min
Delta R.T.: -0.001 min
Response: 16914568
Conc: 216.09 ng/ml




#35 AR-1260-5

R.T.: 8.364 min
Delta R.T.: -0.003 min
Response: 28097300
Conc: 267.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB-24-0-2-20230411

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/20/2023
Supervised By :Ankita Jodhani 04/20/2023



#35 AR-1260-5

R.T.: 7.283 min
Delta R.T.: -0.001 min
Response: 43226758
Conc: 262.62 ng/ml