

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042022\
 Data File : PP046360.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Apr 2022 16:47
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 21 05:25:58 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 19 10:41:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.874	3.135	128.1E6	88359601	52.084	55.608
2) SA Decachlor...	9.962	8.462	73421106	75379349	44.726	52.886
Target Compounds						
21) L5 AR-1248-1	5.119	4.268	24146875	17743510	559.752	517.405
22) L5 AR-1248-2	5.414	4.519	31831692	23814385	532.465	508.345
23) L5 AR-1248-3	5.632	4.561	37302528	25010622	533.259	511.689
24) L5 AR-1248-4	6.071	4.741	40059309	29664067	506.672	510.556
25) L5 AR-1248-5	6.112	5.160	38800709	30063632	495.582	512.200



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042022\
Data File : PP046360.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Apr 2022 16:47
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 21 05:25:58 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041222.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 19 10:41:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.874 min
Delta R.T.: 0.005 min
Response: 128111987
Conc: 52.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.135 min
Delta R.T.: 0.002 min
Response: 88359601
Conc: 55.61 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.962 min
Delta R.T.: -0.010 min
Response: 73421106
Conc: 44.73 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.462 min
Delta R.T.: -0.001 min
Response: 75379349
Conc: 52.89 ng/ml



#21 AR-1248-1


R.T.: 5.119 min
Delta R.T.: -0.005 min
Response: 24146875
Conc: 559.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.268 min
Delta R.T.: -0.008 min
Response: 17743510
Conc: 517.41 ng/ml




#22 AR-1248-2

R.T.: 5.414 min
Delta R.T.: -0.006 min
Response: 31831692
Conc: 532.47 ng/ml



#22 AR-1248-2


R.T.: 4.519 min
Delta R.T.: -0.008 min
Response: 23814385
Conc: 508.35 ng/ml



#23 AR-1248-3


R.T.: 5.632 min
Delta R.T.: -0.006 min
Response: 37302528
Conc: 533.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.561 min
Delta R.T.: -0.009 min
Response: 25010622
Conc: 511.69 ng/ml




#24 AR-1248-4

R.T.: 6.071 min
Delta R.T.: -0.006 min
Response: 40059309
Conc: 506.67 ng/ml



#24 AR-1248-4


R.T.: 4.741 min
Delta R.T.: -0.009 min
Response: 29664067
Conc: 510.56 ng/ml



#25 AR-1248-5

R.T.: 6.112 min
Delta R.T.: -0.006 min
Response: 38800709
Conc: 495.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.160 min
Delta R.T.: -0.010 min
Response: 30063632
Conc: 512.20 ng/ml