

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042022\
 Data File : PP046374.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Apr 2022 21:18
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 21 08:26:57 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 21 08:16:42 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.869	3.134	130.7E6	87216868	53.149	54.889
2) SA Decachlor...	9.955	8.459	77020437	77105168	46.919	54.097
Target Compounds						
16) L4 AR-1242-1	5.113	4.267	27480049	23338203	459.807m	504.417
17) L4 AR-1242-2	5.137	4.285	44577461	33215512	496.581	512.226
18) L4 AR-1242-3	5.203	4.469	26736683	18196365	481.970	525.865
19) L4 AR-1242-4	5.308	4.560	22218901	17518391	489.608	521.172
20) L4 AR-1242-5	6.106	5.115	22061937	22325542	474.982	508.343



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042022\
Data File : PP046374.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Apr 2022 21:18
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 21 08:26:57 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP041222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 21 08:16:42 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.869 min
Delta R.T.: -0.003 min
Response: 130730976
Conc: 53.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.134 min
Delta R.T.: 0.000 min
Response: 87216868
Conc: 54.89 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.955 min
Delta R.T.: -0.005 min
Response: 77020437
Conc: 46.92 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.459 min
Delta R.T.: -0.002 min
Response: 77105168
Conc: 54.10 ng/ml



#16 AR-1242-1


R.T.: 5.113 min
Delta R.T.: -0.013 min
Response: 27480049
Conc: 459.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.267 min
Delta R.T.: -0.009 min
Response: 23338203
Conc: 504.42 ng/ml




#17 AR-1242-2

R.T.: 5.137 min
Delta R.T.: -0.012 min
Response: 44577461
Conc: 496.58 ng/ml



#17 AR-1242-2


R.T.: 4.285 min
Delta R.T.: -0.009 min
Response: 33215512
Conc: 512.23 ng/ml



#18 AR-1242-3


R.T.: 5.203 min
Delta R.T.: -0.011 min
Response: 26736683
Conc: 481.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.469 min
Delta R.T.: -0.010 min
Response: 18196365
Conc: 525.87 ng/ml




#19 AR-1242-4

R.T.: 5.308 min
Delta R.T.: -0.012 min
Response: 22218901
Conc: 489.61 ng/ml



#19 AR-1242-4


R.T.: 4.560 min
Delta R.T.: -0.010 min
Response: 17518391
Conc: 521.17 ng/ml



#20 AR-1242-5

R.T.: 6.106 min
Delta R.T.: -0.013 min
Response: 22061937
Conc: 474.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.115 min
Delta R.T.: -0.011 min
Response: 22325542
Conc: 508.34 ng/ml