

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042023\
 Data File : PP057049.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Apr 2023 11:38
 Operator : YP\AJ
 Sample : AR1242ICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 20 16:09:19 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 20 16:08:10 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.404	3.631	112.4E6	85926527	50.000	50.000
2) SA Decachlor...	10.216	8.685	78653466	89065774	50.000	50.000
Target Compounds						
16) L4 AR-1242-1	5.578	4.727	25642853	23438374	500.000	500.000
17) L4 AR-1242-2	5.601	4.747	37406017	32769221	500.000	500.000
18) L4 AR-1242-3	5.664	4.925	23868697	18064022	500.000	500.000
19) L4 AR-1242-4	5.762	5.009	18928993	19930007	500.000	500.000
20) L4 AR-1242-5	6.503	5.536	17041876	22556694	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042023\
Data File : PP057049.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Apr 2023 11:38
Operator : YP\AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 20 16:09:19 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042023.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 20 16:08:10 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.404 min
Delta R.T.: 0.000 min
Response: 112409154
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.631 min
Delta R.T.: 0.000 min
Response: 85926527
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.216 min
Delta R.T.: 0.000 min
Response: 78653466
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.685 min
Delta R.T.: 0.000 min
Response: 89065774
Conc: 50.00 ng/ml



#16 AR-1242-1


R.T.: 5.578 min
Delta R.T.: 0.000 min
Response: 25642853
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500




#16 AR-1242-1

R.T.: 4.727 min
Delta R.T.: 0.000 min
Response: 23438374
Conc: 500.00 ng/ml




#17 AR-1242-2

R.T.: 5.601 min
Delta R.T.: 0.000 min
Response: 37406017
Conc: 500.00 ng/ml



#17 AR-1242-2


R.T.: 4.747 min
Delta R.T.: 0.000 min
Response: 32769221
Conc: 500.00 ng/ml



#18 AR-1242-3


R.T.: 5.664 min
Delta R.T.: 0.000 min
Response: 23868697
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500




#18 AR-1242-3

R.T.: 4.925 min
Delta R.T.: 0.000 min
Response: 18064022
Conc: 500.00 ng/ml




#19 AR-1242-4

R.T.: 5.762 min
Delta R.T.: 0.000 min
Response: 18928993
Conc: 500.00 ng/ml



#19 AR-1242-4


R.T.: 5.009 min
Delta R.T.: 0.000 min
Response: 19930007
Conc: 500.00 ng/ml



#20 AR-1242-5

R.T.: 6.503 min
Delta R.T.: 0.000 min
Response: 17041876
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500



#20 AR-1242-5

R.T.: 5.536 min
Delta R.T.: 0.000 min
Response: 22556694
Conc: 500.00 ng/ml